

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042622\
 Data File : BG053313.D
 Acq On : 26 Apr 2022 9:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 03:36:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.557	0.553	0.7	83	0.00
3	Pyridine	1.470	1.609	-9.5	86	0.00
4	n-Nitrosodimethylamine	0.728	0.781	-7.3	85	0.00
5 S	2-Fluorophenol	1.167	1.233	-5.7	87	0.00
6	Aniline	2.085	2.402	-15.2	93	0.00
7 S	Phenol-d6	1.799	2.006	-11.5	90	0.00
8	2-Chlorophenol	1.260	1.360	-7.9	87	0.00
9	Benzaldehyde	1.075	1.091	-1.5	93	0.00
10 C	Phenol	1.791	2.046	-14.2	95	0.00
11	bis(2-Chloroethyl)ether	1.291	1.420	-10.0	91	0.00
12	1,3-Dichlorobenzene	1.464	1.560	-6.6	88	0.00
13 C	1,4-Dichlorobenzene	1.492	1.593	-6.8	88	0.00
14	1,2-Dichlorobenzene	1.439	1.529	-6.3	89	0.00
15	Benzyl Alcohol	1.599	1.819	-13.8	91	0.00
16	2,2'-oxybis(1-Chloropropane	2.155	2.367	-9.8	88	0.00
17	2-Methylphenol	1.212	1.406	-16.0	96	0.00
18	Hexachloroethane	0.560	0.579	-3.4	84	0.00
19 P	n-Nitroso-di-n-propylamine	1.317	1.520	-15.4	93	0.00
20	3+4-Methylphenols	1.711	1.988	-16.2	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.555	0.599	-7.9	95	0.00
23 S	Nitrobenzene-d5	0.493	0.506	-2.6	91	0.00
24	Nitrobenzene	0.479	0.487	-1.7	92	0.00
25	Isophorone	0.837	0.925	-10.5	96	0.00
26 C	2-Nitrophenol	0.200	0.218	-9.0	94	0.00
27	2,4-Dimethylphenol	0.287	0.305	-6.3	92	0.00
28	bis(2-Chloroethoxy)methane	0.470	0.498	-6.0	92	0.00
29 C	2,4-Dichlorophenol	0.361	0.386	-6.9	92	0.00
30	1,2,4-Trichlorobenzene	0.408	0.413	-1.2	89	0.00
31	Naphthalene	1.067	1.138	-6.7	94	0.00
32	Benzoic acid	0.155	0.203	-31.0#	93	0.00
33	4-Chloroaniline	0.457	0.503	-10.1	95	0.00
34 C	Hexachlorobutadiene	0.303	0.302	0.3	88	0.00
35	Caprolactam	0.127	0.156	-22.8	105	0.00
36 C	4-Chloro-3-methylphenol	0.420	0.479	-14.0	98	0.00
37	2-Methylnaphthalene	0.790	0.859	-8.7	96	0.00
38	1-Methylnaphthalene	0.771	0.836	-8.4	96	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	0.679	0.661	2.7	96	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.296	15.7	80	0.00
42 S	2,4,6-Tribromophenol	0.318	0.345	-8.5	104	0.00
43 C	2,4,6-Trichlorophenol	0.468	0.460	1.7	94	0.00
44	2,4,5-Trichlorophenol	0.499	0.491	1.6	94	0.00
45 S	2-Fluorobiphenyl	1.450	1.447	0.2	98	0.00
46	1,1'-Biphenyl	1.440	1.440	0.0	99	0.00
47	2-Chloronaphthalene	1.149	1.157	-0.7	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042622\
 Data File : BG053313.D
 Acq On : 26 Apr 2022 9:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 03:36:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.436	0.456	-4.6	98	0.00
49	Acenaphthylene	1.799	1.869	-3.9	100	0.00
50	Dimethylphthalate	1.600	1.695	-5.9	103	0.00
51	2,6-Dinitrotoluene	0.334	0.354	-6.0	103	0.00
52 C	Acenaphthene	1.220	1.275	-4.5	101	0.00
53	3-Nitroaniline	0.353	0.388	-9.9	105	0.00
54 P	2,4-Dinitrophenol	0.212	0.222	-4.7	98	0.00
55	Dibenzofuran	1.910	2.003	-4.9	103	0.00
56 P	4-Nitrophenol	0.269	0.281	-4.5	96	0.00
57	2,4-Dinitrotoluene	0.475	0.515	-8.4	103	0.00
58	Fluorene	1.543	1.612	-4.5	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.485	0.513	-5.8	101	0.00
60	Diethylphthalate	1.676	1.766	-5.4	103	0.00
61	4-Chlorophenyl-phenylether	0.875	0.919	-5.0	102	0.00
62	4-Nitroaniline	0.372	0.408	-9.7	107	0.00
63	Azobenzene	1.690	1.767	-4.6	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.140	0.151	-7.9	107	0.00
66 c	n-Nitrosodiphenylamine	0.575	0.590	-2.6	104	0.00
67	4-Bromophenyl-phenylether	0.256	0.261	-2.0	101	0.00
68	Hexachlorobenzene	0.278	0.282	-1.4	102	0.00
69	Atrazine	0.225	0.225	0.0	103	0.00
70 C	Pentachlorophenol	0.152	0.152	0.0	94	0.00
71	Phenanthrene	1.092	1.127	-3.2	104	0.00
72	Anthracene	1.082	1.117	-3.2	105	0.00
73	Carbazole	1.052	1.082	-2.9	104	0.00
74	Di-n-butylphthalate	1.193	1.214	-1.8	103	0.00
75 C	Fluoranthene	1.400	1.395	0.4	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzydine	0.571	0.598	-4.7	111	0.00
78	Pyrene	1.424	1.470	-3.2	101	0.00
79 S	Terphenyl-d14	1.184	1.210	-2.2	101	0.00
80	Butylbenzylphthalate	0.531	0.537	-1.1	98	0.00
81	Benzo(a)anthracene	1.367	1.386	-1.4	100	0.00
82	3,3'-Dichlorobenzidine	0.503	0.493	2.0	97	0.00
83	Chrysene	1.269	1.282	-1.0	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.718	0.721	-0.4	100	0.00
85 c	Di-n-octyl phthalate	1.201	1.221	-1.7	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	1.486	1.526	-2.7	98	0.00
88	Benzo(b)fluoranthene	1.262	1.307	-3.6	100	-0.01
89	Benzo(k)fluoranthene	1.240	1.262	-1.8	98	0.00
90 C	Benzo(a)pyrene	1.075	1.107	-3.0	99	-0.01
91	Dibenzo(a,h)anthracene	1.223	1.261	-3.1	99	0.00
92	Benzo(g,h,i)perylene	1.205	1.234	-2.4	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042622\
Data File : BG053313.D
Acq On : 26 Apr 2022 9:49
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 27 03:36:05 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0