

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042715\
 Data File : BG016736.D
 Acq On : 27 Apr 2015 14:54
 Operator : TP/IZ
 Sample : G1970-02DL 2X
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 1052-CHAMBERLAND-BOTTOMDL

Quant Time: Apr 27 16:21:34 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 19:19:11 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.58	152	64026	20.00	ng	-0.02
21) Naphthalene-d8	10.38	136	231844	20.00	ng	0.00
38) Acenaphthene-d10	14.27	164	199219	20.00	ng	0.02
63) Phenanthrene-d10	17.02	188	219884	20.00	ng	0.03
75) Chrysene-d12	21.18	240	247405	20.00	ng	0.00
86) Perylene-d12	23.37	264	245898	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.17	112	179387	45.47	ng	-0.02
7) Phenol-d6	6.78	99	269416	48.67	ng	-0.02
23) Nitrobenzene-d5	8.75	82	222875	60.74	ng	-0.02
41) 2,4,6-Tribromophenol	15.77	330	40622	27.49	ng	0.02
44) 2-Fluorobiphenyl	12.87	172	380912	30.72	ng	0.00
78) Terphenyl-d14	19.64	244	319475	29.75	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	10.43	128	1551245	126.90	ng	# 88
37) 2-Methylnaphthalene	12.08	142	4027919	456.22	ng	# 83
45) 1,1'-Biphenyl	13.09	154	888894	56.81	ng	96
51) Acenaphthene	14.33	154	352378	27.87	ng	# 89
57) Fluorene	15.33	166	737918	47.34	ng	# 83
70) Phenanthrene	17.07	178	2019574	159.88	ng	# 84
77) Pyrene	19.44	202	722276	41.56	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042715\
Data File : BG016736.D
Acq On : 27 Apr 2015 14:54
Operator : TP/IZ
Sample : G1970-02DL 2X
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
1052-CHAMBERLAND-BOTTOMDL

Quant Time: Apr 27 16:21:34 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 23 19:19:11 2015
Response via : Initial Calibration

