

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042715\
 Data File : BG016762.D
 Acq On : 28 Apr 2015 10:22
 Operator : TP/IZ
 Sample : G2020-03 2X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SW-018

Quant Time: Apr 28 14:30:38 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 23 19:19:11 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.57	152	62202	20.00	ng	-0.03
21) Naphthalene-d8	10.35	136	268614	20.00	ng	-0.03
38) Acenaphthene-d10	14.23	164	148129	20.00	ng	-0.02
63) Phenanthrene-d10	16.97	188	281787	20.00	ng	-0.02
75) Chrysene-d12	21.16	240	246613	20.00	ng	-0.02
86) Perylene-d12	23.36	264	237850	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.17	112	204711	53.41	ng	-0.02
7) Phenol-d6	6.77	99	288310	53.61	ng	-0.03
23) Nitrobenzene-d5	8.73	82	143611	33.78	ng	-0.03
41) 2,4,6-Tribromophenol	15.72	330	62365	56.75	ng	-0.03
44) 2-Fluorobiphenyl	12.84	172	356497	38.67	ng	-0.03
78) Terphenyl-d14	19.60	244	322691	30.14	ng	-0.03
Target Compounds						Qvalue
70) Phenanthrene	17.01	178	78210	4.83	ng	98
73) Di-n-butylphthalate	17.94	149	120268	6.25	ng	99
74) Fluoranthene	19.03	202	129646	7.56	ng	97
77) Pyrene	19.40	202	115625	6.67	ng	97
80) Benzo(a)anthracene	21.14	228	60833	4.07	ng	97
82) Chrysene	21.19	228	54451	3.80	ng	96
83) Bis(2-ethylhexyl)phthalate	21.08	149	183273	15.81	ng	97
87) Benzo(b)fluoranthene	22.70	252	67359	4.66	ng	# 93
89) Benzo(a)pyrene	23.26	252	46609	3.42	ng	# 90
91) Benzo(g,h,i)perylene	26.27	276	28144	2.09	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042715\
Data File : BG016762.D
Acq On : 28 Apr 2015 10:22
Operator : TP/IZ
Sample : G2020-03 2X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SW-018

Quant Time: Apr 28 14:30:38 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 23 19:19:11 2015
Response via : Initial Calibration

