

Data Path : Z:\HPCHEM1\BNA G\Data\BG042715\
 Data File : BG016745.D
 Acq On : 27 Apr 2015 21:55
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 28 01:10:41 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 28 00:54:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	-0.03
2	1,4-Dioxane	0.534	0.515	3.6	109	-0.04
3	Pyridine	1.516	1.432	5.5	107	-0.03
4	n-Nitrosodimethylamine	0.447	0.437	2.2	114	-0.03
5 S	2-Fluorophenol	1.232	1.239	-0.6	114	-0.02
6	Aniline	2.371	2.353	0.8	111	-0.02
7 S	Phenol-d6	1.729	1.752	-1.3	113	-0.02
8	2-Chlorophenol	1.523	1.538	-1.0	116	-0.03
9	Benzaldehyde	0.706	0.667	5.5	121	-0.02
10 C	Phenol	1.823	1.799	1.3	112	-0.02
11	bis(2-Chloroethyl)ether	1.423	1.381	3.0	113	-0.03
12	1,3-Dichlorobenzene	1.575	1.603	-1.8	118	-0.03
13 C	1,4-Dichlorobenzene	1.610	1.631	-1.3	117	-0.03
14	1,2-Dichlorobenzene	1.537	1.558	-1.4	118	-0.02
15	Benzyl Alcohol	1.085	1.123	-3.5	113	-0.02
16	2,2'-oxybis(1-Chloropropane	1.332	1.217	8.6	105	-0.02
17	2-Methylphenol	1.219	1.275	-4.6	116	-0.02
18	Hexachloroethane	0.568	0.582	-2.5	120	-0.02
19 P	n-Nitroso-di-n-propylamine	1.078	1.082	-0.4	110	-0.02
20	3+4-Methylphenols	1.713	1.775	-3.6	116	-0.02
21 I	Naphthalene-d8	1.000	1.000	0.0	115	-0.03
22						

Data Path : Z:\HPCHEM1\BNA G\Data\BG042715\
 Data File : BG016745.D
 Acq On : 27 Apr 2015 21:55
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 28 01:10:41 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 28 00:54:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.571	1.641	-4.5	116	-0.02
46	2-Chloronaphthalene	1.207	1.257	-4.1	116	-0.02
47	2-Nitroaniline	0.329	0.339	-3.0	108	-0.02
48	Acenaphthylene	2.115	2.231	-5.5	116	-0.02
49	Dimethylphthalate	1.496	1.591	-6.4	117	-0.02
50	2,6-Dinitrotoluene	0.367	0.391	-6.5	115	-0.02
51 C	Acenaphthene	1.269	1.321	-4.1	116	-0.02
52	3-Nitroaniline	0.444	0.470	-5.9	110	-0.02
53 P	2,4-Dinitrophenol	0.164	0.142	13.4	89	-0.02
54	Dibenzofuran	1.790	1.861	-4.0	116	-0.02
55 P	4-Nitrophenol	0.333	0.326	2.1	105	-0.02
56	2,4-Dinitrotoluene	0.503	0.546	-8.5	113	-0.02
57	Fluorene	1.565	1.646	-5.2	117	-0.02
58	2,3,4,6-Tetrachlorophenol	0.279	0.326	-16.8	126	-0.02
59	Diethylphthalate	1.559	1.665	-6.8	116	-0.02
60	4-Chlorophenyl-phenylether	0.654	0.703	-7.5	120	-0.02
61	4-Nitroaniline	0.491	0.473	3.7	99	-0.02
62	Azobenzene	1.424	1.393	2.2	107	-0.02
63 I	Phenanthrene-d10	1.000	1.000	0.0	115	-0.02
64	4,6-Dinitro-2-methylphenol	0.135	0.141	-4.4	111	-0.02
65 c	n-Nitrosodiphenylamine	0.686	0.723	-5.4		

Data Path : Z:\HPCHEM1\BNA G\Data\BG042715\
Data File : BG016745.D
Acq On : 27 Apr 2015 21:55
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: Apr 28 01:10:41 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG042315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 28 00:54:32 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.186	1.226	-3.4	117	-0.03
89 C	Benzo(a)pyrene	1.145	1.178	-2.9	117	-0.03
90	Dibenzo(a,h)anthracene	1.113	1.074	3.5	109	-0.05
91	Benzo(a,h,i)perylene	1.133	1.124	0.8	113	-0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0