

Data File : BG026731.D
Acq On : 28 Apr 2017 3:09
Operator : SJ/MA
Sample : I2807-13
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
JHSN1

Quant Time: Apr 28 03:50:56 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 28 01:45:47 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	160065	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	793134	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	456468	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.36	188	870270	20.00	ng/ul	0.00
75) Chrysene-d12	21.60	240	762258	20.00	ng/ul	0.00
83) Perylene-d12	24.76	264	790030	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.43	96	22319	6.28	ng/uL	0.00
5) Phenol-d5	7.20	99	556284	38.94	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.33	67	316048	37.62	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	439351	36.23	ng/ul	0.00
13) 4-Methylphenol-d8	8.73	113	497627	41.97	ng/ul	0.00
19) Nitrobenzene-d5	9.17	128	252637	38.11	ng/ul	0.00
22) 2-Nitrophenol-d4	9.89	143	289145	40.10	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.44	165	449729	39.25	ng/ul	0.00
29) 4-Chloroaniline-d4	10.95	131	693193	45.20	ng/ul	0.00
43) Dimethylphthalate-d6	14.02	166	1438952	39.16	ng/ul	0.00
46) Acenaphthylene-d8	14.32	160	1830206	39.66	ng/ul	0.00
51) 4-Nitrophenol-d4	14.85	143	264190	39.89	ng/ul	0.00
57) Fluorene-d10	15.61	176	1243956	38.79	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.73	200	207063	39.18	ng/ul	0.00
70) Anthracene-d10	17.46	188	1714841	40.60	ng/ul	0.00
76) Pyrene-d10	19.74	212	1579658	39.26	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.55	264	1575076	42.17	ng/ul	0.01

Target Compounds						Qvalue
44) Dimethylphthalate	14.07	163	268229	7.46	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BG026731.D

Acq On : 28 Apr 2017 3:09

Operator : SJ/MA

Sample : I2807-13

Misc :

ALS Vial : 22 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

JHSN1


Quant Time: Apr 28 03:50:56 2017




Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M

Quant Title : SVOA CALIBRATION

QLast Update : Fri Apr 28 01:45:47 2017

Response via : Initial Calibration





#44

Dimethylphthalate

Concen: 7.46 ng/ul

RT: 14.07 min Scan# 1869

Delta R.T. 0.00 min

Lab File: BG026731.D

Acq: 28 Apr 2017 3:09

Instrument :

BNA_G

ClientSampleId :

JHSN1

Tgt Ion:163 Resp: 268229

Ion Ratio Lower Upper

163 100

194 4.1 3.4 5.0

164 10.7 9.0 13.4

