

Data File : BG026758.D
Acq On : 28 Apr 2017 22:18
Operator : SJ/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTD02062

Quant Time: Apr 29 00:40:46 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 29 00:39:01 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	178363	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	884611	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	515123	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.35	188	1009261	20.00	ng/ul	0.00
75) Chrysene-d12	21.60	240	909719	20.00	ng/ul	0.00
83) Perylene-d12	24.76	264	943222	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.43	96	31680	8.00	ng/uL	0.00
5) Phenol-d5	7.19	99	346309	21.76	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.33	67	192804	20.60	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	280094	20.73	ng/ul	0.00
13) 4-Methylphenol-d8	8.73	113	288853	21.86	ng/ul	0.00
19) Nitrobenzene-d5	9.17	128	148293	20.05	ng/ul	0.00
22) 2-Nitrophenol-d4	9.89	143	166344	20.68	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.44	165	270776	21.19	ng/ul	0.00
29) 4-Chloroaniline-d4	10.94	131	392027	22.92	ng/ul	0.00
43) Dimethylphthalate-d6	14.02	166	810407	19.55	ng/ul	0.00
46) Acenaphthylene-d8	14.32	160	1045186	20.07	ng/ul	0.00
51) 4-Nitrophenol-d4	14.84	143	150021	20.07	ng/ul	0.00
57) Fluorene-d10	15.61	176	719068	19.87	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.73	200	123730	20.19	ng/ul	0.00
70) Anthracene-d10	17.45	188	974661	19.90	ng/ul	0.00
76) Pyrene-d10	19.73	212	925215	19.27	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.54	264	894326	20.05	ng/ul	0.00

Target Compounds

						Qvalue
6) Phenol	7.21	94	352692	21.86	ng/ul#	73
8) Bis(2-Chloroethyl)ether	7.42	93	248881	19.98	ng/ul	86
10) 2-Chlorophenol	7.58	128	271131	20.39	ng/ul#	78
11) 2-Methylphenol	8.46	108	281220	22.02	ng/ul	92
12) 2,2'-oxybis(1-Chloropropan	8.54	45	275128	21.17	ng/ul#	88
14) Acetophenone	8.83	105	410704	21.16	ng/ul#	72
15) N-Nitroso-di-n-propylamine	8.81	70	192792	20.14	ng/ul#	70
16) 4-Methylphenol	8.79	108	309651	22.17	ng/ul	96
17) Hexachloroethane	9.09	117	99325	19.86	ng/ul#	72
20) Nitrobenzene	9.21	77	284323	19.93	ng/ul#	75
21) Isophorone	9.73	82	547874	20.00	ng/ul#	90
23) 2-Nitrophenol	9.92	139	173445	20.51	ng/ul#	60
24) 2,4-Dimethylphenol	9.99	107	324446	21.05	ng/ul#	80
25) Bis(2-Chloroethoxy)methane	10.21	93	368079	19.67	ng/ul#	93
27) 2,4-Dichlorophenol	10.47	162	270175	21.20	ng/ul	94
28) Naphthalene	10.86	128	908407	19.73	ng/ul	97
30) 4-Chloroaniline	10.97	127	366561	22.72	ng/ul	95
31) Hexachlorobutadiene	11.14	225	122119	18.94	ng/ul#	86
32) Caprolactam	11.71	113	105031	22.06	ng/ul#	57
33) 4-Chloro-3-methylphenol	12.10	107	315036	22.80	ng/ul#	81
34) 2-Methylnaphthalene	12.46	142	697922	20.42	ng/ul	96
36) 1,2,4,5-Tetrachlorobenzene	12.82	216	262727	19.34	ng/ul#	93
37) Hexachlorocyclopentadiene	12.81	237	117021	19.56	ng/ul	96

Data File : BG026758.D
Acq On : 28 Apr 2017 22:18
Operator : SJ/MA
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SSTD02062

Quant Time: Apr 29 00:40:46 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 29 00:39:01 2017
Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
38) 2,4,6-Trichlorophenol	13.07	196	196953	21.08	ng/ul	95
39) 2,4,5-Trichlorophenol	13.15	196	209584	21.70	ng/ul	94
40) 1,1'-Biphenyl	13.46	154	850459	19.64	ng/ul	97
41) 2-Chloronaphthalene	13.50	162	632246	19.42	ng/ul	94
44) Dimethylphthalate	14.07	163	786604	19.38	ng/ul	99
45) 2,6-Dinitrotoluene	14.19	165	175154	20.45	ng/ul#	82
47) Acenaphthylene	14.35	152	1060313	19.99	ng/ul	99
48) 3-Nitroaniline	14.53	138	191174	21.21	ng/ul#	72
49) Acenaphthene	14.69	153	715859	19.40	ng/ul	96
52) 4-Nitrophenol	14.85	109	85786	19.57	ng/ul#	32
53) Dibenzofuran	15.01	168	975338	19.86	ng/ul	94
54) 2,4-Dinitrotoluene	14.98	165	252160	20.57	ng/ul#	76
55) 2,3,4,6-Tetrachlorophenol	15.25	232	161839	21.33	ng/ul#	83
56) Diethylphthalate	15.43	149	839690	19.77	ng/ul	100
58) Fluorene	15.66	166	794548	19.82	ng/ul	99
59) 4-Chlorophenyl-phenylether	15.65	204	339743	19.44	ng/ul	96
60) 4-Nitroaniline	15.68	138	172722	17.38	ng/ul#	49
63) 4,6-Dinitro-2-methylphenol	15.74	198	128154	20.08	ng/ul#	95
64) N-Nitrosodiphenylamine	15.87	169	678695	20.46	ng/ul	95
65) 4-Bromophenyl-phenylether	16.54	248	205865	20.44	ng/ul#	85
66) Hexachlorobenzene	16.67	284	214740	19.80	ng/ul#	89
67) Atrazine	16.81	200	215034	20.74	ng/ul	93
68) Pentachlorophenol	17.02	266	104816	21.16	ng/ul	93
69) Phenanthrene	17.40	178	1131784	19.91	ng/ul	99
71) Anthracene	17.49	178	1176973	20.18	ng/ul	100
72) Carbazole	17.76	167	1058748	21.47	ng/ul	97
73) Di-n-butylphthalate	18.31	149	1348331	21.01	ng/ul#	97
74) Fluoranthene	19.40	202	1145597	21.70	ng/ul#	82
77) Pyrene	19.76	202	1180889	19.35	ng/ul#	78
78) Butylbenzylphthalate	20.64	149	612736	21.61	ng/ul#	82
79) 3,3'-Dichlorobenzidine	21.50	252	374196	21.39	ng/ul#	93
80) Benzo(a)anthracene	21.58	228	1054846	19.83	ng/ul	99
81) Bis(2-ethylhexyl)phthalate	21.48	149	883153	21.83	ng/ul#	98
82) Chrysene	21.64	228	975726	19.74	ng/ul	97
84) Di-n-octyl phthalate	22.66	149	1519649	23.44	ng/ul	100
85) Benzo(b)fluoranthene	23.76	252	1064460	19.75	ng/ul#	94
86) Benzo(k)fluoranthene	23.82	252	1059087	20.00	ng/ul#	93
88) Benzo(a)pyrene	24.62	252	1044453	19.69	ng/ul#	92
89) Indeno(1,2,3-cd)pyrene	28.36	276	1268113	20.17	ng/ul#	81
90) Dibenzo(a,h)anthracene	28.41	278	1063194	19.94	ng/ul#	87
91) Benzo(g,h,i)perylene	29.49	276	1045342	20.05	ng/ul#	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

ALS Vial : 2 Sample Multiplier: 1

SSTD02062

Response via : Initial Calibration





#6

Phenol

Concen: 21.86 ng/ul

RT: 7.21 min Scan# 702

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion: 94 Resp: 352692

Ion Ratio Lower Upper


94 100

65 25.6 26.8 40.2#



66 30.4 44.4 66.6#



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



Time--> 7.10 7.20 7.30



#8

Bis(2-Chloroethyl)ether

Concen: 19.98 ng/ul

RT: 7.42 min Scan# 738

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion: 93 Resp: 248881

Ion Ratio Lower Upper


93 100

63 77.5 75.5 113.3


95 33.5 29.2 43.8




Abundance Ion 93.00 (92.70 to 93.70): BGC
Ion 63.00 (62.70 to 63.70): BGC
Ion 95.00 (94.70 to 95.70): BGC



Time--> 7.35 7.40 7.45 7.50









#10
 2-Chlorophenol
 Concen: 20.39 ng/ul
 RT: 7.58 min Scan# 765
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062

Tgt Ion	Ratio	Lower	Upper
128	100		
64	44.5	56.3	84.5#
130	32.1	26.7	40.1




Abundance Ion 128.00 (127.70 to 128.70): E
 Ion 64.00 (63.70 to 64.70): BG
 Ion 130.00 (129.70 to 130.70): E






#11
 2-Methylphenol
 Concen: 22.02 ng/ul
 RT: 8.46 min Scan# 915
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion	Ratio	Lower	Upper
108	100		
107	87.8	76.7	115.1



Abundance Ion 108.00 (107.70 to 108.70): E
 Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxybis(1-Chloropropane)
Concen: 21.17 ng/ul
RT: 8.54 min Scan# 928
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion: 45 Resp: 275128
Ion Ratio Lower Upper
45 100
77 16.9 10.0 15.0#
79 13.6 7.3 10.9#



#14
Acetophenone
Concen: 21.16 ng/ul
RT: 8.83 min Scan# 977
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion: 105 Resp: 410704
Ion Ratio Lower Upper
105 100
77 70.9 76.0 114.0#
51 21.3 34.8 52.2#





#15

N-Nitroso-di-n-propylamine

Concen: 20.14 ng/ul

RT: 8.81 min Scan# 974

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062

Tgt Ion: 70 Resp: 192792


Ion Ratio Lower Upper

70 100

42 43.0 59.0 88.6#

101 10.8 6.6 9.8#

130 27.3 15.6 23.4#






Abundance Ion 70.00 (69.70 to 70.70): BG0

Ion 42.00 (41.70 to 42.70): BG0

Ion 101.00 (100.70 to 101.70): E

Ion 130.00 (129.70 to 130.70): E



#16

4-Methylphenol

Concen: 22.17 ng/ul

RT: 8.79 min Scan# 971

Delta R.T. 0.00 min

Lab File: BG026758.D


Acq: 28 Apr 2017 22:18

Tgt Ion: 108 Resp: 309651

Ion Ratio Lower Upper



108 100


107 110.0 91.7 137.5



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E





#17

Hexachloroethane

Concen: 19.86 ng/ul

RT: 9.09 min Scan# 1022

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion: 117 Resp: 99325

Ion Ratio Lower Upper

117 100

201 84.8 91.9 137.9#




199 55.2 64.3 96.5#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 19.93 ng/ul

RT: 9.21 min Scan# 1042

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion: 77 Resp: 284323

Ion Ratio Lower Upper

77 100

123 53.7 27.4 41.0#



65 13.4 13.3 19.9




Abundance Ion 77.00 (76.70 to 77.70): BGC

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BGC









#21
Isophorone
Concen: 20.00 ng/ul
RT: 9.73 min Scan# 1130
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion: 82 Resp: 547874
Ion Ratio Lower Upper
82 100
95 6.5 7.8 11.6#
138 19.6 12.2 18.2#



#23
2-Nitrophenol
Concen: 20.51 ng/ul
RT: 9.92 min Scan# 1163
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion: 139 Resp: 173445
Ion Ratio Lower Upper
139 100
65 46.6 67.0 100.6#
109 22.7 39.6 59.4#









#24
 2,4-Dimethylphenol
 Concen: 21.05 ng/ul
 RT: 9.99 min Scan# 1174
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion: 107 Resp: 324446
 Ion Ratio Lower Upper
 107 100
 121 56.4 32.3 48.5#
 122 90.5 61.2 91.8



#25
 Bis(2-Chloroethoxy)methane
 Concen: 19.67 ng/ul
 RT: 10.21 min Scan# 1212
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion: 93 Resp: 368079
 Ion Ratio Lower Upper
 93 100
 95 33.1 23.4 35.0
 123 12.0 7.8 11.6#









#27
2,4-Dichlorophenol
Concen: 21.20 ng/ul
RT: 10.47 min Scan# 1256
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Instrument :
BNA_G
ClientSampleId :
SSTD02062

Tgt Ion	Ratio	Lower	Upper
162	100		
164	66.1	50.8	76.2
98	34.4	32.1	48.1




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): B






#28
Naphthalene
Concen: 19.73 ng/ul
RT: 10.86 min Scan# 1323
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.4	14.2
127	14.7	10.3	15.5



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E









#30
4-Chloroaniline
Concen: 22.72 ng/ul
RT: 10.97 min Scan# 1341
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:127 Resp: 366561
Ion Ratio Lower Upper
127 100
129 33.2 28.9 43.3



#31
Hexachlorobutadiene
Concen: 18.94 ng/ul
RT: 11.14 min Scan# 1371
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:225 Resp: 122119
Ion Ratio Lower Upper
225 100
223 65.8 45.9 68.9
227 68.7 44.7 67.1#









#32
 Caprolactam
 Concen: 22.06 ng/ul
 RT: 11.71 min Scan# 1468
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion:113 Resp: 105031
 Ion Ratio Lower Upper
 113 100
 55 145.1 168.6 252.8#
 56 102.6 128.5 192.7#



#33
 4-Chloro-3-methylphenol
 Concen: 22.80 ng/ul
 RT: 12.10 min Scan# 1534
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:107 Resp: 315036
 Ion Ratio Lower Upper
 107 100
 144 26.2 18.2 27.4
 142 87.4 55.0 82.4#





#34

2-Methylnaphthalene

Concen: 20.42 ng/ul

RT: 12.46 min Scan# 1595

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :


SSTD02062

Tgt Ion:142 Resp: 697922

Ion Ratio Lower Upper

142 100

141 89.6 68.6 102.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

12.46

400000



300000

200000

100000

0

Time-->



#36

1,2,4,5-Tetrachlorobenzene

Concen: 19.34 ng/ul

RT: 12.82 min Scan# 1657

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Tgt Ion:216 Resp: 262727


Ion Ratio Lower Upper

216 100

214 77.9 67.0 100.6

179 24.5 18.1 27.1

108 22.4 14.1 21.1#



Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E

12.82

200000


150000


100000

50000

0

Time-->









#37
Hexachlorocyclopentadiene
Concen: 19.56 ng/ul
RT: 12.81 min Scan# 1654
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion: 237 Resp: 117021
Ion Ratio Lower Upper
237 100
235 66.3 50.2 75.4
272 13.1 10.6 16.0



#38
2,4,6-Trichlorophenol
Concen: 21.08 ng/ul
RT: 13.07 min Scan# 1699
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion: 196 Resp: 196953
Ion Ratio Lower Upper
196 100
198 95.0 71.4 107.2
200 30.3 24.6 37.0









#39
 2,4,5-Trichlorophenol
 Concen: 21.70 ng/ul
 RT: 13.15 min Scan# 1712
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion:196 Resp: 209584
 Ion Ratio Lower Upper
 196 100
 198 92.8 78.6 118.0
 200 29.7 27.3 40.9



#40
 1,1'-Biphenyl
 Concen: 19.64 ng/ul
 RT: 13.46 min Scan# 1765
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:154 Resp: 850459
 Ion Ratio Lower Upper
 154 100
 153 42.4 32.2 48.2
 76 13.1 9.6 14.4









#41
 2-Chloronaphthalene
 Concen: 19.42 ng/ul
 RT: 13.50 min Scan# 1773
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion	Ratio	Lower	Upper
162	100		
164	35.1	25.0	37.6
127	41.5	30.7	46.1



#44
 Dimethylphthalate
 Concen: 19.38 ng/ul
 RT: 14.07 min Scan# 1869
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.9	3.4	5.0
164	10.9	9.0	13.4









#45
 2,6-Dinitrotoluene
 Concen: 20.45 ng/ul
 RT: 14.19 min Scan# 1890
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion:165 Resp: 175154
 Ion Ratio Lower Upper
 165 100
 89 49.4 52.9 79.3#
 121 22.6 22.2 33.4



#47
 Acenaphthylene
 Concen: 19.99 ng/ul
 RT: 14.35 min Scan# 1916
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:152 Resp: 1060313
 Ion Ratio Lower Upper
 152 100
 151 21.6 16.7 25.1
 153 14.7 11.4 17.2









#48
 3-Nitroaniline
 Concen: 21.21 ng/ul
 RT: 14.53 min Scan# 1947
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion:138 Resp: 191174
 Ion Ratio Lower Upper
 138 100
 108 8.4 6.3 9.5
 92 98.1 106.4 159.6#



#49
 Acenaphthene
 Concen: 19.40 ng/ul
 RT: 14.69 min Scan# 1974
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:153 Resp: 715859
 Ion Ratio Lower Upper
 153 100
 152 49.1 36.5 54.7
 154 87.3 72.3 108.5





#52

4-Nitrophenol

Concen: 19.57 ng/ul

RT: 14.85 min Scan# 2002

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:109 Resp: 85786

Ion Ratio Lower Upper

109 100

139 184.9 62.6 93.8#




65 145.9 90.4 135.6#



Abundance Ion 109.00 (108.70 to 109.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 65.00 (64.70 to 65.70): BGC



#53

Dibenzofuran

Concen: 19.86 ng/ul

RT: 15.01 min Scan# 2030

Delta R.T. -0.00 min

Lab File: BG026758.D


Acq: 28 Apr 2017 22:18

Tgt Ion:168 Resp: 975338

Ion Ratio Lower Upper



168 100


139 38.7 33.8 50.8



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E









#54
2,4-Dinitrotoluene
Concen: 20.57 ng/ul
RT: 14.98 min Scan# 2024
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Instrument :
BNA_G
ClientSampleId :
SSTD02062

Tgt Ion:165 Resp: 252160
Ion Ratio Lower Upper
165 100
63 40.2 46.8 70.2#
182 2.5 1.8 2.6




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BG
Ion 182.00 (181.70 to 182.70): E






#55
2,3,4,6-Tetrachlorophenol
Concen: 21.33 ng/ul
RT: 15.25 min Scan# 2070
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:232 Resp: 161839
Ion Ratio Lower Upper
232 100
131 53.4 33.3 49.9#
130 3.0 1.8 2.6#
166 35.3 21.9 32.9#



Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E









#56
Diethylphthalate
Concen: 19.77 ng/ul
RT: 15.43 min Scan# 2100
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:149 Resp: 839690
Ion Ratio Lower Upper
149 100
177 22.1 17.8 26.8
150 12.8 10.2 15.4



#58
Fluorene
Concen: 19.82 ng/ul
RT: 15.66 min Scan# 2140
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:166 Resp: 794548
Ion Ratio Lower Upper
166 100
165 98.5 79.7 119.5
167 14.1 11.4 17.2









#59
4-Chlorophenyl-phenylether
Concen: 19.44 ng/ul
RT: 15.65 min Scan# 2138
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:204 Resp: 339743
Ion Ratio Lower Upper
204 100
206 34.8 32.3 48.5
141 71.0 57.5 86.3



#60
4-Nitroaniline
Concen: 17.38 ng/ul
RT: 15.68 min Scan# 2144
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:138 Resp: 172722
Ion Ratio Lower Upper
138 100
92 46.1 56.9 85.3#
108 57.4 104.9 157.3#









#63
 4,6-Dinitro-2-methylphenol
 Concen: 20.08 ng/ul
 RT: 15.74 min Scan# 2154
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062

Tgt Ion:198 Resp: 128154
 Ion Ratio Lower Upper
 198 100
 182 4.1 5.8 6.4#
 77 25.4 22.1 33.9




Abundance Ion 198.00 (197.70 to 198.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 77.00 (76.70 to 77.70): B






#64
 N-Nitrosodiphenylamine
 Concen: 20.46 ng/ul
 RT: 15.87 min Scan# 2175
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:169 Resp: 678695
 Ion Ratio Lower Upper
 169 100
 168 69.4 59.1 88.7
 167 39.4 29.7 44.5



Abundance Ion 169.00 (168.70 to 169.70): E
 Ion 168.00 (167.70 to 168.70): E
 Ion 167.00 (166.70 to 167.70): E









#65
4-Bromophenyl-phenylether
Concen: 20.44 ng/ul
RT: 16.54 min Scan# 2290
Delta R.T. -0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:248 Resp: 205865
Ion Ratio Lower Upper
248 100
250 95.5 71.9 107.9
141 89.5 53.8 80.8#



#66
Hexachlorobenzene
Concen: 19.80 ng/ul
RT: 16.67 min Scan# 2312
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:284 Resp: 214740
Ion Ratio Lower Upper
284 100
142 44.5 26.1 39.1#
249 32.1 25.9 38.9









#67
 Atrazine
 Concen: 20.74 ng/ul
 RT: 16.81 min Scan# 2335
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SST02062


Tgt Ion:200 Resp: 215034
 Ion Ratio Lower Upper
 200 100
 173 27.7 19.2 28.8
 215 45.4 40.3 60.5



#68
 Pentachlorophenol
 Concen: 21.16 ng/ul
 RT: 17.02 min Scan# 2371
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:266 Resp: 104816
 Ion Ratio Lower Upper
 266 100
 264 57.6 47.1 70.7
 268 61.4 56.2 84.4





#69

Phenanthrene

Concen: 19.91 ng/ul

RT: 17.40 min Scan# 2436

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:178 Resp: 1131784

Ion Ratio Lower Upper




178 100

179 15.8 12.4 18.6

176 20.1 16.5 24.7



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E



#71

Anthracene

Concen: 20.18 ng/ul

RT: 17.49 min Scan# 2451

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion:178 Resp: 1176973

Ion Ratio Lower Upper



178 100


179 15.9 12.6 19.0

176 20.0 15.9 23.9



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E





#72

Carbazole

Concen: 21.47 ng/ul

RT: 17.76 min Scan# 2497

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:167 Resp: 1058748

Ion Ratio Lower Upper

167 100

166 24.0 17.9 26.9




139 14.1 10.5 15.7



Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E



#73

Di-n-butylphthalate

Concen: 21.01 ng/ul

RT: 18.31 min Scan# 2590

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion:149 Resp: 1348331

Ion Ratio Lower Upper

149 100

150 10.0 7.9 11.9



104 4.6 5.8 8.8#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E









#74
 Fluoranthene
 Concen: 21.70 ng/ul
 RT: 19.40 min Scan# 2777
 Delta R.T. 0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18




Instrument :
 BNA_G
 ClientSampleId :
 SSTD02062


Tgt Ion:202 Resp: 1145597
 Ion Ratio Lower Upper
 202 100
 101 14.6 6.2 9.2#
 100 11.7 5.2 7.8#



#77
 Pyrene
 Concen: 19.35 ng/ul
 RT: 19.76 min Scan# 2838
 Delta R.T. -0.00 min
 Lab File: BG026758.D
 Acq: 28 Apr 2017 22:18

Tgt Ion:202 Resp: 1180889
 Ion Ratio Lower Upper
 202 100
 101 17.7 6.9 10.3#
 100 15.0 6.6 10.0#









#78
Butylbenzylphthalate
Concen: 21.61 ng/ul
RT: 20.64 min Scan# 2987
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:149 Resp: 612736
Ion Ratio Lower Upper
149 100
91 64.5 65.8 98.6#
206 18.8 20.1 30.1#



#79
3,3'-Dichlorobenzidine
Concen: 21.39 ng/ul
RT: 21.50 min Scan# 3133
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:252 Resp: 374196
Ion Ratio Lower Upper
252 100
254 65.1 49.0 73.6
126 14.6 5.1 7.7#





#80

Benzo(a)anthracene

Concen: 19.83 ng/ul

RT: 21.58 min Scan# 3148

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:228 Resp: 1054846

Ion Ratio Lower Upper




228 100

229 20.6 16.3 24.5

226 28.2 21.9 32.9



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 229.00 (228.70 to 229.70): E
Ion 226.00 (225.70 to 226.70): E



#81

Bis(2-ethylhexyl)phthalate

Concen: 21.83 ng/ul

RT: 21.48 min Scan# 3130

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion:149 Resp: 883153

Ion Ratio Lower Upper



149 100


167 27.8 21.8 32.8

279 3.6 4.5 6.7#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E





#82

Chrysene

Concen: 19.74 ng/ul

RT: 21.64 min Scan# 3158

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:228 Resp: 975726

Ion Ratio Lower Upper

228 100

226 32.0 24.0 36.0




229 20.7 16.9 25.3



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Di-n-octyl phthalate

Concen: 23.44 ng/ul


RT: 22.66 min Scan# 3331

Delta R.T. -0.00 min



Lab File: BG026758.D


Acq: 28 Apr 2017 22:18

Tgt Ion:149 Resp: 1519649



Abundance Ion 149.00 (148.70 to 149.70): E









#85
Benzo(b)fluoranthene
Concen: 19.75 ng/ul
RT: 23.76 min Scan# 3518
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18




Instrument :
BNA_G
ClientSampleId :
SSTD02062


Tgt Ion:252 Resp: 1064460
Ion Ratio Lower Upper
252 100
253 22.6 17.7 26.5
125 15.1 4.0 6.0#



#86
Benzo(k)fluoranthene
Concen: 20.00 ng/ul
RT: 23.82 min Scan# 3529
Delta R.T. 0.00 min
Lab File: BG026758.D
Acq: 28 Apr 2017 22:18

Tgt Ion:252 Resp: 1059087
Ion Ratio Lower Upper
252 100
253 22.5 18.5 27.7
125 16.4 4.1 6.1#





#88

Benzo(a)pyrene

Concen: 19.69 ng/ul

RT: 24.62 min Scan# 3664

Delta R.T. 0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:252 Resp: 1044453

Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.6




125 16.5 5.4 8.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Indeno(1,2,3-cd)pyrene

Concen: 20.17 ng/ul

RT: 28.36 min Scan# 4301

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion:276 Resp: 1268113

Ion Ratio Lower Upper

276 100

138 30.8 8.2 12.4#



277 24.7 18.6 28.0




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#90

Dibenzo(a,h)anthracene

Concen: 19.94 ng/ul

RT: 28.41 min Scan# 4309

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18

Instrument :

BNA_G

ClientSampleId :

SSTD02062


Tgt Ion:278 Resp: 1063194

Ion Ratio Lower Upper

278 100

139 21.9 6.7 10.1#




279 23.4 20.9 31.3



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 20.05 ng/ul

RT: 29.49 min Scan# 4493

Delta R.T. -0.00 min

Lab File: BG026758.D

Acq: 28 Apr 2017 22:18


Tgt Ion:276 Resp: 1045342

Ion Ratio Lower Upper

276 100

138 31.1 8.6 12.8#

277 23.8 17.6 26.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

