

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042817\
 Data File : BG026762.D
 Acq On : 29 Apr 2017 00:54
 Operator : SJ/MA
 Sample : I2816-17
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 JHT89

Quant Time: Apr 29 01:27:19 2017
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG042717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Apr 29 01:26:17 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	126568	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	646390	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	373453	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.35	188	727285	20.00	ng/ul	0.00
75) Chrysene-d12	21.60	240	656404	20.00	ng/ul	0.00
83) Perylene-d12	24.76	264	669745	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.43	96	18656	6.64	ng/uL	0.00
5) Phenol-d5	7.19	99	403277	35.70	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.33	67	224610	33.81	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	327251	34.13	ng/ul	0.00
13) 4-Methylphenol-d8	8.73	113	342488	36.53	ng/ul	0.00
19) Nitrobenzene-d5	9.17	128	177545	32.86	ng/ul	0.00
22) 2-Nitrophenol-d4	9.89	143	203145	34.57	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.44	165	327533	35.08	ng/ul	0.00
29) 4-Chloroaniline-d4	10.94	131	491914	39.36	ng/ul	0.00
43) Dimethylphthalate-d6	14.02	166	1026549	34.15	ng/ul	0.00
46) Acenaphthylene-d8	14.32	160	1318238	34.91	ng/ul	0.00
51) 4-Nitrophenol-d4	14.85	143	148171	27.35	ng/ul	0.00
57) Fluorene-d10	15.61	176	887372	33.82	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.73	200	121185	27.44	ng/ul	0.00
70) Anthracene-d10	17.45	188	1232001	34.90	ng/ul	0.00
76) Pyrene-d10	19.73	212	1193335	34.44	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.54	264	1184011	37.39	ng/ul	0.00

Target Compounds


						Ovalue
6) Phenol	7.21	94	14466	1.26	ng/ul	93
44) Dimethylphthalate	14.07	163	174424	5.93	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG042817\
Data File : BG026762.D
Acq On : 29 Apr 2017 00:54
Operator : SJ/MA
Sample : I2816-17
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
JHT89

Quant Time: Apr 29 01:27:19 2017
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Apr 29 01:26:17 2017
Response via : Initial Calibration





#6

Phenol

Concen: 1.26 ng/ul

RT: 7.21 min Scan# 702

Delta R.T. 0.00 min

Lab File: BG026762.D

Acq: 29 Apr 2017 00:54

Instrument :

BNA_G

ClientSampleId :

JHT89





Tot Ion: 94 Resp: 14466

Ion Ratio Lower Upper

94 100

65 28.0 26.8 40.2

66 51.6 44.4 66.6



#44

Dimethylphthalate

Concen: 5.93 ng/ul

RT: 14.07 min Scan# 1869

Delta R.T. 0.00 min

Lab File: BG026762.D

Acq: 29 Apr 2017 00:54

Tgt Ion:163 Resp: 174424

Ion Ratio Lower Upper

163 100

194 3.7 3.4 5.0

164 10.9 9.0 13.4

