

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042823\
 Data File : BG057219.D
 Acq On : 28 Apr 2023 15:24
 Operator : CG/JU
 Sample : PB152400BL
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SBLK400

Quant Time: Apr 29 00:53:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG042823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 23:37:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.379	152	13737	20.000	ng/ul	0.00
20) Naphthalene-d8	11.210	136	59144	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.987	164	47681	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.725	188	129797	20.000	ng/ul	0.00
79) Chrysene-d12	22.037	240	119636	20.000	ng/ul	0.00
88) Perylene-d12	25.544	264	120819	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.656	96	1699	5.443	ng/uL	0.00
4) Pyridine-d5	4.096	84	27075	27.511	ng/ul	0.00
7) Phenol-d5	7.509	99	37216	28.647	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.680	67	24127	31.576	ng/ul	0.00
11) 2-Chlorophenol-d4	7.903	132	27051	31.598	ng/ul	0.00
15) 4-Methylphenol-d8	9.072	113	29376	29.751	ng/ul	0.00
21) Nitrobenzene-d5	9.553	128	14991	31.828	ng/ul	0.00
24) 2-Nitrophenol-d4	10.282	143	18221	33.148	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.828	165	30466	28.728	ng/ul	0.00
31) 4-Chloroaniline-d4	11.339	131	40782	29.015	ng/ul	0.00
46) Dimethylphthalate-d6	14.376	166	135619	35.896	ng/ul	0.00
49) Acenaphthylene-d8	14.688	160	139153	32.770	ng/ul	0.00
54) 4-Nitrophenol-d4	15.169	143	16119	29.310	ng/ul	0.00
60) Fluorene-d10	15.974	176	115846	32.915	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.086	200	23507	30.567	ng/ul	0.00
73) Anthracene-d10	17.825	188	201661	34.038	ng/ul	0.00
81) Pyrene-d10	20.092	212	247754	34.889	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.303	264	215053	34.937	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042823\
Data File : BG057219.D
Acq On : 28 Apr 2023 15:24
Operator : CG/JU
Sample : PB152400BL
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SBLK400

Quant Time: Apr 29 00:53:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG042823.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Apr 28 23:37:02 2023
Response via : Initial Calibration

