

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042823\
 Data File : BG057224.D
 Acq On : 28 Apr 2023 19:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465

Quant Time: Apr 29 00:54:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG042823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 23:37:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.378	152	14225	20.000	ng/ul	0.00
20) Naphthalene-d8	11.215	136	60856	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.992	164	49842	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.730	188	128814	20.000	ng/ul	0.00
79) Chrysene-d12	22.036	240	117189	20.000	ng/ul	0.00
88) Perylene-d12	25.543	264	123009	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.655	96	2627	8.127	ng/uL	0.00
4) Pyridine-d5	4.095	84	21613	21.208	ng/ul	0.00
7) Phenol-d5	7.514	99	27593	20.511	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.679	67	16278	20.573	ng/ul	0.00
11) 2-Chlorophenol-d4	7.902	132	18686	21.078	ng/ul	0.00
15) 4-Methylphenol-d8	9.077	113	21998	21.515	ng/ul	0.00
21) Nitrobenzene-d5	9.553	128	10541	21.750	ng/ul	0.00
24) 2-Nitrophenol-d4	10.281	143	11908	21.054	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.827	165	24133	22.116	ng/ul	0.00
31) 4-Chloroaniline-d4	11.344	131	33431	23.116	ng/ul	0.00
46) Dimethylphthalate-d6	14.376	166	85176	21.567	ng/ul	0.00
49) Acenaphthylene-d8	14.687	160	94466	21.282	ng/ul	0.00
54) 4-Nitrophenol-d4	15.169	143	11901	20.702	ng/ul	0.00
60) Fluorene-d10	15.973	176	79832	21.699	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	16.085	200	15628	20.477	ng/ul	0.00
73) Anthracene-d10	17.824	188	128467	21.849	ng/ul	0.00
81) Pyrene-d10	20.091	212	152199	21.880	ng/ul	0.00
92) Benzo(a)pyrene-d12	25.296	264	132879	21.203	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.696	88	3135	9.145	ng/uL#	92
5) Pyridine	4.119	79	22937	21.351	ng/ul	91
6) Benzaldehyde	7.503	77	8079	19.663	ng/ul	91
8) Phenol	7.544	94	27746	20.463	ng/ul	93
10) Bis(2-Chloroethyl)ether	7.779	93	21052	20.745	ng/ul	94
12) 2-Chlorophenol	7.937	128	19961	21.436	ng/ul	98
13) 2-Methylphenol	8.807	108	20951	19.938	ng/ul#	79
14) 2,2'-oxybis(1-Chloropr...	8.895	45	26974	20.616	ng/ul#	97
16) Acetophenone	9.206	105	38021	21.610	ng/ul	94
17) N-Nitroso-di-n-propyla...	9.177	70	21561	21.326	ng/ul#	91
18) 4-Methylphenol	9.141	108	23351	20.530	ng/ul	95
19) Hexachloroethane	9.476	117	7855	20.162	ng/ul#	88
22) Nitrobenzene	9.594	77	31439	21.785	ng/ul#	96
23) Isophorone	10.117	82	60760	22.129	ng/ul	98
25) 2-Nitrophenol	10.316	139	12756	22.267	ng/ul#	89
26) 2,4-Dimethylphenol	10.357	107	28319	21.791	ng/ul	95
27) Bis(2-Chloroethoxy)met...	10.592	93	30990	21.713	ng/ul	96
29) 2,4-Dichlorophenol	10.851	162	23325	22.520	ng/ul	99
30) Naphthalene	11.268	128	72159	21.596	ng/ul	99
32) 4-Chloroaniline	11.368	127	34174	22.678	ng/ul	95
33) Hexachlorobutadiene	11.538	225	19566	22.273	ng/ul	97
34) Caprolactam	12.108	113	3895	11.171	ng/ul	94
35) 4-Chloro-3-methylphenol	12.461	107	26062	21.869	ng/ul	97
36) 2-Methylnaphthalene	12.842	142	52365	21.370	ng/ul	92


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042823\
 Data File : BG057224.D
 Acq On : 28 Apr 2023 19:29
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465

Quant Time: Apr 29 00:54:15 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\SFAM-EPA-BG042823.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Apr 28 23:37:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	13.060	142	53930	21.888	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	13.207	216	38674	21.352	ng/ul	95
40) Hexachlorocyclopentadiene	13.177	237	16056	17.764	ng/ul#	93
41) 2,4,6-Trichlorophenol	13.436	196	22619	21.442	ng/ul	94
42) 2,4,5-Trichlorophenol	13.512	196	24517	21.647	ng/ul	91
43) 1,1'-Biphenyl	13.829	154	80031	21.043	ng/ul	97
44) 2-Chloronaphthalene	13.876	162	62683	21.257	ng/ul	95
45) 2-Nitroaniline	14.070	65	20010	22.478	ng/ul	98
47) Dimethylphthalate	14.423	163	86161	21.725	ng/ul	99
48) 2,6-Dinitrotoluene	14.552	165	17807	21.817	ng/ul	97
50) Acenaphthylene	14.716	152	95783	21.235	ng/ul	99
51) 3-Nitroaniline	14.887	138	11588	20.038	ng/ul	85
52) Acenaphthene	15.051	153	67078	20.872	ng/ul	99
53) 2,4-Dinitrophenol	15.098	184	8757	20.147	ng/ul	95
55) 4-Nitrophenol	15.186	109	14069	21.589	ng/ul	97
56) Dibenzofuran	15.386	168	99128	21.359	ng/ul	98
57) 2,4-Dinitrotoluene	15.339	165	26613	22.777	ng/ul#	97
58) 2,3,4,6-Tetrachlorophenol	15.609	232	22794	22.149	ng/ul#	97
59) Diethylphthalate	15.768	149	86980	21.587	ng/ul	98
61) Fluorene	16.032	166	83944	21.390	ng/ul	97
62) 4-Chlorophenyl-phenyle...	16.009	204	50314	21.966	ng/ul	98
63) 4-Nitroaniline	16.038	138	10276	18.764	ng/ul	88
66) 4,6-Dinitro-2-methylph...	16.103	198	16366	21.044	ng/ul#	93
67) N-Nitrosodiphenylamine	16.220	169	71425	21.227	ng/ul	97
68) 4-Bromophenyl-phenylether	16.902	248	33109	21.505	ng/ul	93
69) Hexachlorobenzene	17.031	284	35144	21.677	ng/ul	96
70) Atrazine	17.154	200	32452	21.700	ng/ul	99
71) Pentachlorophenol	17.377	266	13848	18.479	ng/ul	97
72) Phenanthrene	17.771	178	143802	21.437	ng/ul	99
74) Anthracene	17.859	178	146553	21.750	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.800	216	41154	21.241	ng/ul	95
76) Pentachlorobenzene	15.304	250	42483	21.542	ng/ul	99
77) Carbazole	18.123	167	124059	22.031	ng/ul	99
78) Di-n-butylphthalate	18.646	149	142367	22.109	ng/ul	99
80) Fluoranthene	19.757	202	178845	21.644	ng/ul	99
82) Pyrene	20.121	202	180759	21.609	ng/ul	97
83) Butylbenzylphthalate	20.973	149	57409	21.431	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.912	252	58573	22.505	ng/ul#	97
85) Benzo(a)anthracene	22.018	228	179252	21.627	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.865	149	83519	21.437	ng/ul	96
87) Chrysene	22.089	228	165936	21.220	ng/ul	98
89) Di-n-octyl phthalate	23.164	149	141322	20.805	ng/ul	100
90) Benzo(b)fluoranthene	24.421	252	183283	21.516	ng/ul	98
91) Benzo(k)fluoranthene	24.491	252	175513	21.179	ng/ul	100
93) Benzo(a)pyrene	25.378	252	156218	21.243	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	29.596	276	194862	21.197	ng/ul	100
95) Dibenzo(a,h)anthracene	29.649	278	162281	21.289	ng/ul#	95
96) Benzo(g,h,i)perylene	30.853	276	89498	12.030	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 9.145 ng/uL
RT: 3.696 min Scan# 21
Delta R.T. -0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 88 Resp: 3135
Ion Ratio Lower Upper
88 100
43 28.0 28.2 42.4#
58 74.0 63.3 94.9



#5
Pyridine
Concen: 21.351 ng/uL
RT: 4.119 min Scan# 99
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion: 79 Resp: 22937
Ion Ratio Lower Upper
79 100
52 68.6 62.4 93.6
51 38.6 29.0 43.4









#6
Benzaldehyde
Concen: 19.663 ng/ul
RT: 7.503 min Scan# 61
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 77 Resp: 8079
Ion Ratio Lower Upper
77 100
105 93.1 79.8 119.6
106 87.5 78.7 118.1



#8
Phenol
Concen: 20.463 ng/ul
RT: 7.544 min Scan# 682
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion: 94 Resp: 27746
Ion Ratio Lower Upper
94 100
65 37.8 25.5 38.3
66 49.4 36.9 55.3









#10
Bis(2-Chloroethyl)ether
Concen: 20.745 ng/ul
RT: 7.779 min Scan# 721
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 93 Resp: 21052
Ion Ratio Lower Upper
93 100
63 77.5 67.2 100.8
95 34.6 26.3 39.5



#12
2-Chlorophenol
Concen: 21.436 ng/ul
RT: 7.937 min Scan# 749
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:128 Resp: 19961
Ion Ratio Lower Upper
128 100
64 51.9 41.3 61.9
130 33.5 24.6 37.0









#13
2-Methylphenol
Concen: 19.938 ng/ul
RT: 8.807 min Scan# 897
Delta R.T. -0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:108 Resp: 20951
Ion Ratio Lower Upper
108 100
107 79.3 80.0 120.0#



#14
2,2'-oxybis(1-Chloropropane)
Concen: 20.616 ng/ul
RT: 8.895 min Scan# 912
Delta R.T. -0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion: 45 Resp: 26974
Ion Ratio Lower Upper
45 100
77 16.3 12.9 19.3
79 13.9 8.9 13.3#









#16
Acetophenone
Concen: 21.610 ng/ul
RT: 9.206 min Scan# 90
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:105 Resp: 38021
Ion Ratio Lower Upper
105 100
77 82.3 70.6 105.8
51 26.9 24.5 36.7



#17
N-Nitroso-di-n-propylamine
Concen: 21.326 ng/ul
RT: 9.177 min Scan# 960
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion: 70 Resp: 21561
Ion Ratio Lower Upper
70 100
42 49.9 45.0 67.4
101 7.4 6.5 9.7
130 14.2 15.9 23.9#









#18
4-Methylphenol
Concen: 20.530 ng/ul
RT: 9.141 min Scan# 911
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:108 Resp: 23351
Ion Ratio Lower Upper
108 100
107 112.0 85.1 127.7



#19
Hexachloroethane
Concen: 20.162 ng/ul
RT: 9.476 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:117 Resp: 7855
Ion Ratio Lower Upper
117 100
201 146.4 109.4 164.2
199 98.2 65.4 98.2#









#22
Nitrobenzene
Concen: 21.785 ng/ul
RT: 9.594 min Scan# 1031
Delta R.T. -0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion: 77 Resp: 31439
Ion Ratio Lower Upper
77 100
123 35.7 29.6 44.4
65 17.3 11.4 17.0#



#23
Isophorone
Concen: 22.129 ng/ul
RT: 10.117 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion: 82 Resp: 60760
Ion Ratio Lower Upper
82 100
95 10.0 6.8 10.2
138 15.5 12.2 18.2









#25
2-Nitrophenol
Concen: 22.267 ng/ul
RT: 10.316 min Scan# 1154
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:139 Resp: 12756
Ion Ratio Lower Upper
139 100
65 74.6 48.7 73.1#
109 42.5 35.2 52.8



#26
2,4-Dimethylphenol
Concen: 21.791 ng/ul
RT: 10.357 min Scan# 1161
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:107 Resp: 28319
Ion Ratio Lower Upper
107 100
121 44.3 35.4 53.0
122 78.5 57.0 85.4





#27

Bis(2-Chloroethoxy)methane

Concen: 21.713 ng/ul

RT: 10.592 min Scan# 1

Delta R.T. 0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion: 93 Resp: 30990

Ion Ratio Lower Upper

93 100

95 33.4 24.8 37.2

123 9.5 8.1 12.1



#29

2,4-Dichlorophenol

Concen: 22.520 ng/ul

RT: 10.851 min Scan# 1245

Delta R.T. -0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:162 Resp: 23325


Ion Ratio Lower Upper

162 100

164 62.4 50.5 75.7

98 34.4 27.5 41.3









#30
Naphthalene
Concen: 21.596 ng/ul
RT: 11.268 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:128 Resp: 72159
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 13.9 11.4 17.0



#32
4-Chloroaniline
Concen: 22.678 ng/ul
RT: 11.368 min Scan# 1333
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:127 Resp: 34174
Ion Ratio Lower Upper
127 100
129 28.9 25.4 38.0





#33

Hexachlorobutadiene

Concen: 22.273 ng/ul

RT: 11.538 min Scan# 11

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:225 Resp: 19566

Ion Ratio Lower Upper

225 100

223 59.5 47.5 71.3

227 65.3 48.4 72.6



#34

Caprolactam

Concen: 11.171 ng/ul

RT: 12.108 min Scan# 1459

Delta R.T. -0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:113 Resp: 3895


Ion Ratio Lower Upper

113 100

55 147.9 123.1 184.7

56 131.6 98.7 148.1









#35
 4-Chloro-3-methylphenol
 Concen: 21.869 ng/ul
 RT: 12.461 min Scan# 1519
 Delta R.T. 0.000 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465


Tgt Ion:	107	Resp:	26062
Ion Ratio	Lower	Upper	
107	100		
144	23.1	18.3	27.5
142	76.8	58.6	88.0



#36
 2-Methylnaphthalene
 Concen: 21.370 ng/ul
 RT: 12.842 min Scan# 1584
 Delta R.T. 0.000 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29

Tgt Ion:	142	Resp:	52365
Ion Ratio	Lower	Upper	
142	100		
141	92.8	68.7	103.1





#37

1-Methylnaphthalene

Concen: 21.888 ng/ul

RT: 13.060 min Scan# 1

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:142 Resp: 53930

Ion Ratio Lower Upper

142 100

141 90.9 72.9 109.3

116 4.6 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.352 ng/ul

RT: 13.207 min Scan# 1646

Delta R.T. 0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Tgt Ion:216 Resp: 38674




Ion Ratio Lower Upper


216 100

214 74.2 62.9 94.3

179 16.8 15.3 22.9

108 12.5 11.4 17.2









#40
Hexachlorocyclopentadiene
Concen: 17.764 ng/ul
RT: 13.177 min Scan# 1641
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:237 Resp: 16056
Ion Ratio Lower Upper
237 100
235 59.3 51.4 77.0
272 15.0 9.4 14.2#



#41
2,4,6-Trichlorophenol
Concen: 21.442 ng/ul
RT: 13.436 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:196 Resp: 22619
Ion Ratio Lower Upper
196 100
198 97.1 83.8 125.8
200 33.1 26.6 40.0









#42
2,4,5-Trichlorophenol
Concen: 21.647 ng/ul
RT: 13.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:196 Resp: 24517
Ion Ratio Lower Upper
196 100
198 98.4 85.7 128.5
200 31.5 29.9 44.9



#43
1,1'-Biphenyl
Concen: 21.043 ng/ul
RT: 13.829 min Scan# 1752
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:154 Resp: 80031
Ion Ratio Lower Upper
154 100
153 37.9 31.6 47.4
76 11.3 8.0 12.0









#44
2-Chloronaphthalene
Concen: 21.257 ng/ul
RT: 13.876 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:	162	Resp:	62683
Ion Ratio	Lower	Upper	
162	100		
164	36.2	26.2	39.4
127	39.2	29.4	44.2



#45
2-Nitroaniline
Concen: 22.478 ng/ul
RT: 14.070 min Scan# 1793
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:	65	Resp:	20010
Ion Ratio	Lower	Upper	
65	100		
92	66.4	51.8	77.6
138	88.6	71.9	107.9





#47

Dimethylphthalate

Concen: 21.725 ng/ul

RT: 14.423 min Scan# 1853

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:163 Resp: 86161

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.2

164 10.9 8.3 12.5



#48

2,6-Dinitrotoluene

Concen: 21.817 ng/ul

RT: 14.552 min Scan# 1875

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:165 Resp: 17807


Ion Ratio Lower Upper

165 100

89 65.7 50.6 76.0

121 20.1 16.8 25.2









#50
Acenaphthylene
Concen: 21.235 ng/ul
RT: 14.716 min Scan# 1903
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:152 Resp: 95783
Ion Ratio Lower Upper
152 100
151 19.8 16.3 24.5
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 20.038 ng/ul
RT: 14.887 min Scan# 1932
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:138 Resp: 11588
Ion Ratio Lower Upper
138 100
108 13.2 9.2 13.8
92 142.9 99.5 149.3









#52
Acenaphthene
Concen: 20.872 ng/ul
RT: 15.051 min Scan# 1960
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:153 Resp: 67078
Ion Ratio Lower Upper
153 100
152 46.3 36.5 54.7
154 89.0 70.6 106.0



#53
2,4-Dinitrophenol
Concen: 20.147 ng/ul
RT: 15.098 min Scan# 1968
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:184 Resp: 8757
Ion Ratio Lower Upper
184 100
63 65.2 48.3 72.5
154 59.2 50.0 75.0









#55
4-Nitrophenol
Concen: 21.589 ng/ul
RT: 15.186 min Scan# 1983
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:109 Resp: 14069
Ion Ratio Lower Upper
109 100
139 87.4 73.8 110.8
65 103.0 81.7 122.5



#56
Dibenzofuran
Concen: 21.359 ng/ul
RT: 15.386 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:168 Resp: 99128
Ion Ratio Lower Upper
168 100
139 37.9 29.6 44.4





#57

2,4-Dinitrotoluene

Concen: 22.777 ng/ul

RT: 15.339 min Scan# 2009

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:165 Resp: 26613

Ion Ratio Lower Upper

165 100

63 42.7 35.9 53.9

182 3.2 1.9 2.9#



#58

2,3,4,6-Tetrachlorophenol

Concen: 22.149 ng/ul

RT: 15.609 min Scan# 2055

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Tgt Ion:232 Resp: 22794




Ion Ratio Lower Upper


232 100

131 34.2 28.6 43.0

130 2.7 1.8 2.6#

166 29.5 22.2 33.2









#59
Diethylphthalate
Concen: 21.587 ng/ul
RT: 15.768 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:149 Resp: 86980
Ion Ratio Lower Upper
149 100
177 24.4 18.6 28.0
150 13.2 9.8 14.8



#61
Fluorene
Concen: 21.390 ng/ul
RT: 16.032 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:166 Resp: 83944
Ion Ratio Lower Upper
166 100
165 103.2 80.1 120.1
167 12.8 10.2 15.4





#62

4-Chlorophenyl-phenylether

Concen: 21.966 ng/ul

RT: 16.009 min Scan# 2123

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:204 Resp: 50314

Ion Ratio Lower Upper

204 100

206 31.8 26.5 39.7

141 51.0 41.5 62.3



#63

4-Nitroaniline

Concen: 18.764 ng/ul

RT: 16.038 min Scan# 2128

Delta R.T. -0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:138 Resp: 10276


Ion Ratio Lower Upper

138 100

92 49.8 39.9 59.9

108 92.6 89.1 133.7









#66
4,6-Dinitro-2-methylphenol
Concen: 21.044 ng/ul
RT: 16.103 min Scan# 2138
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:198 Resp: 16366
Ion Ratio Lower Upper
198 100
182 2.7 3.0 4.6#
77 21.5 20.1 30.1



#67
N-Nitrosodiphenylamine
Concen: 21.227 ng/ul
RT: 16.220 min Scan# 2159
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:169 Resp: 71425
Ion Ratio Lower Upper
169 100
168 70.9 55.4 83.2
167 36.5 31.0 46.4









#68
4-Bromophenyl-phenylether
Concen: 21.505 ng/ul
RT: 16.902 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:248 Resp: 33109
Ion Ratio Lower Upper
248 100
250 100.9 74.8 112.2
141 65.0 47.8 71.6



#69
Hexachlorobenzene
Concen: 21.677 ng/ul
RT: 17.031 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:284 Resp: 35144
Ion Ratio Lower Upper
284 100
142 31.8 22.6 33.8
249 33.9 26.2 39.2









#70
Atrazine
Concen: 21.700 ng/ul
RT: 17.154 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:200 Resp: 32452
Ion Ratio Lower Upper
200 100
173 21.6 17.8 26.6
215 49.3 40.3 60.5



#71
Pentachlorophenol
Concen: 18.479 ng/ul
RT: 17.377 min Scan# 2356
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:266 Resp: 13848
Ion Ratio Lower Upper
266 100
264 57.3 47.8 71.6
268 58.2 48.2 72.4









#72
Phenanthrene
Concen: 21.437 ng/ul
RT: 17.771 min Scan# 2423
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:178 Resp: 143802
Ion Ratio Lower Upper
178 100
179 15.6 11.8 17.8
176 19.7 15.4 23.0



#74
Anthracene
Concen: 21.750 ng/ul
RT: 17.859 min Scan# 2438
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:178 Resp: 146553
Ion Ratio Lower Upper
178 100
179 15.3 12.5 18.7
176 19.2 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 21.241 ng/uL
RT: 13.800 min Scan# 11
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:216 Resp: 41154
Ion Ratio Lower Upper
216 100
214 74.5 64.3 96.5
218 46.7 38.5 57.7



#76
Pentachlorobenzene
Concen: 21.542 ng/uL
RT: 15.304 min Scan# 2003
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:250 Resp: 42483
Ion Ratio Lower Upper
250 100
248 65.0 52.2 78.2
252 66.6 52.5 78.7









#77
 Carbazole
 Concen: 22.031 ng/ul
 RT: 18.123 min Scan# 2483
 Delta R.T. 0.000 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29




Instrument :
 BNA_G
 ClientSampleId :
 SSTD020465


Tgt Ion:167 Resp: 124059
 Ion Ratio Lower Upper
 167 100
 166 23.1 18.0 27.0
 139 12.2 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 22.109 ng/ul
 RT: 18.646 min Scan# 2572
 Delta R.T. 0.000 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29

Tgt Ion:149 Resp: 142367
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 5.9 5.0 7.6









#80
Fluoranthene
Concen: 21.644 ng/u1
RT: 19.757 min Scan# 21
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:202 Resp: 178845
Ion Ratio Lower Upper
202 100
101 6.6 5.8 8.6
100 5.7 4.6 7.0



#82
Pyrene
Concen: 21.609 ng/u1
RT: 20.121 min Scan# 2823
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:202 Resp: 180759
Ion Ratio Lower Upper
202 100
101 7.8 5.7 8.5
100 6.3 6.0 9.0









#83
Butylbenzylphthalate
Concen: 21.431 ng/ul
RT: 20.973 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:149 Resp: 57409
Ion Ratio Lower Upper
149 100
91 72.8 58.6 87.8
206 24.8 19.8 29.8



#84
3,3'-Dichlorobenzidine
Concen: 22.505 ng/ul
RT: 21.912 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:252 Resp: 58573
Ion Ratio Lower Upper
252 100
254 64.9 53.7 80.5
126 7.8 5.2 7.8#









#85
Benzo(a)anthracene
Concen: 21.627 ng/ul
RT: 22.018 min Scan# 3145
Delta R.T. 0.006 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29




Instrument :
BNA_G
ClientSampleId :
SSTD020465


Tgt Ion:228 Resp: 179252
Ion Ratio Lower Upper
228 100
229 19.4 15.8 23.6
226 26.9 21.7 32.5



#86
Bis(2-ethylhexyl)phthalate
Concen: 21.437 ng/ul
RT: 21.865 min Scan# 3120
Delta R.T. 0.000 min
Lab File: BG057224.D
Acq: 28 Apr 2023 19:29

Tgt Ion:149 Resp: 83519
Ion Ratio Lower Upper
149 100
167 26.6 23.0 34.6
279 5.2 3.6 5.4





#87

Chrysene

Concen: 21.220 ng/ul

RT: 22.089 min Scan# 3158

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:228 Resp: 165936

Ion Ratio Lower Upper

228 100

226 29.7 22.8 34.2

229 20.4 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 20.805 ng/ul




RT: 23.164 min Scan# 3341


Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Tgt Ion:149 Resp: 141322





#90

Benzo(b)fluoranthene

Concen: 21.516 ng/ul

RT: 24.421 min Scan# 3555

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:252 Resp: 183283

Ion Ratio Lower Upper

252 100

253 21.9 16.7 25.1

125 5.2 4.8 7.2



#91

Benzo(k)fluoranthene

Concen: 21.179 ng/ul

RT: 24.491 min Scan# 3567

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:252 Resp: 175513


Ion Ratio Lower Upper

252 100

253 22.0 17.7 26.5

125 5.3 4.5 6.7





#93

Benzo(a)pyrene

Concen: 21.243 ng/ul

RT: 25.378 min Scan# 3718

Delta R.T. 0.000 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29

Instrument :

BNA_G

ClientSampleId :

SSTD020465





Tgt Ion:252 Resp: 156218

Ion Ratio Lower Upper

252 100

253 21.7 17.2 25.8

125 6.2 5.1 7.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 21.197 ng/ul

RT: 29.596 min Scan# 4436

Delta R.T. 0.006 min

Lab File: BG057224.D

Acq: 28 Apr 2023 19:29




Tgt Ion:276 Resp: 194862


Ion Ratio Lower Upper

276 100

138 11.6 9.4 14.2

277 23.3 18.6 28.0









#95
 Dibenzo(a,h)anthracene
 Concen: 21.289 ng/ul
 RT: 29.649 min Scan# 4445
 Delta R.T. -0.012 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29

Instrument :
 BNA_G
 ClientSampleId :
 SST020465

Tgt Ion:278 Resp: 162281
 Ion Ratio Lower Upper
 278 100
 139 8.9 5.8 8.6#
 279 25.0 18.3 27.5



#96
 Benzo(g,h,i)perylene
 Concen: 12.030 ng/ul
 RT: 30.853 min Scan# 4650
 Delta R.T. -0.006 min
 Lab File: BG057224.D
 Acq: 28 Apr 2023 19:29

Tgt Ion:276 Resp: 89498
 Ion Ratio Lower Upper
 276 100
 138 12.0 8.7 13.1
 277 22.3 18.4 27.6

