


Hexachlorocyclopentadiene




$R = 3.11e-002 A^2 + 1.38e-002 A - 2.77e-003$
Coef of Det (r^2) = 0.999 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
Calibration Table Last Updated: Wed Apr 29 16:22:16 2015

2,4-Dinitrophenol

Response Ratio




$R = 2.71e-002 A^2 + 8.00e-002 A - 2.27e-002$
Coef of Det (r^2) = 1.000 Curve Fit: Quadratic

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
Calibration Table Last Updated: Wed Apr 29 16:22:16 2015

4-Nitrophenol


Response Ratio



Resp Ratio = $3.07 \times 10^{-1} \times \text{Amt} - 9.57 \times 10^{-2}$
Coef of Det (r^2) = 0.991 Curve Fit: Linear


Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
Calibration Table Last Updated: Wed Apr 29 16:22:16 2015

4,6-Dinitro-2-methylphenol



Resp Ratio = $1.47 \times 10^{-1} \times \text{Amt} - 4.56 \times 10^{-2}$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
Calibration Table Last Updated: Wed Apr 29 16:22:16 2015



Method Name: Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
Calibration Table Last Updated: Wed Apr 29 16:22:16 2015