


Data Path : Z:\HPCHEM1\BNA G\Data\BG042915\
 Data File : BG016792.D
 Acq On : 29 Apr 2015 11:13
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG042915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Apr 30 02:01:08 2015



AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2500

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.7	29794	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	37.5	35249	PASS
70	69	0.00	2	1.0	343	PASS
127	198	10	80	55.7	52410	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	94077	PASS
199	198	5	9	7.0	6545	PASS
275	198	10	60	20.3	19135	PASS
365	198	1	100	2.2	2045	PASS
441	442	0.01	24	21.6	13091	PASS
442	198	50	100	64.4	60584	PASS
443	442	15	24	19.3	11705	PASS