

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
 Data File : BG053378.D
 Acq On : 29 Apr 2022 15:45
 Operator : CG/JU
 Sample : N2616-10
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MH-20D

Quant Time: May 02 01:02:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	22499	20.000	ng	-0.01
21) Naphthalene-d8	10.913	136	85821	20.000	ng	# 0.00
39) Acenaphthene-d10	14.725	164	64053	20.000	ng	0.00
64) Phenanthrene-d10	17.475	188	152642	20.000	ng	0.00
76) Chrysene-d12	21.757	240	179439	20.000	ng	0.00
86) Perylene-d12	25.047	264	187703	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.667	112	134830	102.726	ng	0.00
7) Phenol-d6	7.259	99	198839	98.239	ng	-0.01
23) Nitrobenzene-d5	9.262	82	139222	65.868	ng	-0.01
42) 2,4,6-Tribromophenol	16.212	330	108438	106.410	ng	-0.01
45) 2-Fluorobiphenyl	13.351	172	303295	65.292	ng	0.00
79) Terphenyl-d14	20.077	244	642987	60.519	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
Data File : BG053378.D
Acq On : 29 Apr 2022 15:45
Operator : CG/JU
Sample : N2616-10
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
MH-20D

Quant Time: May 02 01:02:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

