

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
 Data File : BG053385.D
 Acq On : 29 Apr 2022 20:15
 Operator : CG/JU
 Sample : N2609-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 WC-1

Quant Time: May 02 01:03:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	18854	20.000	ng	-0.01
21) Naphthalene-d8	10.913	136	70820	20.000	ng	# 0.00
39) Acenaphthene-d10	14.726	164	54207	20.000	ng	0.00
64) Phenanthrene-d10	17.469	188	155589	20.000	ng	-0.01
76) Chrysene-d12	21.752	240	198487	20.000	ng	-0.01
86) Perylene-d12	25.053	264	198789	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.667	112	159559	145.069	ng	0.00
7) Phenol-d6	7.265	99	208846	123.131	ng	0.00
23) Nitrobenzene-d5	9.263	82	173403	99.418	ng	-0.01
42) 2,4,6-Tribromophenol	16.212	330	142272	164.970	ng	-0.01
45) 2-Fluorobiphenyl	13.351	172	378478	96.276	ng	0.00
79) Terphenyl-d14	20.077	244	1077715	91.702	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
Data File : BG053385.D
Acq On : 29 Apr 2022 20:15
Operator : CG/JU
Sample : N2609-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
WC-1

Quant Time: May 02 01:03:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

