

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
 Data File : BG053375.D
 Acq On : 29 Apr 2022 13:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 01:02:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	43.744	-9.4	122	0.00
3	Pyridine	40.000	44.284	-10.7	117	0.00
4	n-Nitrosodimethylamine	40.000	44.736	-11.8	119	0.00
5 S	2-Fluorophenol	80.000	86.885	-8.6	120	0.00
6	Aniline	40.000	40.779	-1.9	110	0.00
7 S	Phenol-d6	80.000	84.773	-6.0	114	0.00
8	2-Chlorophenol	40.000	41.067	-2.7	111	0.00
9	Benzaldehyde	40.000	41.738	-4.3	128	0.00
10 C	Phenol	40.000	42.244	-5.6	117	0.00
11	bis(2-Chloroethyl)ether	40.000	41.848	-4.6	117	0.00
12	1,3-Dichlorobenzene	40.000	41.959	-4.9	116	0.00
13 C	1,4-Dichlorobenzene	40.000	41.797	-4.5	115	0.00
14	1,2-Dichlorobenzene	40.000	40.170	-0.4	113	0.00
15	Benzyl Alcohol	40.000	40.644	-1.6	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.465	-3.7	111	-0.01
17	2-Methylphenol	40.000	41.346	-3.4	114	0.00
18	Hexachloroethane	40.000	42.134	-5.3	114	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.832	0.4	108	0.00
20	3+4-Methylphenols	40.000	40.735	-1.8	110	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	41.913	-4.8	110	0.00
23 S	Nitrobenzene-d5	80.000	82.463	-3.1	109	0.00
24	Nitrobenzene	40.000	41.277	-3.2	112	0.00
25	Isophorone	40.000	41.110	-2.8	107	0.00
26 C	2-Nitrophenol	40.000	42.116	-5.3	110	0.00
27	2,4-Dimethylphenol	40.000	41.878	-4.7	109	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.208	-5.5	110	0.00
29 C	2,4-Dichlorophenol	40.000	42.280	-5.7	109	0.00
30	1,2,4-Trichlorobenzene	40.000	41.606	-4.0	109	0.00
31	Naphthalene	40.000	40.637	-1.6	107	0.00
32	Benzoic acid	40.000	38.940	2.7	98	0.00
33	4-Chloroaniline	40.000	40.320	-0.8	104	0.00
34 C	Hexachlorobutadiene	40.000	41.570	-3.9	110	0.00
35	Caprolactam	40.000	38.854	2.9	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.618	-1.5	105	0.00
37	2-Methylnaphthalene	40.000	40.582	-1.5	107	0.00
38	1-Methylnaphthalene	40.000	39.794	0.5	105	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.697	-4.2	107	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.923	0.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	79.568	0.5	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.588	1.0	99	0.00
44	2,4,5-Trichlorophenol	40.000	38.796	3.0	97	0.00
45 S	2-Fluorobiphenyl	80.000	81.459	-1.8	104	0.00
46	1,1'-Biphenyl	40.000	40.644	-1.6	104	0.00
47	2-Chloronaphthalene	40.000	40.336	-0.8	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
 Data File : BG053375.D
 Acq On : 29 Apr 2022 13:38
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 01:02:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 01:13:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.738	-1.8	100	0.00
49	Acenaphthylene	40.000	40.741	-1.9	102	0.00
50	Dimethylphthalate	40.000	39.449	1.4	100	0.00
51	2,6-Dinitrotoluene	40.000	39.927	0.2	101	0.00
52 C	Acenaphthene	40.000	40.609	-1.5	102	0.00
53	3-Nitroaniline	40.000	41.730	-4.3	103	0.00
54 P	2,4-Dinitrophenol	40.000	37.831	5.4	92	0.00
55	Dibenzofuran	40.000	40.174	-0.4	103	0.00
56 P	4-Nitrophenol	40.000	40.211	-0.5	96	0.00
57	2,4-Dinitrotoluene	40.000	41.285	-3.2	102	0.00
58	Fluorene	40.000	39.817	0.5	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.292	-0.7	100	0.00
60	Diethylphthalate	40.000	39.327	1.7	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.962	0.1	101	0.00
62	4-Nitroaniline	40.000	41.603	-4.0	106	0.00
63	Azobenzene	40.000	40.443	-1.1	103	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.321	-3.3	101	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.532	-1.3	101	0.00
67	4-Bromophenyl-phenylether	40.000	41.358	-3.4	102	0.00
68	Hexachlorobenzene	40.000	40.664	-1.7	101	0.00
69	Atrazine	40.000	38.496	3.8	98	0.00
70 C	Pentachlorophenol	40.000	38.624	3.4	90	0.00
71	Phenanthrene	40.000	40.549	-1.4	102	0.00
72	Anthracene	40.000	40.457	-1.1	102	0.00
73	Carbazole	40.000	40.984	-2.5	103	0.00
74	Di-n-butylphthalate	40.000	40.517	-1.3	102	0.00
75 C	Fluoranthene	40.000	40.802	-2.0	104	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	37.909	5.2	108	0.00
78	Pyrene	40.000	38.828	2.9	103	0.00
79 S	Terphenyl-d14	80.000	76.142	4.8	101	0.00
80	Butylbenzylphthalate	40.000	41.205	-3.0	108	0.00
81	Benzo(a)anthracene	40.000	40.985	-2.5	109	0.00
82	3,3'-Dichlorobenzidine	40.000	40.533	-1.3	108	0.00
83	Chrysene	40.000	40.735	-1.8	108	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.955	-4.9	112	0.00
85 c	Di-n-octyl phthalate	40.000	42.661	-6.7	113	0.00
86 I	Perylene-d12	20.000	20.000	0.0	115	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.157	-2.9	114	0.00
88	Benzo(b)fluoranthene	40.000	40.974	-2.4	114	-0.01
89	Benzo(k)fluoranthene	40.000	40.524	-1.3	113	0.00
90 C	Benzo(a)pyrene	40.000	41.405	-3.5	115	0.00
91	Dibenzo(a,h)anthracene	40.000	40.794	-2.0	114	0.00
92	Benzo(g,h,i)perylene	40.000	41.625	-4.1	116	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG042922\
Data File : BG053375.D
Acq On : 29 Apr 2022 13:38
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 02 01:02:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 01:13:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0