




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG043024\
Data File : BG061057.D
Acq On : 30 Apr 2024 10:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 03 13:27:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061057.D\data.ms

(70) Pentachlorophenol (C)




16.984min (-0.002) 44271.60 ng

response 68990

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.80	63.91
264.00	64.40	59.47
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG043024\
Data File : BG061057.D
Acq On : 30 Apr 2024 10:01
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 03 13:27:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061057.D\data.ms

(77) Benzidine

19.563min (-0.002) 285195.57 ng m

response 368805

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	13.32
183.00	13.60	11.95
0.00	0.00	0.00