

Data File : BG026836.D
Acq On : 2 May 2017 10:16
Operator : SJ/MA
Sample : I2908-05
Misc :
ALS Vial : 43 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
SVOC-GPC-BLANK

Quant Time: May 03 00:58:48 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 03 00:58:45 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	107928	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	602132	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	359592	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.36	188	680572	20.00	ng/ul	0.00
75) Chrysene-d12	21.60	240	620552	20.00	ng/ul	0.00
83) Perylene-d12	24.77	264	633879	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.42	96	71	0.03	ng/uL	0.00
5) Phenol-d5	6.98	99	104	0.01	ng/ul	-0.21
7) Bis-(2-Chloroethyl)ether-d	7.33	67	296	0.05	ng/ul	0.00
9) 2-Chlorophenol-d4	7.49	132	124	0.02	ng/ul	-0.06
13) 4-Methylphenol-d8	8.68	113	55	0.01	ng/ul	-0.05
19) Nitrobenzene-d5	9.15	128	58	0.01	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.94	143	59	0.01	ng/ul	0.05
26) 2,4-Dichlorophenol-d3	10.44	165	63	0.01	ng/ul	0.00
29) 4-Chloroaniline-d4	10.91	131	56	0.00	ng/ul	-0.03
43) Dimethylphthalate-d6	0.00	166	0	0.00	ng/ul	
46) Acenaphthylene-d8	0.00	160	0	0.00	ng/ul	
51) 4-Nitrophenol-d4	14.84	143	56	0.01	ng/ul	0.00
57) Fluorene-d10	15.63	176	2077	0.08	ng/ul	0.01
62) 4,6-Dinitro-2-methylphenol	0.00	200	0	0.00	ng/ul	
70) Anthracene-d10	17.47	188	3488	0.11	ng/ul	0.01
76) Pyrene-d10	19.73	212	76	0.00	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.77	264	627210	20.93	ng/ul	0.22

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BG026836.D

Acq On : 2 May 2017 10:16

Operator : SJ/MA

Sample : I2908-05

Misc :

ALS Vial : 43 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SVOC-GPC-BLANK

Quant Time: May 03 00:58:48 2017

Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M

Quant Title : SVOA CALIBRATION

QLast Update : Wed May 03 00:58:45 2017

Response via : Initial Calibration

