

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050118\
 Data File : BG034075.D
 Acq On : 1 May 2018 12:33
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 5/2/2018 4:50:29 PM

Quant Time: May 01 17:16:11 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 01 15:30:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.08	152	54713	20.00	ng	0.00
21) Naphthalene-d8	10.89	136	217548	20.00	ng	0.00
38) Acenaphthene-d10	14.72	164	174242	20.00	ng	0.00
63) Phenanthrene-d10	17.47	188	477986	20.00	ng	0.00
75) Chrysene-d12	21.76	240	568583	20.00	ng	0.00
86) Perylene-d12	25.05	264	542969	20.00	ng	-0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.63	112	18094	5.28	ng	0.00
7) Phenol-d6	7.21	99	27374	5.48	ng	0.00
23) Nitrobenzene-d5	9.23	82	26278	6.87	ng	0.00
41) 2,4,6-Tribromophenol	16.20	330	12498	6.15	ng	0.00
44) 2-Fluorobiphenyl	13.34	172	65472	5.52	ng	0.00
78) Terphenyl-d14	20.09	244	142420	5.63	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.56	88	4220	2.914	ng	# 62
4) n-Nitrosodimethylamine	3.87	42	5961	3.123	ng	# 82
6) Aniline	7.40	93	17777	2.661	ng	# 87
8) 2-Chlorophenol	7.63	128	8779	2.286	ng	# 88
10) Phenol	7.25	94	12513	2.444	ng	# 61
11) bis(2-Chloroethyl)ether	7.49	93	10704	2.733	ng	# 82
12) 1,3-Dichlorobenzene	7.96	146	12837	2.904	ng	# 80
13) 1,4-Dichlorobenzene	8.12	146	12058m	2.693	ng	
14) 1,2-Dichlorobenzene	8.43	146	11016	2.530	ng	90
15) Benzyl Alcohol	8.31	79	13403	3.075	ng	91
16) 2,2'-oxybis(1-Chloropropan	8.61	45	14567	1.907	ng	76
17) 2-Methylphenol	8.50	107	8287	2.184	ng	# 73
18) Hexachloroethane	9.17	117	5305	3.243	ng	# 75
19) n-Nitroso-di-n-propylamine	8.88	70	12211	2.866	ng	# 87
20) 3+4-Methylphenols	8.83	107	12957	2.350	ng	# 78
22) Acetophenone	8.89	105	18383	3.105	ng	# 90
24) Nitrobenzene	9.28	77	14711	3.532	ng	# 79
25) Isophorone	9.80	82	31067	3.615	ng	# 92
26) 2-Nitrophenol	10.00	139	3561	2.164	ng	# 61
27) 2,4-Dimethylphenol	10.04	122	8036	2.610	ng	# 60
28) bis(2-Chloroethoxy)methane	10.29	93	15828	3.485	ng	# 87
29) 2,4-Dichlorophenol	10.52	162	9977	2.907	ng	# 79
30) 1,2,4-Trichlorobenzene	10.75	180	12330	3.208	ng	# 87
31) Naphthalene	10.94	128	28495	2.547	ng	# 85
33) 4-Chloroaniline	11.04	127	11679	2.221	ng	# 74
34) Hexachlorobutadiene	11.23	225	10578	4.600	ng	97
35) Caprolactam	11.78	113	3517	2.410	ng	# 60
36) 4-Chloro-3-methylphenol	12.15	107	12680	3.027	ng	# 74
37) 2-Methylnaphthalene	12.54	142	24440m	2.853	ng	
39) 1,2,4,5-Tetrachlorobenzene	12.91	216	17399	3.007	ng	# 91
40) Hexachlorocyclopentadiene	12.89	237	10090	3.172	ng	92
42) 2,4,6-Trichlorophenol	13.15	196	10480	2.971	ng	# 73
43) 2,4,5-Trichlorophenol	13.21	196	11058	2.727	ng	# 94
45) 1,1'-Biphenyl	13.55	154	36278	2.572	ng	95

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050118\
 Data File : BG034075.D
 Acq On : 1 May 2018 12:33
 Operator : SJ/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Manual Integrations
 APPROVED

Sohil
 5/2/2018 4:50:29 PM

Quant Time: May 01 17:16:11 2018
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050118.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 01 15:30:49 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) 2-Chloronaphthalene	13.59	162	26107	2.441	nq	91
47) 2-Nitroaniline	13.78	65	8712	2.609	nq #	77
48) Acenaphthylene	14.44	152	44756	2.518	nq	95
49) Dimethylphthalate	14.17	163	42073	2.745	nq	97
50) 2,6-Dinitrotoluene	14.28	165	5905	2.033	nq	92
51) Acenaphthene	14.78	154	28265	2.531	nq	92
52) 3-Nitroaniline	14.60	138	7290	2.311	nq #	54
54) Dibenzofuran	15.12	168	43494	2.627	nq #	91
55) 4-Nitrophenol	14.90	139	5357	2.166	nq #	53
56) 2,4-Dinitrotoluene	15.06	165	7517	1.946	nq #	83
57) Fluorene	15.77	166	40202	2.806	nq	97
58) 2,3,4,6-Tetrachlorophenol	15.34	232	11788	3.175	nq	96
59) Diethylphthalate	15.54	149	41626	2.569	nq #	92
60) 4-Chlorophenyl-phenylether	15.76	204	22501	3.003	nq	94
61) 4-Nitroaniline	15.76	138	7341	2.192	nq #	65
62) Azobenzene	16.06	77	48013	3.325	nq	96
65) n-Nitrosodiphenylamine	15.97	169	33023	2.376	nq	96
66) 4-Bromophenyl-phenylether	16.65	248	15631	2.978	nq #	85
67) Hexachlorobenzene	16.77	284	14376	2.568	nq #	76
68) Atrazine	16.92	200	15726	2.889	nq	88
69) Pentachlorophenol	17.11	266	9324	2.425	nq	91
70) Phenanthrene	17.51	178	66302	2.586	nq	96
71) Anthracene	17.60	178	66398m	2.589	nq	
72) Carbazole	17.87	167	62074	2.509	nq	97
73) Di-n-butylphthalate	18.45	149	76372	2.482	nq #	97
74) Fluoranthene	19.52	202	87531	2.822	nq	94
76) Benzidine	19.70	184	32766	2.134	nq	97
77) Pvrene	19.88	202	89946	2.412	nq	97
79) Butylbenzylphthalate	20.79	149	33582	2.051	nq #	84
80) Benzo(a)anthracene	21.74	228	97280m	2.713	nq	
81) 3,3'-Dichlorobenzidine	21.65	252	35176	2.631	nq #	92
82) Chrysene	21.81	228	92802	2.711	nq	93
83) Bis(2-ethylhexyl)phthalate	21.68	149	47328	2.046	nq #	99
84) Di-n-octyl phthalate	22.94	149	73775	2.010	nq #	94
85) Indeno(1,2,3-cd)pvrene	28.79	276	88389m	2.268	nq	
87) Benzo(b)fluoranthene	24.01	252	89901	2.714	nq #	87
88) Benzo(k)fluoranthene	24.08	252	84254	2.561	nq #	90
89) Benzo(a)pyrene	24.89	252	79012	2.491	nq	95
90) Dibenzo(a,h)anthracene	28.88	278	79773	2.599	nq #	80
91) Benzo(g,h,i)perylene	29.97	276	76277m	2.525	nq	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG050118\
Data File : BG034075.D
Acq On : 1 May 2018 12:33
Operator : SJ/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM

Quant Time: May 01 17:16:11 2018
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050118.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 01 15:30:49 2018
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.000 ng

RT: 8.08 min Scan# 815

Delta R.T. 0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
152	100		
150	190.0	126.6	189.8#
115	65.8	53.0	79.4

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#2

1,4-Dioxane

Concen: 2.914 ng




RT: 3.56 min Scan# 46


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
88	100		
58	87.0	79.6	119.4
43	98.6	27.4	41.0#





#4





n-Nitrosodimethylamine
 Concen: 3.123 ng
 RT: 3.87 min Scan# 99
 Delta R.T. 0.01 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
42	100		
74	105.9	102.5	153.7
44	29.8	18.9	28.3

Manual Integrations
 APPROVED




Sohil
 5/2/2018 4:50:29 PM




#5

2-Fluorophenol
 Concen: 5.280 ng
 RT: 5.63 min Scan# 399
 Delta R.T. -0.00 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
112	100		
64	61.1	56.3	84.5
63	41.2	30.1	45.1





#6

Aniline

Concen: 2.661 ng

RT: 7.40 min Scan# 699

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	93	Resp:	17777
Ion Ratio	Lower	Upper	
93	100		
66	30.6	33.7	50.5#
65	19.5	16.1	24.1

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#7

Phenol-d6

Concen: 5.476 ng




RT: 7.21 min Scan# 668


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	99	Resp:	27374
Ion Ratio	Lower	Upper	
99	100		
42	25.3	14.1	21.1#
71	45.7	26.1	39.1#





#8

2-Chlorophenol

Concen: 2.286 ng

RT: 7.63 min Scan# 739

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	128	Resp:	8779
Ion Ratio	Lower	Upper	
128	100		
130	38.8	14.0	54.0
64	68.8	37.7	77.7

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#10

Phenol

Concen: 2.444 ng




RT: 7.25 min Scan# 674


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	94	Resp:	12513
Ion Ratio	Lower	Upper	
94	100		
65	48.0	11.9	51.9
66	71.6	22.2	62.2#





#11

bis(2-Chloroethyl)ether

Concen: 2.733 ng

RT: 7.49 min Scan# 716

Delta R.T. 0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	93	Resp:	10704
Ion	Ratio	Lower	Upper
93	100		
63	108.4	67.1	107.1#
95	28.6	11.5	51.5

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#12

1,3-Dichlorobenzene

Concen: 2.904 ng




RT: 7.96 min Scan# 796


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	146	Resp:	12837
Ion	Ratio	Lower	Upper
146	100		
148	43.1	51.4	77.2#
75	37.3	26.6	39.8





#13

1,4-Dichlorobenzene

Concen: 2.693 ng m

RT: 8.12 min Scan# 822

Delta R.T. 0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	146	Resp:	12058
Ion Ratio	Lower	Upper	
146	100		
148	51.0	53.7	80.5#
111	29.1	35.2	52.8#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#14

1,2-Dichlorobenzene

Concen: 2.530 ng




RT: 8.43 min Scan# 875


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	146	Resp:	11016
Ion Ratio	Lower	Upper	
146	100		
148	69.7	49.3	73.9
111	48.7	34.3	51.5





#15

Benzyl Alcohol

Concen: 3.075 ng

RT: 8.31 min Scan# 854

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampled :

SSTDICC2.5

Tot Ion:	79	Resp:	13403
Ion Ratio	Lower	Upper	
79	100		
108	57.8	57.2	85.8
77	57.1	46.1	69.1

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#16

2,2'-oxybis(1-chloropropane)

Concen: 1.907 ng




RT: 8.61 min Scan# 905


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	45	Resp:	14567
Ion Ratio	Lower	Upper	
45	100		
77	20.4	0.0	30.2
79	15.2	0.0	27.9









#17
2-Methylphenol
Concen: 2.184 ng
RT: 8.50 min Scan# 887
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
107	100			
108	70.5	83.9	125.9#	
77	62.1	37.2	55.8#	
79	56.4	36.2	54.2#	




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#18
Hexachloroethane
Concen: 3.243 ng
RT: 9.17 min Scan# 1001
Delta R.T. 0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
117	100			
119	70.2	76.7	115.1#	
201	92.2	95.7	143.5#	





#19

n-Nitroso-di-n-propylamine

Concen: 2.866 ng

RT: 8.88 min Scan# 951

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
70	100		
42	48.2	49.1	73.7#
101	7.3	8.5	12.7#
130	15.7	13.4	20.0

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#20

3+4-Methylphenols

Concen: 2.350 ng




RT: 8.83 min Scan# 943


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
107	100		
108	107.2	61.6	101.6#
77	41.4	13.9	53.9
79	35.2	8.8	48.8









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.89 min Scan# 1294
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
136	100	217548		
137	11.5		9.2	13.8
54	11.2		10.3	15.5
68	3.5		4.5	6.7#




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#22
Acetophenone
Concen: 3.105 ng
RT: 8.89 min Scan# 954
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
105	100	18383		
71	2.8		3.0	4.4#
51	37.9		26.1	39.1
120	27.2		17.4	26.2#









#23
 Nitrobenzene-d5
 Concen: 6.874 ng
 RT: 9.23 min Scan# 1012
 Delta R.T. -0.01 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
82	100		
128	20.7	29.7	44.5#
54	46.1	43.1	64.7




Manual Integrations
 APPROVED


Sohil
 5/2/2018 4:50:29 PM



#24
 Nitrobenzene
 Concen: 3.532 ng
 RT: 9.28 min Scan# 1019
 Delta R.T. -0.01 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
77	100		
123	28.1	33.0	49.6#
65	7.5	13.0	19.6#





#25

Isophorone

Concen: 3.615 ng

RT: 9.80 min Scan# 1109

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	82	Resp:	31067
Ion Ratio	Lower	Upper	
82	100		
95	4.8	6.2	9.2#
138	11.8	12.1	18.1#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#26

2-Nitrophenol

Concen: 2.164 ng




RT: 10.00 min Scan# 1142


Delta R.T. 0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	139	Resp:	3561
Ion Ratio	Lower	Upper	
139	100		
109	85.8	24.2	36.2#
65	54.2	47.4	71.0





#27

2,4-Dimethylphenol

Concen: 2.610 ng

RT: 10.04 min Scan# 1150

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
122	100		
107	186.7	100.2	150.2#
121	61.7	44.4	66.6

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#28

bis(2-Chloroethoxy)methane

Concen: 3.485 ng




RT: 10.29 min Scan# 1192


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
93	100		
95	22.2	24.3	36.5#
123	8.1	8.4	12.6#









#29
 2,4-Dichlorophenol
 Concen: 2.907 ng
 RT: 10.52 min Scan# 1231
 Delta R.T. -0.01 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
162	100	9977		
164	86.8	48.0	88.0	
98	29.6	21.1	61.1	




Manual Integrations
 APPROVED


Sohil
 5/2/2018 4:50:29 PM



#30
 1,2,4-Trichlorobenzene
 Concen: 3.208 ng
 RT: 10.75 min Scan# 1270
 Delta R.T. -0.00 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
180	100	12330		
182	107.6	77.5	116.3	
145	39.1	23.4	35.2	









#31
Naphthalene
Concen: 2.547 ng
RT: 10.94 min Scan# 1303
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
128	100	28495		
129	6.5		8.6	12.8#
127	21.6		11.6	17.4#




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#33
4-Chloroaniline
Concen: 2.221 ng
RT: 11.04 min Scan# 1320
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
127	100	11679		
129	35.2		24.0	36.0
65	64.6		27.0	40.4#
92	19.7		16.6	25.0





#34

Hexachlorobutadiene

Concen: 4.600 ng

RT: 11.23 min Scan# 1352

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	225	Resp:	10578
Ion	Ratio	Lower	Upper
225	100		
223	63.7	49.7	74.5
227	57.8	48.1	72.1

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#35

Caprolactam

Concen: 2.410 ng




RT: 11.78 min Scan# 1446


Delta R.T. -0.03 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	113	Resp:	3517
Ion	Ratio	Lower	Upper
113	100		
55	165.8	186.9	226.9#
56	76.0	130.9	170.9#









#36
4-Chloro-3-methylphenol
Concen: 3.027 ng
RT: 12.15 min Scan# 1509
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
107	100		
144	14.6	18.2	27.4#
142	49.9	59.0	88.4#




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#37
2-Methylnaphthalene
Concen: 2.853 ng m
RT: 12.54 min Scan# 1575
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
142	100		
141	67.6	67.1	100.7
115	65.5	30.7	46.1#





#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.72 min Scan# 1945

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	164	Resp:	174242
Ion Ratio	Lower	Upper	
164	100		
162	98.0	78.8	118.2
160	43.1	35.4	53.0

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#39

1,2,4,5-Tetrachlorobenzene

Concen: 3.007 ng




RT: 12.91 min Scan# 1637


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	216	Resp:	17399
Ion Ratio	Lower	Upper	
216	100		
214	86.4	63.0	94.4
179	19.3	17.0	25.6
108	21.8	12.8	19.2#





#40

Hexachlorocyclopentadiene

Concen: 3.172 ng

RT: 12.89 min Scan# 1635

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	237	Resp:	10090
Ion Ratio		Lower	Upper
237	100		
235	63.0	50.4	90.4
272	10.4	0.0	31.8

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#41

2,4,6-Tribromophenol

Concen: 6.148 ng




RT: 16.20 min Scan# 2198


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	330	Resp:	12498
Ion Ratio		Lower	Upper
330	100		
332	93.6	77.0	115.6
141	44.8	25.2	37.8#









#42
 2,4,6-Trichlorophenol
 Concen: 2.971 ng
 RT: 13.15 min Scan# 1678
 Delta R.T. 0.00 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
196	100		
198	65.0	78.2	117.2#
200	25.4	24.6	36.8




Manual Integrations
 APPROVED


Sohil
 5/2/2018 4:50:29 PM



#43
 2,4,5-Trichlorophenol
 Concen: 2.727 ng
 RT: 13.21 min Scan# 1689
 Delta R.T. -0.00 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
196	100		
198	98.5	76.2	114.4
97	53.3	38.7	58.1
132	19.6	20.2	30.2#









#44
 2-Fluorobiphenyl
 Concen: 5.520 ng
 RT: 13.34 min Scan# 1711
 Delta R.T. -0.01 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Instrument :
 BNA_G
 ClientSampleId :
 SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
172	100		
171	35.7	30.0	45.0
170	27.1	19.5	29.3




Manual Integrations
 APPROVED


Sohil
 5/2/2018 4:50:29 PM



#45
 1,1'-Biphenyl
 Concen: 2.572 ng
 RT: 13.55 min Scan# 1747
 Delta R.T. -0.00 min
 Lab File: BG034075.D
 Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
154	100		
153	43.4	19.8	59.8
76	11.9	0.0	31.9









#46
2-Chloronaphthalene
Concen: 2.441 ng
RT: 13.59 min Scan# 1754
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion:162	Resp:	26107
Ion Ratio	Lower	Upper
162 100		
127 30.8	30.7	46.1
164 35.2	26.0	39.0




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#47
2-Nitroaniline
Concen: 2.609 ng
RT: 13.78 min Scan# 1786
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion: 65	Resp:	8712
Ion Ratio	Lower	Upper
65 100		
92 43.5	54.6	82.0#
138 74.4	72.9	109.3





#48

Acenaphthylene

Concen: 2.518 ng

RT: 14.44 min Scan# 1898

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	152	Resp:	44756
Ion Ratio	Lower	Upper	
152	100		
151	18.2	17.2	25.8
153	11.6	10.2	15.4

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#49

Dimethylphthalate

Concen: 2.745 ng




RT: 14.17 min Scan# 1852


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	163	Resp:	42073
Ion Ratio	Lower	Upper	
163	100		
194	4.7	3.4	5.2
164	8.7	8.2	12.4









#50
2,6-Dinitrotoluene
Concen: 2.033 ng
RT: 14.28 min Scan# 1870
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
165	100		
63	56.7	52.7	79.1
89	58.4	49.2	73.8




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#51
Acenaphthene
Concen: 2.531 ng
RT: 14.78 min Scan# 1956
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
154	100		
153	102.7	90.4	135.6
152	54.4	41.6	62.4









#52
3-Nitroaniline
Concen: 2.311 ng
RT: 14.60 min Scan# 1926
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
138	100		
108	30.3	17.5	26.3#
92	191.6	105.8	158.8#




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#54
Dibenzofuran
Concen: 2.627 ng
RT: 15.12 min Scan# 2013
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
168	100		
139	41.3	28.6	43.0
169	11.1	11.2	16.8#









#55
4-Nitrophenol
Concen: 2.166 ng
RT: 14.90 min Scan# 1976
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
139	100			
109	121.9	62.2	102.2#	
65	169.8	97.2	137.2#	




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#56
2,4-Dinitrotoluene
Concen: 1.946 ng
RT: 15.06 min Scan# 2004
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
165	100			
63	37.5	42.0	63.0#	
89	70.7	66.9	100.3	
182	0.0	2.6	3.8#	





#57

Fluorene

Concen: 2.806 ng

RT: 15.77 min Scan# 2124

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
166	100		
165	97.7	80.6	121.0
167	11.0	10.4	15.6

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#58

2,3,4,6-Tetrachlorophenol

Concen: 3.175 ng




RT: 15.34 min Scan# 2051


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
232	100		
131	38.5	33.5	50.3
130	2.7	2.2	3.2
166	29.9	24.8	37.2









#59
Diethylphthalate
Concen: 2.569 ng
RT: 15.54 min Scan# 2085
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
149	100		
177	17.4	18.5	27.7#
150	12.0	10.2	15.2




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#60
4-Chlorophenyl-phenylether
Concen: 3.003 ng
RT: 15.76 min Scan# 2123
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
204	100		
206	26.6	26.2	39.2
141	59.9	45.8	68.6









#61
4-Nitroaniline
Concen: 2.192 ng
RT: 15.76 min Scan# 2123
Delta R.T. -0.02 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
138	100		
92	67.0	40.1	80.1
108	134.6	65.4	105.4




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#62
Azobenzene
Concen: 3.325 ng
RT: 16.06 min Scan# 2173
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
77	100		
182	24.4	7.4	47.4
105	17.9	0.3	40.3
51	32.2	11.6	51.6





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.47 min Scan# 2414

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	188	Resp:	477986
Ion Ratio	Lower	Upper	
188	100		
94	6.1	6.9	10.3#
80	6.9	7.7	11.5#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#65

n-Nitrosodiphenylamine

Concen: 2.376 ng




RT: 15.97 min Scan# 2159


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	169	Resp:	33023
Ion Ratio	Lower	Upper	
169	100		
168	73.6	55.7	83.5
167	35.4	30.1	45.1









#66
4-Bromophenyl-phenylether
Concen: 2.978 ng
RT: 16.65 min Scan# 2275
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
248	100		
250	73.0	77.1	115.7#
141	64.9	50.8	76.2




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#67
Hexachlorobenzene
Concen: 2.568 ng
RT: 16.77 min Scan# 2295
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
284	100		
142	27.8	26.8	40.2
249	54.5	26.3	39.5#









#68
Atrazine
Concen: 2.889 ng
RT: 16.92 min Scan# 2320
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
200	100		
173	21.9	5.2	45.2
215	40.6	30.4	70.4




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#69
Pentachlorophenol
Concen: 2.425 ng
RT: 17.11 min Scan# 2353
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
266	100		
268	68.5	51.1	76.7
264	71.5	49.6	74.4









#70
Phenanthrene
Concen: 2.586 ng
RT: 17.51 min Scan# 2421
Delta R.T. -0.00 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
178	100		
176	20.2	15.7	23.5
179	12.6	12.5	18.7




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#71
Anthracene
Concen: 2.589 ng m
RT: 17.60 min Scan# 2436
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
178	100		
176	24.2	16.0	24.0#
179	16.1	12.5	18.7





#72

Carbazole

Concen: 2.509 ng

RT: 17.87 min Scan# 2481

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	167	Resp:	62074
Ion Ratio	Lower	Upper	
167	100		
166	24.2	18.2	27.4
139	13.4	9.4	14.2

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#73

Di-n-butylphthalate

Concen: 2.482 ng




RT: 18.45 min Scan# 2580


Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	149	Resp:	76372
Ion Ratio	Lower	Upper	
149	100		
150	10.8	7.8	11.6
104	6.1	3.9	5.9





#74

Fluoranthene

Concen: 2.822 ng

RT: 19.52 min Scan# 2763

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion	Ratio	Resp	Lower	Upper
202	100	87531		
101	7.4	0.0	28.9	
203	20.2	0.0	37.5	

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#75

Chrysene-d12

Concen: 20.000 ng




RT: 21.76 min Scan# 3144


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	568583		
120	4.6	6.8	10.2#	
236	26.4	20.9	31.3	





#76

Benzidine

Concen: 2.134 ng

RT: 19.70 min Scan# 2793

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	184	Resp:	32766
Ion Ratio	Lower	Upper	
184	100		
185	13.3	11.1	16.7
183	11.3	10.6	16.0

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#77

Pyrene

Concen: 2.412 ng




RT: 19.88 min Scan# 2824


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	202	Resp:	89946
Ion Ratio	Lower	Upper	
202	100		
200	21.9	16.9	25.3
203	19.1	13.9	20.9





#78

Terphenyl-d14

Concen: 5.627 ng

RT: 20.09 min Scan# 2859

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	244	Resp:	142420
Ion Ratio	Lower	Upper	
244	100		
212	6.3	5.8	8.8
122	5.1	7.0	10.4#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#79

Butylbenzylphthalate

Concen: 2.051 ng




RT: 20.79 min Scan# 2978


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	149	Resp:	33582
Ion Ratio	Lower	Upper	
149	100		
91	91.0	62.2	93.2
206	32.0	18.6	27.8#









#80
Benzo(a)anthracene
Concen: 2.713 ng/ml
RT: 21.74 min Scan# 3141
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
228	100		
226	27.1	22.7	34.1
229	15.0	16.2	24.2




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#81
3,3'-Dichlorobenzidine
Concen: 2.631 ng/ml
RT: 21.65 min Scan# 3126
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
252	100		
254	68.8	50.8	76.2
126	4.8	7.4	11.2





#82

Chrysene

Concen: 2.711 ng

RT: 21.81 min Scan# 3153

Delta R.T. -0.00 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	228	Resp:	92802
Ion Ratio	Lower	Upper	
228	100		
226	26.9	24.9	37.3
229	16.6	15.0	22.4

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#83

Bis(2-ethylhexyl)phthalate

Concen: 2.046 ng




RT: 21.68 min Scan# 3130


Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	149	Resp:	47328
Ion Ratio	Lower	Upper	
149	100		
167	30.0	24.3	36.5
279	3.5	4.0	6.0#





#84

Di-n-octyl phthalate

Concen: 2.010 ng

RT: 22.94 min Scan# 3344

Delta R.T. -0.01 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	149	Resp:	73775
Ion Ratio	Lower	Upper	
149	100		
167	1.7	1.4	2.0
43	13.9	8.9	13.3#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#85

Indeno(1,2,3-cd)pyrene

Concen: 2.268 ng m




RT: 28.79 min Scan# 4340


Delta R.T. -0.05 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	276	Resp:	88389
Ion Ratio	Lower	Upper	
276	100		
138	1.2	16.0	24.0#
277	19.3	20.0	30.0#









#86
Pervlene-d12
Concen: 20.000 ng
RT: 25.05 min Scan# 3704
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampleId :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.5	17.4	26.0
265	21.1	17.7	26.5




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#87
Benzo(b)fluoranthene
Concen: 2.714 ng
RT: 24.01 min Scan# 3527
Delta R.T. -0.01 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	16.0	18.2	27.4#
125	5.1	7.2	10.8#









#88
Benzo(k)fluoranthene
Concen: 2.561 ng
RT: 24.08 min Scan# 3538
Delta R.T. -0.02 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Instrument :
BNA_G
ClientSampled :
SSTDICC2.5

Tot Ion	Ratio	Lower	Upper
252	100		
253	17.7	17.4	26.2
125	3.3	6.6	9.8#




Manual Integrations
APPROVED


Sohil
5/2/2018 4:50:29 PM



#89
Benzo(a)pyrene
Concen: 2.491 ng
RT: 24.89 min Scan# 3676
Delta R.T. -0.03 min
Lab File: BG034075.D
Acq: 1 May 2018 12:33

Tgt Ion	Ratio	Lower	Upper
252	100		
253	19.7	18.3	27.5
125	8.3	7.3	10.9





#90

Dibenzo(a,h)anthracene

Concen: 2.599 ng

RT: 28.88 min Scan# 4356

Delta R.T. -0.04 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Instrument :

BNA_G





ClientSampleId :

SSTDICC2.5

Tot Ion:	278	Resp:	79773
Ion Ratio	Lower	Upper	
278	100		
139	3.1	9.8	14.6#
279	14.4	18.4	27.6#

Manual Integrations
APPROVED

Sohil
5/2/2018 4:50:29 PM



#91

Benzo(g,h,i)perylene

Concen: 2.525 ng m

RT: 29.97 min Scan# 4542

Delta R.T. -0.06 min

Lab File: BG034075.D

Acq: 1 May 2018 12:33

Tgt Ion:	276	Resp:	76277
Ion Ratio	Lower	Upper	
276	100		
277	23.2	18.3	27.5
138	10.0	13.9	20.9#

