




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050224\
Data File : BG061085.D
Acq On : 2 May 2024 9:10
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 02 12:18:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061085.D\data.ms

(70) Pentachlorophenol (C)




16.980min (-0.006) 152357.80 ng

response 72737

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.80	65.37
264.00	64.40	64.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050224\
Data File : BG061085.D
Acq On : 2 May 2024 9:10
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 02 12:18:41 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061085.D\data.ms

(77) Benzidine

19.560min (-0.005) 268689.53 ng

response 369176

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.44
183.00	13.60	12.70
0.00	0.00	0.00