

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050317\
 Data File : BG026853.D
 Acq On : 3 May 2017 16:06
 Operator : SJ/MA
 Sample : I2871-07
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 YAB18

Manual Integrations
 APPROVED

mohammad
 5/4/2017 1:47:13 PM

Quant Time: May 04 01:18:40 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 04 00:50:13 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.01	152	157619	20.00	ng/ul	0.00
18) Naphthalene-d8	10.81	136	781026	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.62	164	437281	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.35	188	853653	20.00	ng/ul	0.00
75) Chrysene-d12	21.60	240	771164	20.00	ng/ul	0.00
83) Perylene-d12	24.76	264	787198	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.42	96	20484	5.85	ng/uL	0.00
5) Phenol-d5	7.18	99	476020	33.84	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.33	67	285966	34.57	ng/ul	0.00
9) 2-Chlorophenol-d4	7.55	132	386778	32.39	ng/ul	0.00
13) 4-Methylphenol-d8	8.72	113	403349	34.55	ng/ul	0.00
19) Nitrobenzene-d5	9.16	128	199472	30.55	ng/ul	0.00
22) 2-Nitrophenol-d4	9.89	143	225582	31.77	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.44	165	356862m	31.63	ng/ul	0.00
29) 4-Chloroaniline-d4	10.94	131	543586	35.99	ng/ul	0.00
43) Dimethylphthalate-d6	14.02	166	1122637	31.90	ng/ul	0.00
46) Acenaphthylene-d8	14.32	160	1463986	33.11	ng/ul	0.00
51) 4-Nitrophenol-d4	14.84	143	175876	27.72	ng/ul	0.00
57) Fluorene-d10	15.61	176	955017	31.09	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.73	200	126200	24.35	ng/ul	0.00
70) Anthracene-d10	17.45	188	1335833	32.24	ng/ul	0.00
76) Pyrene-d10	19.73	212	1313988	32.28	ng/ul	0.00
87) Benzo(a)pyrene-d12	24.54	264	1229203	33.03	ng/ul	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Dimethylphthalate	14.07	163	418882	12.16	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050317\
Data File : BG026853.D
Acq On : 3 May 2017 16:06
Operator : SJ/MA
Sample : I2871-07
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
YAB18

Manual Integrations
APPROVED

mohammad
5/4/2017 1:47:13 PM

Quant Time: May 04 01:18:40 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 04 00:50:13 2017
Response via : Initial Calibration





#44
 Dimethylphthalate
 Concen: 12.16 ng/ul
 RT: 14.07 min Scan# 1869
 Delta R.T. -0.00 min
 Lab File: BG026853.D
 Acq: 3 May 2017 16:06

Instrument :
 BNA_G
 ClientSampleId :
 YAB18

Tgt Ion	Ratio	Resp Lower	Resp Upper
163	100		
194	4.3	3.4	5.0
164	10.6	9.0	13.4

Manual Integrations
 APPROVED

mohammad
 5/4/2017 1:47:13 PM

