


Data Path : Z:\HPCHEM1\BNA_G\Data\BG050317\
Data File : BG026846.D
Acq On : 3 May 2017 11:24
Operator : SJ/MA
Sample : DFTPP55
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP55

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Title : SVOA CALIBRATION
Last Update : Thu May 04 00:50:13 2017



AutoFind: Scans 2501, 2502, 2503; Background Corrected with Scan 2493

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.8	58261	PASS
68	69	0.00	2	0.8	481	PASS
69	198	0.00	100	38.1	60386	PASS
70	69	0.00	2	0.5	286	PASS
127	198	10	80	57.9	91650	PASS
197	198	0.00	2	0.5	715	PASS
198	198	100	100	100.0	158330	PASS
199	198	5	9	6.9	10988	PASS
275	198	10	60	24.1	38098	PASS
365	198	1	100	3.4	5372	PASS
441	442	0.01	100	17.1	18492	PASS
442	198	50	100	68.2	107954	PASS
443	442	15	24	19.8	21378	PASS

m/z	Abundance
39.90	502.0
43.10	166.0
44.00	860.0
44.90	220.0
49.00	324.0
55.00	174.0
56.90	317.0
70.90	157.0
79.00	175.0
84.00	368.0
85.90	172.0
95.90	241.0
108.90	170.0
121.30	170.0
151.90	167.0
207.00	325.0
400.00	188.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

m/z	Abundance
36.90	224.0
37.90	719.0
39.00	4569.0
39.90	795.0
41.00	321.0
43.10	411.0
44.00	1898.0
44.90	237.0
46.80	248.0
49.00	816.0
50.00	18928.0
51.00	75528.0
52.00	3833.0
53.10	168.0
55.00	574.0
56.00	2695.0
57.00	6638.0
58.00	255.0
60.00	154.0
61.00	1019.0
62.00	1079.0
63.00	4187.0
64.00	353.0
65.00	1866.0
67.00	275.0
67.90	1445.0
68.90	77432.0
70.00	257.0
71.10	354.0
72.90	893.0
74.00	8368.0
75.00	12342.0
76.00	3913.0
77.00	96424.0
78.00	6063.0
79.00	5989.0
79.90	3704.0
81.00	6244.0
81.90	1597.0
83.00	2008.0
83.90	432.0
85.00	945.0
86.00	1397.0
87.00	651.0
87.90	237.0
89.00	176.0
90.20	179.0
91.00	1998.0
92.00	1155.0
93.00	9677.0
93.90	1065.0
96.00	664.0
97.10	200.0
98.00	7255.0
99.00	6301.0
100.00	688.0
101.00	3554.0
101.90	348.0
102.90	883.0
104.00	3045.0
105.00	1782.0
106.00	682.0
107.00	26552.0
108.00	4774.0
109.90	54408.0
111.00	7757.0
111.90	852.0
112.90	292.0
114.00	213.0
114.90	179.0
116.00	1832.0
117.00	20072.0
118.00	1379.0
119.00	246.0
120.00	362.0
121.20	178.0
122.00	2064.0
123.00	2782.0
123.90	1210.0
124.90	1436.0
127.00	109896.0
128.00	8523.0
129.00	38928.0
130.00	2794.0
131.00	610.0
131.90	487.0
133.00	236.0
133.90	1030.0
135.00	3347.0
136.00	1279.0
137.00	2430.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

137.90	360.0
139.90	331.0
140.90	5308.0
141.90	2067.0
142.90	1241.0
144.00	558.0
145.00	419.0
145.90	856.0
147.00	3028.0
148.00	4877.0
149.00	1340.0
149.90	328.0
151.00	742.0
151.40	637.0
151.90	506.0
152.90	1198.0
153.90	1511.0
155.00	2992.0
156.00	3895.0
157.00	966.0
158.00	927.0
159.00	664.0
160.00	1300.0
161.00	1828.0
161.70	510.0
161.90	564.0
162.80	201.0
164.00	152.0
165.00	1615.0
166.00	1424.0
167.00	8420.0
168.00	4532.0
169.00	830.0
169.90	178.0
171.00	559.0
171.90	837.0
173.00	1078.0
174.00	2248.0
175.00	3535.0
176.00	1115.0
177.00	1958.0
178.00	601.0
178.90	6551.0
180.00	4875.0
180.90	2141.0
181.90	507.0
182.70	228.0
184.00	593.0
184.90	3161.0
186.00	24864.0
187.00	7261.0
188.00	697.0
188.90	1468.0
190.00	245.0
191.00	967.0
192.00	2131.0
193.00	2265.0
194.10	656.0
194.90	854.0
196.00	6363.0
196.70	2147.0
197.90	180288.0
198.90	13315.0
200.00	1315.0
201.40	622.0
203.00	1459.0
204.00	6216.0
205.00	10139.0
206.00	49088.0
207.00	6606.0
208.00	1514.0
208.80	589.0
211.00	1542.0
211.70	425.0
213.10	181.0
215.00	420.0
216.00	1166.0
216.90	11601.0
217.90	1424.0
218.90	162.0
220.90	6628.0
221.80	1469.0
222.90	2744.0
224.00	23400.0
225.00	5983.0
226.00	814.0
227.00	10126.0
227.90	1279.0
228.90	1510.0
230.10	161.0
230.90	1196.0
232.00	316.0
233.00	222.0
233.90	580.0
235.00	881.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

236.00	283.0
236.90	757.0
238.90	271.0
239.90	306.0
241.00	584.0
242.00	1043.0
243.00	1874.0
244.00	17064.0
245.00	3395.0
246.00	2899.0
246.90	646.0
247.90	312.0
249.00	646.0
249.90	250.0
251.00	297.0
251.90	332.0
252.90	460.0
255.00	94784.0
256.00	15003.0
257.00	1212.0
258.00	5878.0
258.90	581.0
259.90	172.0
260.80	156.0
263.90	261.0
265.00	2355.0
265.90	279.0
269.90	371.0
270.90	373.0
271.80	645.0
273.00	3061.0
274.00	7005.0
275.00	41968.0
275.90	5175.0
277.00	3336.0
278.00	696.0
283.00	294.0
284.00	333.0
285.00	820.0
288.10	211.0
288.90	243.0
289.70	249.0
291.90	206.0
293.00	1033.0
294.00	396.0
296.00	11225.0
297.00	1172.0
298.00	193.0
301.90	160.0
303.00	1329.0
304.00	331.0
310.10	269.0
314.00	788.0
314.90	1127.0
315.90	1131.0
317.00	212.0
321.10	345.0
322.10	194.0
323.00	3498.0
324.00	868.0
326.90	668.0
328.00	477.0
329.00	182.0
331.90	177.0
333.00	614.0
334.00	2828.0
335.00	553.0
336.10	208.0
341.00	243.0
342.00	279.0
345.90	772.0
346.80	232.0
351.00	169.0
352.00	1188.0
353.00	990.0
354.00	1028.0
364.90	5131.0
366.00	826.0
371.00	278.0
372.00	2031.0
373.00	442.0
383.00	676.0
390.80	180.0
391.90	202.0
402.00	699.0
403.00	1025.0
404.00	300.0
405.00	165.0
420.90	918.0
422.00	748.0
423.00	5146.0
424.00	1475.0
432.90	212.0
437.10	178.0
438.10	202.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

439.30	191.0
439.90	269.0
441.00	16712.0
442.00	98512.0
443.00	20192.0
444.00	1833.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

m/z	Abundance
38.10	584.0
39.00	4328.0
40.00	1303.0
41.00	422.0
42.80	243.0
44.00	1666.0
45.00	354.0
49.00	847.0
50.00	16008.0
51.00	67344.0
52.00	3315.0
53.00	322.0
54.90	310.0
55.90	1849.0
57.00	6287.0
57.90	163.0
61.00	1079.0
61.90	1107.0
62.90	2432.0
64.00	528.0
65.00	1780.0
66.10	163.0
68.90	66776.0
70.00	602.0
70.90	288.0
73.00	690.0
74.00	6776.0
75.00	11108.0
76.00	3905.0
77.00	82704.0
78.00	6040.0
79.00	5844.0
80.00	3918.0
81.00	6658.0
82.00	1283.0
82.90	1342.0
83.90	522.0
85.00	1181.0
85.90	1564.0
87.00	915.0
88.10	449.0
88.90	246.0
90.90	1584.0
92.00	1274.0
93.00	8770.0
94.00	702.0
94.90	195.0
96.00	758.0
97.00	328.0
98.00	6645.0
99.00	5450.0
100.00	558.0
101.00	4075.0
102.00	335.0
103.00	1419.0
103.90	2356.0
104.90	2371.0
107.00	26432.0
108.00	4045.0
109.90	51320.0
111.00	7675.0
112.00	758.0
113.00	542.0
116.00	1821.0
117.00	19168.0
118.00	1518.0
118.90	320.0
119.90	501.0
121.90	2030.0
123.00	2548.0
123.90	1023.0
125.00	1425.0
127.00	107600.0
128.00	7584.0
128.90	39984.0
130.00	3465.0
131.00	802.0
131.90	443.0
133.90	1071.0
135.00	3501.0
136.00	1049.0
136.90	1821.0
138.00	331.0
139.00	192.0
140.10	301.0
141.00	4424.0
141.90	1642.0
142.90	1010.0
143.90	321.0
144.80	154.0
145.90	784.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

146.90	2115.0
148.00	4731.0
148.90	1362.0
149.90	162.0
151.10	393.0
151.80	535.0
152.90	1616.0
154.00	1212.0
155.00	2793.0
156.00	3993.0
157.00	737.0
157.90	974.0
158.90	551.0
159.90	1283.0
161.00	2253.0
161.90	771.0
162.90	360.0
164.00	242.0
164.90	1816.0
166.00	1944.0
166.90	8116.0
168.00	4717.0
168.90	738.0
170.20	210.0
170.80	334.0
171.90	625.0
173.00	1024.0
173.90	1718.0
175.00	3428.0
176.10	1025.0
177.00	1935.0
178.00	702.0
178.90	6801.0
180.00	5371.0
181.00	1894.0
181.80	371.0
183.00	285.0
184.00	499.0
185.00	3696.0
186.00	24816.0
187.00	7511.0
188.00	1079.0
188.90	1398.0
190.10	319.0
191.00	950.0
192.00	1989.0
193.00	2946.0
194.00	487.0
194.70	390.0
196.00	5625.0
197.90	186048.0
198.90	12447.0
200.00	791.0
201.40	919.0
202.90	909.0
204.00	6528.0
205.00	12383.0
206.00	47160.0
207.00	7295.0
208.00	1490.0
209.00	613.0
210.20	553.0
211.00	1951.0
211.80	496.0
214.90	666.0
216.00	887.0
216.90	11875.0
217.90	1451.0
219.00	182.0
220.90	7550.0
222.00	1663.0
223.00	2756.0
224.00	25800.0
225.00	6881.0
226.00	1051.0
226.90	9535.0
228.00	1347.0
229.00	1916.0
229.90	217.0
231.00	898.0
233.00	362.0
233.90	545.0
235.00	771.0
236.00	970.0
237.00	776.0
239.00	419.0
240.10	319.0
240.90	516.0
241.90	1583.0
243.00	1992.0
244.00	20416.0
245.00	2299.0
246.00	3962.0
247.00	1025.0
248.00	337.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

248.90	498.0
250.90	314.0
251.80	339.0
253.00	727.0
255.00	102896.0
256.00	14979.0
256.90	1173.0
258.00	5437.0
258.90	916.0
260.90	213.0
263.80	170.0
265.00	1653.0
266.00	289.0
268.90	172.0
271.10	462.0
272.00	414.0
273.00	3328.0
273.90	8437.0
275.00	45792.0
276.00	6207.0
277.00	4085.0
277.90	387.0
283.10	332.0
283.90	288.0
285.00	968.0
286.10	152.0
288.70	177.0
291.10	238.0
292.00	425.0
293.00	873.0
293.90	377.0
296.00	12547.0
297.00	2175.0
298.00	160.0
301.00	307.0
302.10	178.0
303.00	1020.0
303.90	414.0
304.90	179.0
307.90	348.0
308.90	163.0
309.90	256.0
314.00	672.0
314.90	1503.0
315.90	704.0
318.20	159.0
321.00	458.0
321.90	196.0
323.00	4126.0
324.00	294.0
327.00	764.0
328.00	246.0
331.70	158.0
333.00	721.0
334.00	3229.0
334.90	458.0
335.90	275.0
338.70	163.0
339.90	188.0
341.00	291.0
341.90	192.0
346.00	605.0
346.90	192.0
352.00	1226.0
353.00	1170.0
354.00	1557.0
355.00	334.0
363.90	263.0
365.00	5805.0
366.00	841.0
370.00	160.0
370.90	237.0
372.00	2266.0
373.00	322.0
382.90	779.0
383.90	161.0
390.00	322.0
391.80	169.0
400.90	221.0
401.90	807.0
403.00	1142.0
404.00	364.0
420.90	1067.0
422.00	1129.0
423.00	7660.0
424.00	1731.0
434.30	166.0
436.90	183.0
437.40	326.0
437.90	430.0
438.80	237.0
439.50	301.0
441.00	22816.0
442.00	127960.0
443.00	25032.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

444.002333.0

453.30181.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

m/z	Abundance
37.90	222.0
39.00	2265.0
40.00	682.0
43.10	308.0
44.00	1986.0
49.00	684.0
50.00	7629.0
51.00	31912.0
52.00	1405.0
54.80	267.0
56.00	1133.0
57.00	3117.0
58.00	240.0
60.90	516.0
62.00	430.0
63.00	1687.0
64.00	311.0
65.00	631.0
68.90	36952.0
71.00	229.0
73.10	382.0
74.00	3545.0
74.90	5682.0
76.00	2365.0
77.00	46064.0
78.00	2714.0
78.90	2546.0
79.90	1752.0
80.90	2803.0
81.90	1083.0
82.90	860.0
84.00	304.0
85.00	528.0
86.00	770.0
86.90	514.0
91.00	446.0
92.00	554.0
93.00	4560.0
93.90	315.0
96.00	241.0
97.10	173.0
97.90	3610.0
99.00	2877.0
100.00	335.0
100.90	2209.0
102.90	792.0
103.90	1096.0
104.90	1259.0
105.90	541.0
107.00	13527.0
108.00	2450.0
109.90	27472.0
111.00	3721.0
111.90	654.0
112.90	327.0
116.10	695.0
117.00	10072.0
118.00	592.0
120.00	296.0
122.00	1097.0
123.00	1428.0
123.90	854.0
124.90	887.0
127.00	57456.0
128.00	4569.0
128.90	20184.0
130.00	1461.0
130.90	601.0
132.10	203.0
133.90	823.0
135.00	1622.0
136.00	666.0
136.90	1563.0
138.40	182.0
140.00	335.0
140.90	2300.0
142.00	1009.0
142.90	740.0
144.00	355.0
144.90	539.0
146.00	567.0
147.00	1245.0
147.90	2547.0
149.00	677.0
149.90	198.0
151.20	327.0
152.00	208.0
153.00	1217.0
154.00	656.0
155.00	1761.0
156.00	2437.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

157.20	338.0
158.00	536.0
158.90	265.0
160.00	665.0
160.90	1441.0
162.00	334.0
163.80	276.0
164.90	1052.0
166.00	772.0
166.90	5115.0
168.00	3321.0
169.00	464.0
169.90	289.0
170.80	303.0
171.90	698.0
172.90	745.0
173.90	1072.0
175.00	2080.0
175.90	439.0
177.00	815.0
178.00	243.0
179.00	3582.0
179.90	2417.0
180.90	1441.0
182.00	304.0
183.00	545.0
184.00	493.0
185.00	2150.0
186.00	14282.0
187.00	3999.0
187.90	383.0
188.90	876.0
190.00	177.0
190.90	326.0
191.20	242.0
192.00	1374.0
193.00	1582.0
194.00	433.0
196.00	3076.0
197.90	108656.0
198.90	7204.0
200.00	665.0
201.50	700.0
202.20	165.0
203.00	822.0
204.00	4089.0
205.00	6548.0
206.00	29192.0
207.00	4567.0
207.90	781.0
208.80	296.0
209.90	648.0
211.00	881.0
211.90	257.0
213.10	212.0
214.80	279.0
215.90	754.0
216.90	7027.0
217.90	1012.0
219.80	250.0
221.00	3580.0
222.00	753.0
223.00	1642.0
224.00	15460.0
225.00	3205.0
226.00	408.0
227.00	6365.0
228.00	754.0
229.00	1477.0
230.20	197.0
230.90	734.0
232.90	194.0
234.00	420.0
234.90	473.0
236.00	394.0
237.00	488.0
238.90	358.0
239.90	252.0
240.90	611.0
242.00	1008.0
242.90	829.0
244.00	12941.0
245.00	1814.0
246.00	2102.0
246.90	470.0
248.00	181.0
248.90	886.0
251.00	206.0
252.10	322.0
252.90	335.0
255.00	63272.0
256.00	10303.0
256.90	584.0
258.00	3825.0
259.00	771.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55

261.10	178.0
265.00	1382.0
266.00	216.0
270.00	170.0
270.90	233.0
271.30	253.0
272.10	332.0
273.00	1777.0
274.00	4907.0
275.00	26536.0
276.00	4064.0
277.00	1827.0
278.00	376.0
280.90	351.0
282.10	288.0
283.00	225.0
284.00	204.0
285.00	541.0
286.00	172.0
291.90	206.0
292.80	521.0
294.00	186.0
295.00	301.0
295.90	8310.0
296.90	968.0
303.00	764.0
304.00	236.0
309.00	413.0
311.20	152.0
313.90	280.0
315.00	1031.0
316.00	580.0
321.00	540.0
321.90	260.0
323.00	2299.0
324.00	341.0
326.90	549.0
328.00	282.0
333.10	318.0
334.00	1809.0
334.90	435.0
341.00	450.0
345.90	573.0
351.90	1089.0
352.90	589.0
354.00	929.0
355.00	228.0
365.00	5180.0
366.00	474.0
370.20	161.0
371.00	173.0
372.00	1431.0
372.90	434.0
373.40	165.0
382.90	585.0
390.00	185.0
391.00	168.0
391.90	215.0
402.00	803.0
403.00	709.0
404.00	368.0
420.90	1052.0
422.00	891.0
423.00	5373.0
424.00	1102.0
431.20	163.0
434.60	175.0
435.30	150.0
437.90	313.0
439.30	266.0
441.00	15950.0
442.00	97392.0
443.00	18912.0
444.00	1334.0
444.90	181.0

Instrument :
BNA_G
ClientSampleId :
DFTPP55