

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050422\
 Data File : BG053425.D
 Acq On : 4 May 2022 10:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 18:07:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 18:06:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	117	0.00
2	1,4-Dioxane	0.557	0.614	-10.2	123	0.00
3	Pyridine	1.470	1.617	-10.0	116	0.00
4	n-Nitrosodimethylamine	0.728	0.848	-16.5	124	0.00
5 S	2-Fluorophenol	1.167	1.242	-6.4	117	0.00
6	Aniline	2.085	2.115	-1.4	109	0.00
7 S	Phenol-d6	1.799	1.894	-5.3	114	0.00
8	2-Chlorophenol	1.260	1.283	-1.8	110	0.00
9	Benzaldehyde	1.075	1.038	3.4	119	0.00
10 C	Phenol	1.791	1.862	-4.0	116	0.00
11	bis(2-Chloroethyl)ether	1.291	1.363	-5.6	118	0.00
12	1,3-Dichlorobenzene	1.464	1.514	-3.4	114	0.00
13 C	1,4-Dichlorobenzene	1.492	1.542	-3.4	114	0.00
14	1,2-Dichlorobenzene	1.439	1.433	0.4	112	0.00
15	Benzyl Alcohol	1.599	1.618	-1.2	109	0.00
16	2,2'-oxybis(1-Chloropropane	2.155	2.260	-4.9	113	0.00
17	2-Methylphenol	1.212	1.231	-1.6	113	0.00
18	Hexachloroethane	0.560	0.570	-1.8	110	0.00
19 P	n-Nitroso-di-n-propylamine	1.317	1.286	2.4	106	0.00
20	3+4-Methylphenols	1.711	1.751	-2.3	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
22	Acetophenone	0.555	0.589	-6.1	109	0.00
23 S	Nitrobenzene-d5	0.493	0.515	-4.5	108	0.00
24	Nitrobenzene	0.479	0.502	-4.8	111	0.00
25	Isophorone	0.837	0.866	-3.5	105	0.00
26 C	2-Nitrophenol	0.200	0.219	-9.5	111	0.00
27	2,4-Dimethylphenol	0.287	0.303	-5.6	107	0.00
28	bis(2-Chloroethoxy)methane	0.470	0.493	-4.9	107	0.00
29 C	2,4-Dichlorophenol	0.361	0.380	-5.3	106	0.00
30	1,2,4-Trichlorobenzene	0.408	0.421	-3.2	106	0.00
31	Naphthalene	1.067	1.123	-5.2	108	0.00
32	Benzoic acid	0.155	0.179	-15.5	96	0.00
33	4-Chloroaniline	0.457	0.465	-1.8	103	0.00
34 C	Hexachlorobutadiene	0.303	0.315	-4.0	107	0.00
35	Caprolactam	0.127	0.129	-1.6	102	0.00
36 C	4-Chloro-3-methylphenol	0.420	0.440	-4.8	106	0.00
37	2-Methylnaphthalene	0.790	0.828	-4.8	108	0.00
38	1-Methylnaphthalene	0.771	0.804	-4.3	107	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.679	0.694	-2.2	105	0.00
41 P	Hexachlorocyclopentadiene	0.351	0.322	8.3	91	0.00
42 S	2,4,6-Tribromophenol	0.318	0.317	0.3	100	0.00
43 C	2,4,6-Trichlorophenol	0.468	0.461	1.5	99	0.00
44	2,4,5-Trichlorophenol	0.499	0.490	1.8	99	0.00
45 S	2-Fluorobiphenyl	1.450	1.467	-1.2	105	0.00
46	1,1'-Biphenyl	1.440	1.471	-2.2	106	0.00
47	2-Chloronaphthalene	1.149	1.146	0.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050422\
 Data File : BG053425.D
 Acq On : 4 May 2022 10:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 18:07:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 18:06:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.436	0.448	-2.8	101	0.00
49	Acenaphthylene	1.799	1.854	-3.1	104	0.00
50	Dimethylphthalate	1.600	1.601	-0.1	102	0.00
51	2,6-Dinitrotoluene	0.334	0.332	0.6	102	0.00
52 C	Acenaphthene	1.220	1.235	-1.2	102	0.00
53	3-Nitroaniline	0.353	0.379	-7.4	107	0.00
54 P	2,4-Dinitrophenol	0.212	0.196	7.5	90	0.00
55	Dibenzofuran	1.910	1.917	-0.4	104	0.00
56 P	4-Nitrophenol	0.269	0.260	3.3	93	0.00
57	2,4-Dinitrotoluene	0.475	0.494	-4.0	104	0.00
58	Fluorene	1.543	1.546	-0.2	103	0.00
59	2,3,4,6-Tetrachlorophenol	0.485	0.471	2.9	97	0.00
60	Diethylphthalate	1.676	1.662	0.8	101	0.00
61	4-Chlorophenyl-phenylether	0.875	0.872	0.3	102	0.00
62	4-Nitroaniline	0.372	0.389	-4.6	107	0.00
63	Azobenzene	1.690	1.714	-1.4	104	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.140	0.146	-4.3	104	0.00
66 c	n-Nitrosodiphenylamine	0.575	0.585	-1.7	103	0.00
67	4-Bromophenyl-phenylether	0.256	0.259	-1.2	101	0.00
68	Hexachlorobenzene	0.278	0.278	0.0	101	0.00
69	Atrazine	0.225	0.220	2.2	100	0.00
70 C	Pentachlorophenol	0.152	0.150	1.3	93	0.00
71	Phenanthrene	1.092	1.119	-2.5	104	0.00
72	Anthracene	1.082	1.101	-1.8	103	0.00
73	Carbazole	1.052	1.088	-3.4	105	0.00
74	Di-n-butylphthalate	1.193	1.239	-3.9	106	0.00
75 C	Fluoranthene	1.400	1.441	-2.9	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	113	0.00
77	Benzydine	0.571	0.508	11.0	106	0.00
78	Pyrene	1.424	1.386	2.7	107	0.00
79 S	Terphenyl-d14	1.184	1.142	3.5	107	0.00
80	Butylbenzylphthalate	0.531	0.552	-4.0	113	0.00
81	Benzo(a)anthracene	1.367	1.401	-2.5	114	0.00
82	3,3'-Dichlorobenzidine	0.503	0.503	0.0	111	0.00
83	Chrysene	1.269	1.308	-3.1	115	0.00
84	Bis(2-ethylhexyl)phthalate	0.718	0.761	-6.0	118	0.00
85 c	Di-n-octyl phthalate	1.201	1.304	-8.6	121	0.00
86 I	Perylene-d12	1.000	1.000	0.0	120	0.00
87	Indeno(1,2,3-cd)pyrene	1.486	1.539	-3.6	120	0.00
88	Benzo(b)fluoranthene	1.262	1.284	-1.7	119	0.00
89	Benzo(k)fluoranthene	1.240	1.262	-1.8	119	0.00
90 C	Benzo(a)pyrene	1.075	1.116	-3.8	120	0.00
91	Dibenzo(a,h)anthracene	1.223	1.245	-1.8	119	0.00
92	Benzo(g,h,i)perylene	1.205	1.256	-4.2	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050422\
Data File : BG053425.D
Acq On : 4 May 2022 10:55
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 04 18:07:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG040822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 18:06:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0