



Data Path : Z:\HPCHEM1\BNA_G\Data\BG050515\
Data File : BG016859.D
Acq On : 5 May 2015 10:36
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 06 07:09:23 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 05 16:13:58 2015
Response via : Initial Calibration



TIC: BG016859.D

(69) Pentachlorophenol (C)



16.836min (+0.004) 0.00ng

response 37018

Ion	Exp%	Act%
265.70	100	100
268.00	69.00	66.47
264.00	64.20	66.53
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG050515\
Data File : BG016859.D
Acq On : 5 May 2015 10:36
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 06 07:09:23 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 05 16:13:58 2015
Response via : Initial Calibration



TIC: BG016859.D

(76) Benzidine

19.439min (+0.004) 0.00ng

response 271911

Ion	Exp%	Act%
184.00	100	100
185.00	14.70	13.99
183.00	13.20	13.01
0.00	0.00	0.00