

Data Path : Z:\HPCHEM1\BNA G\DATA\BG050515\
 Data File : BG016867.D
 Acq On : 5 May 2015 16:11
 Operator : TP/IZ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG050515

Quant Time: May 05 17:06:17 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 05 16:13:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.481	0.472	1.9	108	0.00
3	Pyridine	1.488	1.484	0.3	106	0.00
4	n-Nitrosodimethylamine	0.888	0.886	0.2	112	0.00
5 S	2-Fluorophenol	1.149	1.128	1.8	107	0.00
6	Aniline	2.304	2.286	0.8	108	0.00
7 S	Phenol-d6	1.641	1.608	2.0	107	0.00
8	2-Chlorophenol	1.331	1.291	3.0	107	0.00
9	Benzaldehyde	1.144	1.170	-2.3	108	0.00
10 C	Phenol	1.760	1.748	0.7	109	0.00
11	bis(2-Chloroethyl)ether	1.361	1.335	1.9	107	0.00
12	1,3-Dichlorobenzene	1.467	1.427	2.7	109	0.00
13 C	1,4-Dichlorobenzene	1.486	1.420	4.4	104	0.00
14	1,2-Dichlorobenzene	1.426	1.414	0.8	109	0.00
15	Benzyl Alcohol	1.499	1.488	0.7	109	0.00
16	2,2'-oxybis(1-Chloropropane	2.519	2.418	4.0	107	0.00
17	2-Methylphenol	1.239	1.235	0.3	109	0.00
18	Hexachloroethane	0.611	0.588	3.8	106	0.00
19 P	n-Nitroso-di-n-propylamine	1.439	1.424	1.0	110	0.00
20	3+4-Methylphenols	1.708	1.694	0.8	108	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
22	Acetophenone	0.476	0.488	-2.5	109	0.00
23 S	Nitrobenzene-d5	0.409	0.422	-3.2	111	0.00
24	Nitrobenzene	0.413	0.417	-1.0	108	0.00
25	Isophorone	0.827	0.844	-2.1	110	0.00
26 C	2-Nitrophenol	0.168	0.176	-4.8	111	0.00
27	2,4-Dimethylphenol	0.305	0.316	-3.6	109	0.00
28	bis(2-Chloroethoxy)methane	0.417	0.423	-1.4	112	0.00
29 C	2,4-Dichlorophenol	0.288	0.289	-0.3	107	0.00
30	1,2,4-Trichlorobenzene	0.305	0.311	-2.0	109	0.00
31	Naphthalene	0.995	0.990	0.5	108	0.00
32	Benzoic acid	0.179	0.198	-10.6	120	0.00
33	4-Chloroaniline	0.459	0.475	-3.5	109	0.00
34 C	Hexachlorobutadiene	0.191	0.192	-0.5	110	0.00
35	Caprolactam	0.150	0.155	-3.3	111	0.00
36 C	4-Chloro-3-methylphenol	0.369	0.378	-2.4	110	0.00
37	2-Methylnaphthalene	0.758	0.762	-0.5	110	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
39	1,2,4,5-Tetrachlorobenzene	0.541	0.552	-2.0	108	0.00
40 P	Hexachlorocyclopentadiene	0.349	0.364	-4.3	111	0.00
41 S	2,4,6-Tribromophenol	0.201	0.211	-5.0	112	0.00
42 C	2,4,6-Trichlorophenol	0.383	0.402	-5.0	106	0.00
43	2,4,5-Trichlorophenol	0.407	0.424	-4.2	108	0.00
44 S	2-Fluorobiphenyl	1.325	1.370	-3.4	109	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG050515\
 Data File : BG016867.D
 Acq On : 5 May 2015 16:11
 Operator : TP/IZ
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 ICVBG050515

Quant Time: May 05 17:06:17 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 05 16:13:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.432	1.490	-4.1	110	0.00
46	2-Chloronaphthalene	1.134	1.168	-3.0	109	0.00
47	2-Nitroaniline	0.395	0.434	-9.9	114	0.00
48	Acenaphthylene	1.975	2.029	-2.7	108	0.00
49	Dimethylphthalate	1.493	1.572	-5.3	112	0.00
50	2,6-Dinitrotoluene	0.310	0.334	-7.7	110	0.00
51 C	Acenaphthene	1.175	1.192	-1.4	108	0.00
52	3-Nitroaniline	0.353	0.380	-7.6	108	0.00
53 P	2,4-Dinitrophenol	0.142	0.154	-8.5	120	0.00
54	Dibenzofuran	1.669	1.742	-4.4	109	0.00
55 P	4-Nitrophenol	0.299	0.325	-8.7	113	0.00
56	2,4-Dinitrotoluene	0.397	0.452	-13.9	117	0.00
57	Fluorene	1.478	1.555	-5.2	111	0.00
58	2,3,4,6-Tetrachlorophenol	0.351	0.375	-6.8	108	0.00
59	Diethylphthalate	1.577	1.644	-4.2	111	0.00
60	4-Chlorophenyl-phenylether	0.742	0.768	-3.5	112	0.00
61	4-Nitroaniline	0.409	0.441	-7.8	112	0.00
62	Azobenzene	1.633	1.694	-3.7	110	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	100	0.00
64	4,6-Dinitro-2-methylphenol	0.099	0.107	-8.1	115	0.00
65 c	n-Nitrosodiphenylamine	0.611	0.626	-2.5	109	0.00
66	4-Bromophenyl-phenylether	0.201	0.209	-4.0	111	0.00
67	Hexachlorobenzene	0.212	0.219	-3.3	112	0.00
68	Atrazine	0.220	0.228	-3.6	108	0.00
69 C	Pentachlorophenol	0.137	0.147	-7.3	110	0.00
70	Phenanthrene	1.029	1.062	-3.2	109	0.00
71	Anthracene	1.038	1.075	-3.6	110	0.00
72	Carbazole	0.987	1.021	-3.4	108	0.00
73	Di-n-butylphthalate	1.280	1.318	-3.0	108	0.00
74 C	Fluoranthene	1.194	1.244	-4.2	108	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
76	Benzidine	0.680	0.699	-2.8	100	0.00
77	Pyrene	1.180	1.239	-5.0	108	0.00
78 S	Terphenyl-d14	0.853	0.899	-5.4	106	0.00
79	Butylbenzylphthalate	0.561	0.588	-4.8	106	0.00
80	Benzo(a)anthracene	1.073	1.119	-4.3	106	0.00
81	3,3'-Dichlorobenzidine	0.419	0.434	-3.6	105	0.00
82	Chrysene	1.005	1.050	-4.5	106	0.00
83	Bis(2-ethylhexyl)phthalate	0.767	0.811	-5.7	106	0.00
84 c	Di-n-octyl phthalate	1.289	1.366	-6.0	106	0.00
85	Indeno(1,2,3-cd)pyrene	1.198	1.243	-3.8	109	0.02
86 I	Perylene-d12	1.000	1.000	0.0	97	0.00
87	Benzo(b)fluoranthene	1.114	1.154	-3.6	108	0.00

Data Path : Z:\HPCHEM1\BNA G\DATA\BG050515\
Data File : BG016867.D
Acq On : 5 May 2015 16:11
Operator : TP/IZ
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
ICVBG050515

Quant Time: May 05 17:06:17 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 05 16:13:58 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.072	1.080	-0.7	105	0.00
89 C	Benzo(a)pyrene	1.039	1.054	-1.4	105	0.00
90	Dibenzo(a,h)anthracene	1.061	1.074	-1.2	107	0.01
91	Benzo(a,h,i)perylene	1.011	1.017	-0.6	109	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0