


Data Path : Z:\HPCHEM1\BNA G\DATA\BG050515\
 Data File : BG016877.D
 Acq On : 5 May 2015 23:10
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2542, 2543, 2544; Background Corrected with Scan 2536

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	54.7	38662	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	54.6	38629	PASS
70	69	0.00	2	0.7	275	PASS
127	198	10	80	57.1	40376	PASS
197	198	0.00	2	0.7	484	PASS
198	198	100	100	100.0	70736	PASS
199	198	5	9	6.4	4497	PASS
275	198	10	60	27.7	19610	PASS
365	198	1	100	4.2	2946	PASS
441	442	0.01	24	14.6	6723	PASS
442	198	50	100	65.0	45960	PASS
443	442	15	24	21.2	9754	PASS