

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050522\
 Data File : BG053441.D
 Acq On : 5 May 2022 00:04
 Operator : CG/JU
 Sample : PB144564BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB144564BL

Quant Time: May 05 02:23:07 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 02:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.099	152	23353	20.000	ng	0.00
21) Naphthalene-d8	10.907	136	96173	20.000	ng	0.00
39) Acenaphthene-d10	14.720	164	73476	20.000	ng	0.00
64) Phenanthrene-d10	17.463	188	197902	20.000	ng	0.00
76) Chrysene-d12	21.745	240	226597	20.000	ng	0.00
86) Perylene-d12	25.035	264	225267	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.667	112	196019	142.453	ng	0.00
7) Phenol-d6	7.259	99	285532	136.727	ng	0.00
23) Nitrobenzene-d5	9.262	82	197597	87.461	ng	0.00
42) 2,4,6-Tribromophenol	16.206	330	145107	133.901	ng	0.00
45) 2-Fluorobiphenyl	13.345	172	456571	90.802	ng	0.00
79) Terphenyl-d14	20.065	244	1094689	89.976	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050522\
Data File : BG053441.D
Acq On : 5 May 2022 00:04
Operator : CG/JU
Sample : PB144564BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB144564BL

Quant Time: May 05 02:23:07 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 02:13:53 2022
Response via : Initial Calibration

