


Data Path : Z:\HPCHEM1\BNA G\Data\BG050615\
 Data File : BG016893.D
 Acq On : 6 May 2015 11:19
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2542, 2543, 2544; Background Corrected with Scan 2535

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	50.0	44019	PASS
68	69	0.00	2	1.0	447	PASS
69	198	0.00	100	49.9	43960	PASS
70	69	0.00	2	0.1	51	PASS
127	198	10	80	54.8	48261	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	88125	PASS
199	198	5	9	6.6	5794	PASS
275	198	10	60	27.3	24080	PASS
365	198	1	100	4.2	3734	PASS
441	442	0.01	24	15.4	8790	PASS
442	198	50	100	64.9	57213	PASS
443	442	15	24	19.6	11206	PASS