


Data Path : Z:\HPCHEM1\BNA G\Data\BG050715\
 Data File : BG016931.D
 Acq On : 7 May 2015 21:14
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2541, 2542, 2543; Background Corrected with Scan 2535

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	55.3	42046	PASS
68	69	0.00	2	0.5	214	PASS
69	198	0.00	100	53.7	40837	PASS
70	69	0.00	2	0.2	91	PASS
127	198	10	80	56.6	43024	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	75994	PASS
199	198	5	9	7.1	5376	PASS
275	198	10	60	26.7	20319	PASS
365	198	1	100	4.9	3716	PASS
441	442	0.01	24	15.1	8155	PASS
442	198	50	100	71.0	53946	PASS
443	442	15	24	19.0	10242	PASS