


Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050724\
Data File : BG061128.D
Acq On : 7 May 2024 11:15
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed May 01 05:11:25 2024



AutoFind: Scans 2513, 2514, 2515; Background Corrected with Scan 2506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.9	22501	PASS
68	69	0.00	2	0.9	152	PASS
69	198	0.00	100	31.6	16992	PASS
70	69	0.00	2	0.3	51	PASS
127	198	10	80	38.4	20633	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	53741	PASS
199	198	5	9	7.2	3846	PASS
275	198	10	60	34.5	18527	PASS
365	198	1	100	6.1	3274	PASS
441	198	0.01	100	20.2	10851	PASS
442	442	50	100	100.0	65821	PASS
443	442	15	24	19.6	12896	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
5/7/2024	BNA_G	<u>BG061128.D</u>
Compound Name	Response	Retention Time
DDT	206704	20.79
DDD	8012	20.35
DDE	541	19.85
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8553	215257	3.97

Instrument :
BNA_G
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050724\
Data File : BG061128.D
Acq On : 7 May 2024 11:15
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 07 12:14:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061128.D\data.ms

(70) Pentachlorophenol (C)

16.977min (-0.009) 12438.97 ng m




response 50309

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.80	67.05
264.00	64.40	66.62
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050724\
Data File : BG061128.D
Acq On : 7 May 2024 11:15
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 07 12:14:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061128.D\data.ms

(77) Benzidine

19.556min (-0.009) 108931.01 ng

response 319144

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.47
183.00	13.60	12.71
0.00	0.00	0.00