


Data Path : Z:\HPCHEM1\BNA\_G\Data\BG050815\  
 Data File : BG017029.D  
 Acq On : 10 May 2015 18:22  
 Operator : TP/IZ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_G  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_G\METHODS\8270-BG050515.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2533

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 52.5      | 28546   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 69          | 198          | 0.00         | 100          | 48.4      | 26317   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.5       | 119     | PASS             |
| 127         | 198          | 10           | 80           | 55.0      | 29941   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.2       | 136     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 54405   | PASS             |
| 199         | 198          | 5            | 9            | 6.6       | 3601    | PASS             |
| 275         | 198          | 10           | 60           | 29.6      | 16131   | PASS             |
| 365         | 198          | 1            | 100          | 5.3       | 2869    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.7      | 6862    | PASS             |
| 442         | 198          | 50           | 100          | 80.3      | 43666   | PASS             |
| 443         | 442          | 15           | 24           | 18.5      | 8081    | PASS             |