

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050815\
 Data File : BG017020.D
 Acq On : 10 May 2015 11:25
 Operator : TP/IZ
 Sample : G2171-04
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DCW-WW-134-5615

Manual Integrations
 APPROVED

apatel
 5/12/2015 4:01:55 PM

Quant Time: May 11 07:49:25 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 05 16:13:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	58358	20.00	ng	-0.02
21) Naphthalene-d8	10.58	136	271140	20.00	ng	-0.02
38) Acenaphthene-d10	14.42	164	184009	20.00	ng	-0.02
63) Phenanthrene-d10	17.17	188	384677	20.00	ng	-0.02
75) Chrysene-d12	21.34	240	380496	20.00	ng	-0.02
86) Perylene-d12	23.64	264	370753	20.00	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	238099	71.04	ng	0.00
7) Phenol-d6	6.97	99	210606	43.98	ng	0.00
23) Nitrobenzene-d5	8.95	82	507939	91.52	ng	-0.02
41) 2,4,6-Tribromophenol	15.92	330	311081	168.40	ng	-0.01
44) 2-Fluorobiphenyl	13.05	172	1082538	88.78	ng	-0.02
78) Terphenyl-d14	19.79	244	1528733	94.19	ng	-0.02
Target Compounds						Qvalue
3) Pyridine	3.69	79	9050	2.08	ng	# 88
13) 1,4-Dichlorobenzene	7.83	146	26401	6.09	ng	91
14) 1,2-Dichlorobenzene	8.14	146	14132	3.40	ng	96
17) 2-Methylphenol	8.24	107	13134	3.63	ng	90
22) Acetophenone	8.60	105	20993	3.25	ng	# 92
27) 2,4-Dimethylphenol	9.78	122	16560m	4.01	ng	
30) 1,2,4-Trichlorobenzene	10.45	180	148212	35.82	ng	88
31) Naphthalene	10.64	128	2341064	173.59	ng	# 93
33) 4-Chloroaniline	10.69	127	42986	6.90	ng	95
37) 2-Methylnaphthalene	12.24	142	287713	28.01	ng	# 95
51) Acenaphthene	14.49	154	35321	3.27	ng	96
56) 2,4-Dinitrotoluene	14.79	165	32606	8.93	ng	# 96
57) Fluorene	15.48	166	42094	3.10	ng	97
59) Diethylphthalate	15.25	149	36872	2.54	ng	98
73) Di-n-butylphthalate	18.14	149	55716	2.26	ng	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050815\
Data File : BG017020.D
Acq On : 10 May 2015 11:25
Operator : TP/IZ
Sample : G2171-04
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DCW-WW-134-5615

Manual Integrations
APPROVED

apatel
5/12/2015 4:01:55 PM

Quant Time: May 11 07:49:25 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 05 16:13:58 2015
Response via : Initial Calibration

