

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050815\
 Data File : BG016980.D
 Acq On : 9 May 2015 7:04
 Operator : TP/IZ
 Sample : G2059-03MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Manual Integrations
 APPROVED

apatel
 5/12/2015 8:30:17 AM

Quant Time: May 11 01:49:04 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 05 16:13:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	57371	20.00	ng	-0.02
21) Naphthalene-d8	10.58	136	265260	20.00	ng	-0.01
38) Acenaphthene-d10	14.43	164	180460	20.00	ng	-0.01
63) Phenanthrene-d10	17.17	188	408057	20.00	ng	-0.02
75) Chrysene-d12	21.35	240	404247	20.00	ng	-0.02
86) Perylene-d12	23.65	264	394441	20.00	ng	-0.03

System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	249550	75.74	ng	-0.01
7) Phenol-d6	6.96	99	232547	49.40	ng	0.00
23) Nitrobenzene-d5	8.95	82	502684	92.58	ng	-0.01
41) 2,4,6-Tribromophenol	15.92	330	309119	170.63	ng	-0.01
44) 2-Fluorobiphenyl	13.05	172	1060578	88.69	ng	-0.02
78) Terphenyl-d14	19.80	244	1629275	94.48	ng	-0.02

Target Compounds						Qvalue
2) 1,4-Dioxane	3.29	88	24111	17.46	ng	# 1
3) Pyridine	3.68	79	55576	13.02	ng	99
4) n-Nitrosodimethylamine	3.60	42	49681	19.50	ng	93
6) Aniline	7.12	93	113939	17.24	ng	95
8) 2-Chlorophenol	7.36	128	161861	42.40	ng	92
9) Benzaldehyde	6.93	77	16346	4.15	ng	97
10) Phenol	6.99	94	85816	17.00	ng	99
11) bis(2-Chloroethyl)ether	7.22	93	183221	46.94	ng	98
12) 1,3-Dichlorobenzene	7.68	146	163569	38.86	ng	97
13) 1,4-Dichlorobenzene	7.82	146	175225	41.10	ng	94
14) 1,2-Dichlorobenzene	8.14	146	168830	41.26	ng	95
15) Benzyl Alcohol	8.03	79	136508	31.75	ng	90
16) 2,2'-oxybis(1-Chloropropan	8.33	45	313015	43.32	ng	98
17) 2-Methylphenol	8.23	107	126372	35.55	ng	98
18) Hexachloroethane	8.87	117	67127	38.30	ng	99
19) n-Nitroso-di-n-propylamine	8.59	70	183510	44.46	ng	92
20) 3+4-Methylphenols	8.56	107	158014	32.26	ng	89
22) Acetophenone	8.61	105	306681	48.54	ng	97
24) Nitrobenzene	8.99	77	255852	46.66	ng	98
25) Isophorone	9.52	82	497237	45.35	ng	99
26) 2-Nitrophenol	9.70	139	109758	49.16	ng	92
27) 2,4-Dimethylphenol	9.76	122	180824	44.75	ng	98
28) bis(2-Chloroethoxy)methane	10.00	93	267437	48.32	ng	98
29) 2,4-Dichlorophenol	10.23	162	190570	49.92	ng	97
30) 1,2,4-Trichlorobenzene	10.45	180	178221	44.03	ng	97
31) Naphthalene	10.63	128	577340	43.76	ng	99
32) Benzoic acid	9.85	122	37950	19.47	ng	97
33) 4-Chloroaniline	10.74	127	101715	16.69	ng	95
34) Hexachlorobutadiene	10.92	225	99576	39.31	ng	97
35) Caprolactam	11.51	113	17313m	8.73	ng	
36) 4-Chloro-3-methylphenol	11.87	107	233848	47.84	ng	96
37) 2-Methylnaphthalene	12.25	142	478200	47.59	ng	98
39) 1,2,4,5-Tetrachlorobenzene	12.62	216	217354	44.56	ng	# 100
40) Hexachlorocyclopentadiene	12.60	237	253140	80.35	ng	95

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG050815\
 Data File : BG016980.D
 Acq On : 9 May 2015 7:04
 Operator : TP/IZ
 Sample : G2059-03MSD
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD


Manual Integrations
 APPROVED

apatel
 5/12/2015 8:30:17 AM

Quant Time: May 11 01:49:04 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 05 16:13:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	12.86	196	168265	48.67	ng	93
43) 2,4,5-Trichlorophenol	12.93	196	185231	50.48	ng	98
45) 1,1'-Biphenyl	13.26	154	619586	47.96	ng	100
46) 2-Chloronaphthalene	13.30	162	489892	47.87	ng	98
47) 2-Nitroaniline	13.51	65	210236	59.02	ng	100
48) Acenaphthylene	14.15	152	835926	46.90	ng	99
49) Dimethylphthalate	13.89	163	676097	50.17	ng	100
50) 2,6-Dinitrotoluene	14.01	165	159542	57.03	ng	94
51) Acenaphthene	14.50	154	504924	47.61	ng	98
52) 3-Nitroaniline	14.34	138	82939	26.06	ng	89
53) 2,4-Dinitrophenol	14.54	184	150867	117.90	ng	88
54) Dibenzofuran	14.83	168	766869	50.92	ng	95
55) 4-Nitrophenol	14.65	139	113543	42.07	ng	94
56) 2,4-Dinitrotoluene	14.80	165	222543	62.18	ng	99
57) Fluorene	15.48	166	667752	50.08	ng	99
58) 2,3,4,6-Tetrachlorophenol	15.06	232	165350	52.22	ng	# 77
59) Diethylphthalate	15.26	149	723142	50.84	ng	98
60) 4-Chlorophenyl-phenylether	15.47	204	335243	50.04	ng	99
61) 4-Nitroaniline	15.50	138	174875	47.39	ng	94
62) Azobenzene	15.77	77	765944	51.99	ng	96
64) 4,6-Dinitro-2-methylphenol	15.56	198	114307	56.50	ng	91
65) n-Nitrosodiphenylamine	15.69	169	596187	47.86	ng	96
66) 4-Bromophenyl-phenylether	16.37	248	205922	50.28	ng	98
67) Hexachlorobenzene	16.48	284	219906	50.91	ng	96
68) Atrazine	16.64	200	234867	52.39	ng	98
69) Pentachlorophenol	16.82	266	271890	96.96	ng	96
70) Phenanthrene	17.22	178	1017386	48.45	ng	99
71) Anthracene	17.31	178	1030833	48.66	ng	99
72) Carbazole	17.58	167	991545	49.23	ng	100
73) Di-n-butylphthalate	18.14	149	1267143	48.53	ng	99
74) Fluoranthene	19.23	202	1152245	47.29	ng	98
76) Benzidine	19.42	184	410340	29.87	ng	99
77) Pyrene	19.59	202	1179295	49.43	ng	99
79) Butylbenzylphthalate	20.49	149	564168	49.75	ng	99
80) Benzo(a)anthracene	21.33	228	1080357	49.82	ng	98
81) 3,3'-Dichlorobenzidine	21.27	252	221110	26.11	ng	100
82) Chrysene	21.39	228	1002685	49.36	ng	98
83) Bis(2-ethylhexyl)phthalate	21.27	149	771248	49.72	ng	97
84) Di-n-octyl phthalate	22.15	149	1286536	49.37	ng	# 100
85) Indeno(1,2,3-cd)pyrene	26.01	276	1239279	51.20	ng	# 100
87) Benzo(b)fluoranthene	22.95	252	1123590	51.13	ng	98
88) Benzo(k)fluoranthene	23.00	252	1028688	48.66	ng	100
89) Benzo(a)pyrene	23.55	252	1044127	50.96	ng	97
90) Dibenzo(a,h)anthracene	26.02	278	1061671	50.73	ng	97
91) Benzo(g,h,i)perylene	26.73	276	1009860	50.67	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 7.79 min Scan# 868

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :





MWHR3-14MSD

Manual Integrations
APPROVED

apatel

5/12/2015 8:30:17 AM

Tgt Ion:	152	Resp:	57371
Ion Ratio	Lower	Upper	
152	100		
150	142.7	121.0	181.4
115	64.9	50.5	75.7



#2

1,4-Dioxane

Concen: 17.46 ng




RT: 3.29 min Scan# 102

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	88	Resp:	24111
Ion Ratio	Lower	Upper	
88	100		
58	76.4	0.1	0.1#
43	41.2	1.3	1.9#





#3

Pyridine

Concen: 13.02 ng

RT: 3.68 min Scan# 169

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :





MWHR3-14MSD

Manual Integrations
APPROVED

apatel

5/12/2015 8:30:17 AM

Tgt Ion:	79	Resp:	55576
Ion	Ratio	Lower	Upper
79	100		
52	87.0	68.5	102.7
51	47.9	37.8	56.8



#4

n-Nitrosodimethylamine

Concen: 19.50 ng




RT: 3.60 min Scan# 155


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	42	Resp:	49681
Ion	Ratio	Lower	Upper
42	100		
74	88.0	76.2	114.2
44	6.5	4.9	7.3





#5

2-Fluorophenol

Concen: 75.74 ng

RT: 5.37 min Scan# 457

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G


ClientSampleId :

MWHR3-14MSD




Manual Integrations
APPROVED

apatel

5/12/2015 8:30:17 AM



Tgt Ion	Ratio	Resp	Lower	Upper
112	100	249550		
64	71.7	51.4	77.2	
63	37.6	28.1	42.1	



#6

Aniline

Concen: 17.24 ng




RT: 7.12 min Scan# 754




Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	113939		
66	40.4	33.8	50.6	
65	19.1	18.7	28.1	





#7

Phenol-d6

Concen: 49.40 ng

RT: 6.96 min Scan# 727

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	99	Resp:	232547
Ion Ratio	Lower	Upper	
99	100		
42	29.1	22.1	33.1
71	43.2	33.4	50.0

Instrument :





BNA_G

ClientSampleId :

MWHR3-14MSD

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#8

2-Chlorophenol

Concen: 42.40 ng


RT: 7.36 min Scan# 795


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	128	Resp:	161861
Ion Ratio	Lower	Upper	
128	100		
130	32.6	14.4	54.4
64	50.8	38.4	78.4





#9

Benzaldehyde

Concen: 4.15 ng

RT: 6.93 min Scan# 722

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :





MWHR3-14MSD

Manual Integrations
APPROVED

apatel

5/12/2015 8:30:17 AM

Tgt Ion:	77	Resp:	16346
Ion Ratio	Lower	Upper	
77	100		
105	87.7	71.6	111.6
106	89.2	71.5	111.5



#10

Phenol

Concen: 17.00 ng




RT: 6.99 min Scan# 732

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	94	Resp:	85816
Ion Ratio	Lower	Upper	
94	100		
65	32.8	14.3	54.3
66	46.8	26.9	66.9









#11
bis(2-Chloroethyl)ether
Concen: 46.94 ng
RT: 7.22 min Scan# 771
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion:	93	Resp:	183221
Ion Ratio	Lower	Upper	
93	100		
63	82.0	61.2	101.2
95	31.5	15.0	55.0




Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#12
1,3-Dichlorobenzene
Concen: 38.86 ng
RT: 7.68 min Scan# 850
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion:	146	Resp:	163569
Ion Ratio	Lower	Upper	
146	100		
148	66.7	51.9	77.9
75	37.8	28.2	42.2









#13
 1,4-Dichlorobenzene
 Concen: 41.10 ng
 RT: 7.82 min Scan# 874
 Delta R.T. -0.02 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	175225		
148	59.3	52.8	79.2	
111	40.7	33.9	50.9	




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#14
 1,2-Dichlorobenzene
 Concen: 41.26 ng
 RT: 8.14 min Scan# 928
 Delta R.T. -0.02 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
146	100	168830		
148	63.7	49.9	74.9	
111	46.6	32.1	48.1	





#15

Benzyl Alcohol

Concen: 31.75 ng

RT: 8.03 min Scan# 909

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G





ClientSampleId :

MWHR3-14MSD

Tgt Ion:	79	Resp:	136508
Ion	Ratio	Lower	Upper
79	100		
108	76.1	53.4	80.2
77	72.0	52.2	78.2

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#16

2,2'-oxybis(1-Chloropropane)

Concen: 43.32 ng




RT: 8.33 min Scan# 960


Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	45	Resp:	313015
Ion	Ratio	Lower	Upper
45	100		
77	10.8	0.0	31.8
79	9.2	0.0	29.1






#17
2-Methylphenol
Concen: 35.55 ng
RT: 8.23 min Scan# 944
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD




Tgt Ion	Ratio	Resp	Lower	Upper
107	100	126372		
108	108.4	85.9	128.9	
77	52.9	43.8	65.6	
79	52.3	39.8	59.6	

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BG
Ion 79.00 (78.70 to 79.70): BG






#18
Hexachloroethane
Concen: 38.30 ng
RT: 8.87 min Scan# 1052
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
117	100	67127		
119	95.4	74.8	112.2	
201	90.0	72.1	108.1	



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E









#19
n-Nitroso-di-n-propylamine
Concen: 44.46 ng
RT: 8.59 min Scan# 1005
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
70	100		183510		
42	68.8	61.5	92.3		
101	11.6	9.8	14.6		
130	22.5	20.2	30.4		




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#20
3+4-Methylphenols
Concen: 32.26 ng
RT: 8.56 min Scan# 999
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
107	100		158014		
108	81.3	76.6	116.6		
77	36.9	19.0	59.0		
79	31.2	15.0	55.0		









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.58 min Scan# 1344
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
136	100		
137	11.8	9.8	14.8
54	10.9	8.7	13.1
68	5.9	5.0	7.4




Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#22
Acetophenone
Concen: 48.54 ng
RT: 8.61 min Scan# 1008
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp Lower	Resp Upper
105	100		
71	3.9	2.8	4.2
51	38.7	29.4	44.2
120	28.9	22.2	33.4









#23
 Nitrobenzene-d5
 Concen: 92.58 ng
 RT: 8.95 min Scan# 1066
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	502684		
128	40.1	33.2	49.8	
54	58.5	48.6	72.8	




Manual Integrations
 APPROVED

apatel
 5/12/2015 8:30:17 AM



#24
 Nitrobenzene
 Concen: 46.66 ng
 RT: 8.99 min Scan# 1073
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
77	100	255852		
123	42.3	34.0	51.0	
65	16.3	11.2	16.8	





#25

Isophorone

Concen: 45.35 ng

RT: 9.52 min Scan# 1162

Delta R.T. -0.01 min

Lab File: BG016980.D


Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :




MWHR3-14MSD



Tgt Ion	Ratio	Resp	Lower	Upper
82	100	497237		
95	8.0	6.7	10.1	
138	17.8	14.7	22.1	

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#26

2-Nitrophenol

Concen: 49.16 ng




RT: 9.70 min Scan# 1193


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	109758		
109	37.1	27.4	41.2	
65	66.0	47.1	70.7	









#27
 2,4-Dimethylphenol
 Concen: 44.75 ng
 RT: 9.76 min Scan# 1204
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
122	100	180824		
107	123.5	97.6	146.4	
121	58.0	48.4	72.6	




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#28
 bis(2-Chloroethoxy)methane
 Concen: 48.32 ng
 RT: 10.00 min Scan# 1244
 Delta R.T. -0.02 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	267437		
95	31.8	24.8	37.2	
123	14.4	12.6	19.0	









#29
 2,4-Dichlorophenol
 Concen: 49.92 ng
 RT: 10.23 min Scan# 1284
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion:	162	Resp:	190570
Ion Ratio	Lower	Upper	
162	100		
164	65.9	42.9	82.9
98	39.8	18.4	58.4




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#30
 1,2,4-Trichlorobenzene
 Concen: 44.03 ng
 RT: 10.45 min Scan# 1321
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion:	180	Resp:	178221
Ion Ratio	Lower	Upper	
180	100		
182	94.7	73.8	110.8
145	29.7	25.4	38.0









#31
Naphthalene
Concen: 43.76 ng
RT: 10.63 min Scan# 1352
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
128	100	577340		
129	12.2		9.1	13.7
127	13.5		10.9	16.3




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#32
Benzoic acid
Concen: 19.47 ng
RT: 9.85 min Scan# 1219
Delta R.T. -0.04 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
122	100	37950		
105	140.1		123.2	163.2
77	109.2		93.0	133.0






#33
4-Chloroaniline
Concen: 16.69 ng
RT: 10.74 min Scan# 1371
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD




Tgt Ion	Ratio	Resp	Lower	Upper
127	100	101715		
129	31.5	25.8	38.6	
65	38.7	28.2	42.2	
92	22.7	15.6	23.4	

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM




Abundance Ion 127.00 (126.70 to 127.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 65.00 (64.70 to 65.70): BG
Ion 92.00 (91.70 to 92.70): BG






#34
Hexachlorobutadiene
Concen: 39.31 ng
RT: 10.92 min Scan# 1401
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
225	100	99576		
223	59.3	45.8	68.6	
227	60.7	50.7	76.1	



Abundance Ion 225.00 (224.70 to 225.70): E
Ion 223.00 (222.70 to 223.70): E
Ion 227.00 (226.70 to 227.70): E





#35

Caprolactam

Concen: 8.73 ng m

RT: 11.51 min Scan# 1502

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G





ClientSampleId :

MWHR3-14MSD

Tgt Ion:	113	Resp:	17313
Ion Ratio	Lower	Upper	
113	100		
55	115.8	137.1	177.1#
56	94.3	94.3	134.3

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#36

4-Chloro-3-methylphenol

Concen: 47.84 ng




RT: 11.87 min Scan# 1563

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	107	Resp:	233848
Ion Ratio	Lower	Upper	
107	100		
144	25.2	20.2	30.2
142	75.5	64.2	96.2









#37
 2-Methylnaphthalene
 Concen: 47.59 ng
 RT: 12.25 min Scan# 1627
 Delta R.T. -0.02 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
142	100	478200		
141	86.6	69.9	104.9	
115	40.2	29.8	44.8	




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 14.43 min Scan# 1999
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	180460		
162	102.3	74.0	111.0	
160	44.4	34.7	52.1	





#39

1,2,4,5-Tetrachlorobenzene

Concen: 44.56 ng

RT: 12.62 min Scan# 1690

Delta R.T. -0.01 min

Lab File: BG016980.D


Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :




MWHR3-14MSD



Tgt Ion	Ratio	Resp Lower	Upper
216	100		
214	80.9	0.0	0.0#
179	22.1	0.0	0.0#
108	26.4	0.0	0.0#

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#40

Hexachlorocyclopentadiene

Concen: 80.35 ng




RT: 12.60 min Scan# 1687


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp Lower	Upper
237	100		
235	60.9	45.5	85.5
272	12.7	0.0	33.7









#41
 2,4,6-Tribromophenol
 Concen: 170.63 ng
 RT: 15.92 min Scan# 2252
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
330	100		
332	95.8	0.0	0.0#
141	49.7	0.0	0.0#




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#42
 2,4,6-Trichlorophenol
 Concen: 48.67 ng
 RT: 12.86 min Scan# 1731
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04


Tgt Ion	Ratio	Resp Lower	Resp Upper
196	100		
198	94.9	70.7	106.1
200	32.9	23.0	34.4





#43
 2,4,5-Trichlorophenol
 Concen: 50.48 ng
 RT: 12.93 min Scan# 1743
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04




Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD



Tgt Ion:	196	Resp:	185231
Ion Ratio	Lower	Upper	
196	100		
198	93.8	77.5	116.3
97	58.5	45.9	68.9
132	30.0	23.7	35.5




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#44
 2-Fluorobiphenyl
 Concen: 88.69 ng
 RT: 13.05 min Scan# 1764
 Delta R.T. -0.02 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion:	172	Resp:	1060578
Ion Ratio	Lower	Upper	
172	100		
171	36.3	29.0	43.4
170	24.9	19.0	28.4









#45
 1,1'-Biphenyl
 Concen: 47.96 ng
 RT: 13.26 min Scan# 1800
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
154	100		
153	42.0	22.0	62.0
76	13.2	0.0	33.3




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#46
 2-Chloronaphthalene
 Concen: 47.87 ng
 RT: 13.30 min Scan# 1807
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp Lower	Resp Upper
162	100		
127	40.8	31.2	46.8
164	33.6	27.8	41.6





#47

2-Nitroaniline

Concen: 59.02 ng

RT: 13.51 min Scan# 1842

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :


MWHR3-14MSD

Tgt Ion: 65 Resp: 210236
Ion Ratio Lower Upper




65	100		
92	63.3	51.0	76.6
138	95.0	76.2	114.2

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



Abundance Ion 65.00 (64.70 to 65.70): BGC
Ion 92.00 (91.70 to 92.70): BGC
Ion 138.00 (137.70 to 138.70): BGC



#48

Acenaphthylene

Concen: 46.90 ng

RT: 14.15 min Scan# 1951


Delta R.T. -0.02 min

Lab File: BG016980.D



Acq: 9 May 2015 7:04


Tgt Ion: 152 Resp: 835926
Ion Ratio Lower Upper

152	100		
151	20.6	17.0	25.4
153	13.1	10.7	16.1



Abundance Ion 152.00 (151.70 to 152.70): BGC
Ion 151.00 (150.70 to 151.70): BGC
Ion 153.00 (152.70 to 153.70): BGC









#49
Dimethylphthalate
Concen: 50.17 ng
RT: 13.89 min Scan# 1907
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
163	100	676097		
194	6.6	5.2	7.8	
164	10.7	8.4	12.6	




Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#50
2,6-Dinitrotoluene
Concen: 57.03 ng
RT: 14.01 min Scan# 1927
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	159542		
63	72.0	50.9	76.3	
89	63.4	51.4	77.2	









#51
Acenaphthene
Concen: 47.61 ng
RT: 14.50 min Scan# 2010
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	504924		
153	103.0	84.2	126.4	
152	50.0	40.2	60.2	




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#52
3-Nitroaniline
Concen: 26.06 ng
RT: 14.34 min Scan# 1983
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
138	100	82939		
108	11.2	9.5	14.3	
92	124.3	89.6	134.4	









#53
2,4-Dinitrophenol
Concen: 117.90 ng
RT: 14.54 min Scan# 2018
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
184	100	150867		
63	77.0	53.4	80.0	
154	58.8	41.0	61.6	




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#54
Dibenzofuran
Concen: 50.92 ng
RT: 14.83 min Scan# 2067
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	766869		
139	44.2	32.5	48.7	
169	13.5	10.4	15.6	









#55
4-Nitrophenol
Concen: 42.07 ng
RT: 14.65 min Scan# 2036
Delta R.T. 0.00 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
139	100	113543		
109	85.0	61.8	101.8	
65	108.3	96.8	136.8	




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#56
2,4-Dinitrotoluene
Concen: 62.18 ng
RT: 14.80 min Scan# 2061
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	222543		
63	57.5	44.2	66.4	
89	84.0	67.2	100.8	
182	4.9	4.2	6.4	





#57

Fluorene

Concen: 50.08 ng

RT: 15.48 min Scan# 2177

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G





ClientSampleId :

MWHR3-14MSD

Tgt Ion:	166	Resp:	667752
Ion Ratio	Lower	Upper	
166	100		
165	95.5	76.1	114.1
167	13.6	11.4	17.2

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#58

2,3,4,6-Tetrachlorophenol

Concen: 52.22 ng




RT: 15.06 min Scan# 2105


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	232	Resp:	165350
Ion Ratio	Lower	Upper	
232	100		
131	53.0	0.1	0.1#
130	3.8	0.0	0.0#
166	33.8	0.0	0.0#









#59
Diethylphthalate
Concen: 50.84 ng
RT: 15.26 min Scan# 2139
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	723142		
177	22.8	18.6	27.8	
150	13.6	9.8	14.8	




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#60
4-Chlorophenyl-phenylether
Concen: 50.04 ng
RT: 15.47 min Scan# 2176
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
204	100	335243		
206	31.5	24.8	37.2	
141	64.3	52.4	78.6	









#61
4-Nitroaniline
Concen: 47.39 ng
RT: 15.50 min Scan# 2181
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Lower	Upper
138	100		
92	54.6	31.8	71.8
108	97.9	70.6	110.6




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#62
Azobenzene
Concen: 51.99 ng
RT: 15.77 min Scan# 2226
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Lower	Upper
77	100		
182	27.8	7.4	47.4
105	16.9	0.0	36.6
51	37.6	13.4	53.4





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 17.17 min Scan# 2465

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G


ClientSampleId :

MWHR3-14MSD

Tgt	Ion:188	Resp:	408057
Ion	Ratio	Lower	Upper
188	100		
94	9.0	8.2	12.4
80	11.2	8.8	13.2

Manual Integrations
APPROVED




apatel
5/12/2015 8:30:17 AM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BGC

Ion 80.00 (79.70 to 80.70): BGC



#64

4,6-Dinitro-2-methylphenol

Concen: 56.50 ng


RT: 15.56 min Scan# 2190

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04



Tgt	Ion:198	Resp:	114307
Ion	Ratio	Lower	Upper
198	100		
51	58.7	31.2	71.2
105	45.7	22.0	62.0




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 51.00 (50.70 to 51.70): BGC

Ion 105.00 (104.70 to 105.70): E









#65
n-Nitrosodiphenylamine
Concen: 47.86 ng
RT: 15.69 min Scan# 2213
Delta R.T. -0.01 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
169	100	596187		
168	67.9	50.6	76.0	
167	36.2	29.0	43.4	




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#66
4-Bromophenyl-phenylether
Concen: 50.28 ng
RT: 16.37 min Scan# 2328
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
248	100	205922		
250	98.2	78.6	117.8	
141	87.1	66.3	99.5	





#67

Hexachlorobenzene

Concen: 50.91 ng

RT: 16.48 min Scan# 2347

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G





ClientSampleId :

MWHR3-14MSD

Tgt Ion:	284	Resp:	219906
Ion Ratio		Lower	Upper
284	100		
142	42.7	37.4	56.0
249	31.5	25.5	38.3

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#68

Atrazine

Concen: 52.39 ng




RT: 16.64 min Scan# 2375


Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	200	Resp:	234867
Ion Ratio		Lower	Upper
200	100		
173	26.4	6.1	46.1
215	52.2	30.8	70.8






#69
 Pentachlorophenol
 Concen: 96.96 ng
 RT: 16.82 min Scan# 2406
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	271890		
268	65.4	55.2	82.8	
264	61.5	51.4	77.0	

Manual Integrations
 APPROVED

apatel
 5/12/2015 8:30:17 AM






Abundance

Ion 265.70 (265.40 to 266.40): E


Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#70
 Phenanthrene
 Concen: 48.45 ng
 RT: 17.22 min Scan# 2473
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	1017386		
176	20.0	16.0	24.0	
179	15.2	13.0	19.6	






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 48.66 ng

RT: 17.31 min Scan# 2488

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G


ClientSampleId :

MWHR3-14MSD

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.4	16.2	24.4
179	16.9	13.1	19.7

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM




Abundance Ion 178.00 (177.70 to 178.70): E


Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E

17.31



Time-->



#72

Carbazole

Concen: 49.23 ng


RT: 17.58 min Scan# 2534

Delta R.T. -0.01 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Lower	Upper
167	100		
166	22.8	18.4	27.6
139	15.2	12.2	18.4




Abundance Ion 167.00 (166.70 to 167.70): E


Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E

17.58



Time-->



#73

Di-n-butylphthalate

Concen: 48.53 ng

RT: 18.14 min Scan# 2630

Delta R.T. -0.03 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :

MWHR3-14MSD

Tgt Ion:149 Resp: 1267143

Ion Ratio Lower Upper

149 100


150 9.9 7.7 11.5

104 6.8 5.4 8.0




Manual Integrations
APPROVED

apatel

5/12/2015 8:30:17 AM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 47.29 ng

RT: 19.23 min Scan# 2815

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04


Tgt Ion:202 Resp: 1152245

Ion Ratio Lower Upper



202 100


101 12.5 0.0 31.5

203 19.1 0.0 38.1



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E









#75
Chrysene-d12
Concen: 20.00 ng
RT: 21.35 min Scan# 3176
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Instrument :
BNA_G
ClientSampleId :
MWHR3-14MSD

Tgt Ion	Ratio	Lower	Upper
240	100		
120	10.9	9.1	13.7
236	26.3	20.2	30.4




Manual Integrations
APPROVED


apatel
5/12/2015 8:30:17 AM



#76
Benzidine
Concen: 29.87 ng
RT: 19.42 min Scan# 2847
Delta R.T. -0.02 min
Lab File: BG016980.D
Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Lower	Upper
184	100		
185	14.3	11.8	17.6
183	12.7	10.6	15.8





#77

Pyrene

Concen: 49.43 ng

RT: 19.59 min Scan# 2877

Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G


ClientSampleId :

MWHR3-14MSD




Tgt Ion:	202	Resp:	1179295
Ion Ratio	Lower	Upper	
202	100		
200	23.4	17.9	26.9
203	19.0	15.2	22.8

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 94.48 ng


RT: 19.80 min Scan# 2912

Delta R.T. -0.02 min



Lab File: BG016980.D


Acq: 9 May 2015 7:04

Tgt Ion:	244	Resp:	1629275
Ion Ratio	Lower	Upper	
244	100		
212	9.2	7.0	10.6
122	13.1	10.4	15.6



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E





#79

Butylbenzylphthalate

Concen: 49.75 ng

RT: 20.49 min Scan# 3030

Delta R.T. -0.03 min

Lab File: BG016980.D


Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :




MWHR3-14MSD



Tgt Ion:	149	Resp:	564168
Ion Ratio	Lower	Upper	
149	100		
91	75.0	60.6	91.0
206	23.2	18.4	27.6

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#80

Benzo(a)anthracene

Concen: 49.82 ng




RT: 21.33 min Scan# 3173


Delta R.T. -0.02 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	228	Resp:	1080357
Ion Ratio	Lower	Upper	
228	100		
226	27.1	22.7	34.1
229	20.9	16.1	24.1









#81
 3,3'-Dichlorobenzidine
 Concen: 26.11 ng
 RT: 21.27 min Scan# 3162
 Delta R.T. -0.03 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	221110		
254	64.6	51.7	77.5	
126	12.2	8.9	13.3	




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#82
 Chrysene
 Concen: 49.36 ng
 RT: 21.39 min Scan# 3183
 Delta R.T. -0.01 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
228	100	1002685		
226	32.5	25.4	38.0	
229	21.9	16.5	24.7	









#83
 Bis(2-ethylhexyl)phthalate
 Concen: 49.72 ng
 RT: 21.27 min Scan# 3162
 Delta R.T. -0.03 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	771248		
167	33.0	25.4	38.0	
279	9.2	6.4	9.6	




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#84
 Di-n-octyl phthalate
 Concen: 49.37 ng
 RT: 22.15 min Scan# 3313
 Delta R.T. -0.04 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Resp	Lower	Upper
149	100	1286536		
167	1.7	0.0	0.0#	
43	14.3	0.0	0.0#	









#85
 Indeno(1,2,3-cd)pyrene
 Concen: 51.20 ng
 RT: 26.01 min Scan# 3969
 Delta R.T. -0.04 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Lower	Upper
276	100		
138	24.8	0.0	0.0#
277	26.7	0.0	0.0#




Manual Integrations
 APPROVED


apatel
 5/12/2015 8:30:17 AM



#86
 Perylene-d12
 Concen: 20.00 ng
 RT: 23.65 min Scan# 3567
 Delta R.T. -0.03 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Tgt Ion	Ratio	Lower	Upper
264	100		
260	22.1	19.5	29.3
265	20.2	18.5	27.7





#87

Benzo(b)fluoranthene

Concen: 51.13 ng

RT: 22.95 min Scan# 3449

Delta R.T. -0.03 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Instrument :

BNA_G





ClientSampleId :

MWHR3-14MSD

Tgt Ion:	252	Resp:	1123590
Ion	Ratio	Lower	Upper
252	100		
253	22.5	17.4	26.0
125	10.2	8.6	13.0

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#88

Benzo(k)fluoranthene

Concen: 48.66 ng




RT: 23.00 min Scan# 3457


Delta R.T. -0.03 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	252	Resp:	1028688
Ion	Ratio	Lower	Upper
252	100		
253	23.0	18.4	27.6
125	10.1	7.9	11.9





#89

Benzo(a)pyrene

Concen: 50.96 ng

RT: 23.55 min Scan# 3550

Delta R.T. -0.03 min

Lab File: BG016980.D


Acq: 9 May 2015 7:04

Instrument :

BNA_G

ClientSampleId :



MWHR3-14MSD



Tgt Ion:	252	Resp:	1044127
Ion Ratio	Lower	Upper	
252	100		
253	23.9	17.7	26.5
125	10.8	8.5	12.7

Manual Integrations
APPROVED

apatel
5/12/2015 8:30:17 AM



#90

Dibenzo(a,h)anthracene

Concen: 50.73 ng




RT: 26.02 min Scan# 3971


Delta R.T. -0.05 min

Lab File: BG016980.D

Acq: 9 May 2015 7:04

Tgt Ion:	278	Resp:	1061671
Ion Ratio	Lower	Upper	
278	100		
139	16.0	14.2	21.2
279	24.4	18.9	28.3





#91
 Benzo(g,h,i)perylene
 Concen: 50.67 ng
 RT: 26.73 min Scan# 4092
 Delta R.T. -0.04 min
 Lab File: BG016980.D
 Acq: 9 May 2015 7:04

Instrument :
 BNA_G
 ClientSampleId :
 MWHR3-14MSD

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.6	29.4
138	21.0	17.5	26.3

Manual Integrations
 APPROVED

apatel
 5/12/2015 8:30:17 AM

