


Data Path : Z:\HPCHEM1\BNA_G\Data\BG050815\
 Data File : BG017010.D
 Acq On : 10 May 2015 5:03
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2539, 2540, 2541; Background Corrected with Scan 2532

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.8	28680	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	49.0	26616	PASS
70	69	0.00	2	0.9	243	PASS
127	198	10	80	55.6	30248	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	54362	PASS
199	198	5	9	7.3	3949	PASS
275	198	10	60	28.4	15455	PASS
365	198	1	100	5.2	2814	PASS
441	442	0.01	24	14.1	5792	PASS
442	198	50	100	75.6	41114	PASS
443	442	15	24	18.6	7662	PASS