

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050922\
 Data File : BG053487.D
 Acq On : 9 May 2022 11:03
 Operator : CG/JU
 Sample : PB144630BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB144630BL

Quant Time: May 09 16:20:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 02:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.085	152	23444	20.000	ng	-0.01
21) Naphthalene-d8	10.905	136	98496	20.000	ng	# 0.00
39) Acenaphthene-d10	14.717	164	78722	20.000	ng	0.00
64) Phenanthrene-d10	17.460	188	204228	20.000	ng	0.00
76) Chrysene-d12	21.737	240	227544	20.000	ng	# 0.00
86) Perylene-d12	25.021	264	226153	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.659	112	203815	147.544	ng	0.00
7) Phenol-d6	7.251	99	300202	143.194	ng	0.00
23) Nitrobenzene-d5	9.254	82	207395	89.633	ng	0.00
42) 2,4,6-Tribromophenol	16.203	330	148183	127.628	ng	0.00
45) 2-Fluorobiphenyl	13.337	172	476503	88.451	ng	0.00
79) Terphenyl-d14	20.063	244	1080848	88.468	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050922\
Data File : BG053487.D
Acq On : 9 May 2022 11:03
Operator : CG/JU
Sample : PB144630BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB144630BL

Quant Time: May 09 16:20:05 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 02:13:53 2022
Response via : Initial Calibration

