

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050922\
 Data File : BG053486.D
 Acq On : 9 May 2022 10:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 09 16:19:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 02:13:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.618	0.590	4.5	101	0.00
3	Pyridine	1.510	1.584	-4.9	105	-0.02
4	n-Nitrosodimethylamine	0.736	0.734	0.3	103	0.00
5 S	2-Fluorophenol	1.178	1.161	1.4	103	-0.02
6	Aniline	1.981	2.065	-4.2	106	0.00
7 S	Phenol-d6	1.788	1.799	-0.6	104	0.00
8	2-Chlorophenol	1.232	1.197	2.8	101	0.00
9	Benzaldehyde	1.136	1.070	5.8	105	0.00
10 C	Phenol	1.749	1.755	-0.3	102	0.00
11	bis(2-Chloroethyl)ether	1.305	1.280	1.9	100	0.00
12	1,3-Dichlorobenzene	1.453	1.396	3.9	101	0.00
13 C	1,4-Dichlorobenzene	1.467	1.412	3.7	103	0.00
14	1,2-Dichlorobenzene	1.377	1.372	0.4	104	0.00
15	Benzyl Alcohol	1.495	1.512	-1.1	103	0.00
16	2,2'-oxybis(1-Chloropropane	2.123	2.115	0.4	104	0.00
17	2-Methylphenol	1.182	1.157	2.1	102	-0.02
18	Hexachloroethane	0.557	0.526	5.6	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.283	1.259	1.9	102	0.00
20	3+4-Methylphenols	1.670	1.668	0.1	101	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.547	0.559	-2.2	105	0.00
23 S	Nitrobenzene-d5	0.470	0.487	-3.6	103	0.00
24	Nitrobenzene	0.455	0.470	-3.3	102	0.00
25	Isophorone	0.791	0.825	-4.3	102	0.00
26 C	2-Nitrophenol	0.184	0.192	-4.3	103	-0.02
27	2,4-Dimethylphenol	0.277	0.281	-1.4	101	0.00
28	bis(2-Chloroethoxy)methane	0.455	0.462	-1.5	102	0.00
29 C	2,4-Dichlorophenol	0.335	0.347	-3.6	100	0.00
30	1,2,4-Trichlorobenzene	0.380	0.390	-2.6	104	0.00
31	Naphthalene	1.017	1.042	-2.5	104	0.00
32	Benzoic acid	0.147	0.150	-2.0	96	0.00
33	4-Chloroaniline	0.426	0.447	-4.9	102	0.00
34 C	Hexachlorobutadiene	0.282	0.286	-1.4	102	0.00
35	Caprolactam	0.122	0.129	-5.7	105	0.00
36 C	4-Chloro-3-methylphenol	0.400	0.423	-5.7	102	0.00
37	2-Methylnaphthalene	0.759	0.772	-1.7	101	0.00
38	1-Methylnaphthalene	0.741	0.739	0.3	102	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
40	1,2,4,5-Tetrachlorobenzene	0.630	0.639	-1.4	102	0.00
41 P	Hexachlorocyclopentadiene	0.200	0.208	-4.0	100	0.00
42 S	2,4,6-Tribromophenol	0.295	0.302	-2.4	100	0.00
43 C	2,4,6-Trichlorophenol	0.412	0.430	-4.4	101	0.00
44	2,4,5-Trichlorophenol	0.438	0.451	-3.0	100	0.00
45 S	2-Fluorobiphenyl	1.369	1.373	-0.3	103	0.00
46	1,1'-Biphenyl	1.363	1.375	-0.9	103	0.00
47	2-Chloronaphthalene	1.068	1.065	0.3	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050922\
 Data File : BG053486.D
 Acq On : 9 May 2022 10:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 09 16:19:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu May 05 02:13:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.398	0.421	-5.8	104	0.00
49	Acenaphthylene	1.704	1.722	-1.1	102	0.00
50	Dimethylphthalate	1.520	1.536	-1.1	102	0.00
51	2,6-Dinitrotoluene	0.310	0.318	-2.6	102	0.00
52 C	Acenaphthene	1.016	1.028	-1.2	102	0.00
53	3-Nitroaniline	0.335	0.353	-5.4	103	0.00
54 P	2,4-Dinitrophenol	0.166	0.177	-6.6	103	0.00
55	Dibenzofuran	1.824	1.829	-0.3	103	0.00
56 P	4-Nitrophenol	0.239	0.244	-2.1	99	0.00
57	2,4-Dinitrotoluene	0.449	0.472	-5.1	103	0.00
58	Fluorene	1.464	1.467	-0.2	101	0.00
59	2,3,4,6-Tetrachlorophenol	0.444	0.447	-0.7	99	0.00
60	Diethylphthalate	1.570	1.585	-1.0	101	0.00
61	4-Chlorophenyl-phenylether	0.817	0.843	-3.2	105	0.00
62	4-Nitroaniline	0.347	0.366	-5.5	105	0.00
63	Azobenzene	1.608	1.644	-2.2	105	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.126	-5.9	102	0.00
66 c	n-Nitrosodiphenylamine	0.525	0.519	1.1	100	0.00
67	4-Bromophenyl-phenylether	0.234	0.237	-1.3	101	0.00
68	Hexachlorobenzene	0.256	0.250	2.3	100	0.00
69	Atrazine	0.221	0.221	0.0	102	0.00
70 C	Pentachlorophenol	0.121	0.123	-1.7	100	0.00
71	Phenanthrene	1.017	1.022	-0.5	103	0.00
72	Anthracene	1.000	0.993	0.7	103	0.00
73	Carbazole	0.980	0.981	-0.1	103	0.00
74	Di-n-butylphthalate	1.102	1.104	-0.2	102	0.00
75 C	Fluoranthene	1.317	1.301	1.2	101	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
77	Benzidine	0.431	0.423	1.9	101	0.00
78	Pyrene	1.295	1.327	-2.5	102	0.00
79 S	Terphenyl-d14	1.074	1.074	0.0	101	0.00
80	Butylbenzylphthalate	0.475	0.498	-4.8	102	0.00
81	Benzo(a)anthracene	1.276	1.305	-2.3	103	0.00
82	3,3'-Dichlorobenzidine	0.323	0.329	-1.9	99	0.00
83	Chrysene	1.191	1.214	-1.9	102	0.00
84	Bis(2-ethylhexyl)phthalate	0.655	0.691	-5.5	104	0.00
85 c	Di-n-octyl phthalate	1.108	1.163	-5.0	103	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	-0.02
87	Indeno(1,2,3-cd)pyrene	1.393	1.429	-2.6	103	0.00
88	Benzo(b)fluoranthene	1.184	1.216	-2.7	102	0.00
89	Benzo(k)fluoranthene	1.163	1.159	0.3	103	0.00
90 C	Benzo(a)pyrene	1.008	1.015	-0.7	101	0.00
91	Dibenzo(a,h)anthracene	1.148	1.172	-2.1	104	-0.02
92	Benzo(g,h,i)perylene	1.138	1.158	-1.8	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG050922\
Data File : BG053486.D
Acq On : 9 May 2022 10:19
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 09 16:19:22 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu May 05 02:13:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0