

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051017\
 Data File : BG026945.D
 Acq On : 10 May 2017 19:41
 Operator : SJ/MA
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02084

Quant Time: May 11 09:07:20 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 10 01:49:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.507	0.444	12.4	90	0.00
3 S	1,4-Dioxane-d8	0.444	0.383	13.7	84	0.00
4	Benzaldehyde	0.863	0.840	2.7	114	0.00
5 S	Phenol-d5	1.785	1.909	-6.9	104	0.01
6	Phenol	1.809	1.929	-6.6	102	0.01
7 S	Bis-(2-Chloroethyl)ether-d8	1.050	1.149	-9.4	105	0.00
8	Bis(2-Chloroethyl)ether	1.397	1.409	-0.9	95	0.00
9 S	2-Chlorophenol-d4	1.515	1.557	-2.8	101	0.00
10	2-Chlorophenol	1.491	1.529	-2.5	98	0.00
11	2-Methylphenol	1.432	1.529	-6.8	104	0.00
12	2,2'-oxybis(1-Chloropropane	1.458	1.818	-24.7#	119	0.00
13 S	4-Methylphenol-d8	1.481	1.547	-4.5	102	0.01
14	Acetophenone	2.176	2.270	-4.3	98	0.00
15 P	N-Nitroso-di-n-propylamine	1.073	1.079	-0.6	99	0.00
16	4-Methylphenol	1.566	1.660	-6.0	103	0.00
17	Hexachloroethane	0.561	0.573	-2.1	104	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
19 S	Nitrobenzene-d5	0.167	0.167	0.0	102	0.00
20	Nitrobenzene	0.323	0.340	-5.3	105	0.00
21	Isophorone	0.619	0.632	-2.1	101	0.00
22 S	2-Nitrophenol-d4	0.182	0.183	-0.5	100	0.00
23 C	2-Nitrophenol	0.191	0.190	0.5	101	0.00
24	2,4-Dimethylphenol	0.349	0.363	-4.0	106	0.00
25	Bis(2-Chloroethoxy)methane	0.423	0.416	1.7	98	0.00
26 S	2,4-Dichlorophenol-d3	0.289	0.295	-2.1	102	0.00
27 C	2,4-Dichlorophenol	0.288	0.291	-1.0	101	0.01
28	Naphthalene	1.041	1.050	-0.9	100	0.01
29 S	4-Chloroaniline-d4	0.387	0.424	-9.6	100	0.00
30	4-Chloroaniline	0.365	0.402	-10.1	102	0.00
31 C	Hexachlorobutadiene	0.146	0.137	6.2	95	0.00
32	Caprolactam	0.108	0.110	-1.9	105	0.01
33 C	4-Chloro-3-methylphenol	0.312	0.336	-7.7	106	0.01
34	2-Methylnaphthalene	0.773	0.764	1.2	99	0.01
35 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
36	1,2,4,5-Tetrachlorobenzene	0.527	0.518	1.7	96	0.01
37	Hexachlorocyclopentadiene	0.232	0.131	43.5#	59	0.00
38 C	2,4,6-Trichlorophenol	0.363	0.371	-2.2	98	0.00
39	2,4,5-Trichlorophenol	0.375	0.369	1.6	94	0.01
40	1,1'-Biphenyl	1.681	1.694	-0.8	97	0.00
41	2-Chloronaphthalene	1.264	1.259	0.4	97	0.01
42	2-Nitroaniline	0.361	0.425	-17.7	113	0.00
43 S	Dimethylphthalate-d6	1.610	1.566	2.7	93	0.00
44	Dimethylphthalate	1.576	1.515	3.9	93	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051017\
 Data File : BG026945.D
 Acq On : 10 May 2017 19:41
 Operator : SJ/MA
 Sample : SSTDC0020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTD02084

Quant Time: May 11 09:07:20 2017
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed May 10 01:49:01 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.333	0.332	0.3	95	0.00
46 S	Acenaphthylene-d8	2.022	2.056	-1.7	97	0.00
47	Acenaphthylene	2.060	2.078	-0.9	97	0.00
48	3-Nitroaniline	0.350	0.373	-6.6	105	0.00
49 C	Acenaphthene	1.433	1.428	0.3	96	0.00
50	2,4-Dinitrophenol	0.170	0.122	28.2#	63	0.01
51 S	4-Nitrophenol-d4	0.290	0.243	16.2	84	0.01
52	4-Nitrophenol	0.170	0.146	14.1	84	0.02
53	Dibenzofuran	1.907	1.854	2.8	93	0.00
54	2,4-Dinitrotoluene	0.476	0.458	3.8	91	0.01
55	2,3,4,6-Tetrachlorophenol	0.295	0.263	10.8	84	0.00
56	Diethylphthalate	1.649	1.611	2.3	94	0.00
57 S	Fluorene-d10	1.405	1.342	4.5	92	0.00
58	Fluorene	1.557	1.490	4.3	92	0.00
59	4-Chlorophenyl-phenylether	0.679	0.634	6.6	91	0.00
60	4-Nitroaniline	0.386	0.385	0.3	104	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.01
62 S	4,6-Dinitro-2-methylphenol-	0.121	0.085	29.8#	65	0.01
63	4,6-Dinitro-2-methylphenol	0.126	0.088	30.2#	64	0.01
64	N-Nitrosodiphenylamine	0.657	0.682	-3.8	91	0.00
65	4-Bromophenyl-phenylether	0.200	0.196	2.0	86	0.00
66	Hexachlorobenzene	0.215	0.201	6.5	84	0.00
67	Atrazine	0.205	0.211	-2.9	88	0.00
68 C	Pentachlorophenol	0.098	0.083	15.3	87	0.00
69	Phenanthrene	1.127	1.128	-0.1	87	0.00
70 S	Anthracene-d10	0.971	0.975	-0.4	88	0.00
71	Anthracene	1.156	1.172	-1.4	88	0.00
72	Carbazole	0.977	1.069	-9.4	88	0.00
73	Di-n-butylphthalate	1.272	1.393	-9.5	92	0.00
74 C	Fluoranthene	1.046	1.127	-7.7	85	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	85	0.01
76 S	Pyrene-d10	1.056	1.055	0.1	84	0.01
77	Pyrene	1.342	1.360	-1.3	85	0.01
78	Butylbenzylphthalate	0.623	0.743	-19.3	100	0.00
79	3,3'-Dichlorobenzidine	0.385	0.399	-3.6	86	0.00
80	Benzo(a)anthracene	1.170	1.186	-1.4	86	0.00
81	Bis(2-ethylhexyl)phthalate	0.889	1.048	-17.9	97	0.00
82	Chrysene	1.087	1.085	0.2	85	0.00
83 I	Perylene-d12	1.000	1.000	0.0	85	0.02
84	Di-n-octyl phthalate	1.374	1.751	-27.4#	100	0.00
85	Benzo(b)fluoranthene	1.143	1.115	2.4	85	0.02
86	Benzo(k)fluoranthene	1.123	1.168	-4.0	88	0.02
87 S	Benzo(a)pyrene-d12	0.946	0.965	-2.0	88	0.02

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051017\
Data File : BG026945.D
Acq On : 10 May 2017 19:41
Operator : SJ/MA
Sample : SSTDCCC020EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTD02084

Quant Time: May 11 09:07:20 2017
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed May 10 01:49:01 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.125	1.122	0.3	85	0.02
89	Indeno(1,2,3-cd)pyrene	1.333	1.074	19.4	70	0.02
90	Dibenzo(a,h)anthracene	1.130	0.936	17.2	71	0.03
91	Benzo(g,h,i)perylene	1.106	0.738	33.3#	58	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0