



Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051024\
Data File : BG061159.D
Acq On : 10 May 2024 11:06
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed May 01 05:11:25 2024



AutoFind: Scans 2512, 2513, 2514; Background Corrected with Scan 2505

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	25536	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	29.8	18782	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	37.7	23787	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	63085	PASS
199	198	5	9	6.8	4282	PASS
275	198	10	60	36.9	23280	PASS
365	198	1	100	6.3	3975	PASS
441	198	0.01	100	21.2	13382	PASS
442	442	50	100	100.0	82811	PASS
443	442	15	24	20.0	16586	PASS

DDT Breakdown




Date	Instrument Name	DFTPP Data File
5/10/2024	BNA_G	<u>BG061159.D</u>
Compound Name	Response	Retention Time
DDT	234114	20.782
DDD	6036	20.341
DDE	380	19.842
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
6416	240530	2.67

Instrument :
BNA_G
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051024\
Data File : BG061159.D
Acq On : 10 May 2024 11:06
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 10 13:01:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061159.D\data.ms

(70) Pentachlorophenol (C)

16.968min (-0.018) 14669.05 ng




response 59857

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.80	68.72
264.00	64.40	62.87
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051024\
Data File : BG061159.D
Acq On : 10 May 2024 11:06
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP

Quant Time: May 10 13:01:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061159.D\data.ms

(77) Benzidine

19.548min (-0.017) 231168.22 ng

response 350318

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	14.98
183.00	13.60	11.85
0.00	0.00	0.00