

Data Path : Z:\HPCHEM1\BNA G\Data\BG051115\
 Data File : BG017049.D
 Acq On : 11 May 2015 12:35
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 12 04:13:19 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 04:00:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	-0.05
2	1,4-Dioxane	0.481	0.467	2.9	85	-0.07
3	Pyridine	1.488	1.465	1.5	83	-0.06
4	n-Nitrosodimethylamine	0.888	0.937	-5.5	94	-0.06
5 S	2-Fluorophenol	1.149	1.181	-2.8	89	-0.04
6	Aniline	2.304	2.250	2.3	85	-0.05
7 S	Phenol-d6	1.641	1.645	-0.2	87	-0.02
8	2-Chlorophenol	1.331	1.378	-3.5	91	-0.05
9	Benzaldehyde	1.144	1.051	8.1	77	-0.05
10 C	Phenol	1.760	1.740	1.1	86	-0.02
11	bis(2-Chloroethyl)ether	1.361	1.306	4.0	83	-0.05
12	1,3-Dichlorobenzene	1.467	1.483	-1.1	90	-0.05
13 C	1,4-Dichlorobenzene	1.486	1.526	-2.7	89	-0.05
14	1,2-Dichlorobenzene	1.426	1.424	0.1	87	-0.05
15	Benzyl Alcohol	1.499	1.505	-0.4	87	-0.05
16	2,2'-oxybis(1-Chloropropane	2.519	2.090	17.0	74	-0.05
17	2-Methylphenol	1.239	1.245	-0.5	87	-0.02
18	Hexachloroethane	0.611	0.609	0.3	87	-0.06
19 P	n-Nitroso-di-n-propylamine	1.439	1.407	2.2	86	-0.05
20	3+4-Methylphenols	1.708	1.719	-0.6	87	-0.03
21 I	Naphthalene-d8	1.000	1.000	0.0	81	-0.05
22	Acetophenone	0.476	0.477	-0.2	86	-0.05
23 S	Nitrobenzene-d5	0.409	0.421	-2.9	90	-0.05
24	Nitrobenzene	0.413	0.419	-1.5	88	-0.05
25	Isophorone	0.827	0.823	0.5	87	-0.05
26 C	2-Nitrophenol	0.168	0.190	-13.1	97	-0.05
27	2,4-Dimethylphenol	0.305	0.312	-2.3	87	-0.03
28	bis(2-Chloroethoxy)methane	0.417	0.412	1.2	88	-0.05
29 C	2,4-Dichlorophenol	0.288	0.299	-3.8	90	-0.03
30	1,2,4-Trichlorobenzene	0.305	0.313	-2.6	89	-0.05
31	Naphthalene	0.995	0.972	2.3	86	-0.05
32	Benzoic acid	0.179	0.235	-31.3#	116	0.00
33	4-Chloroaniline	0.459	0.476	-3.7	89	-0.04
34 C	Hexachlorobutadiene	0.191	0.195	-2.1	90	-0.05
35	Caprolactam	0.150	0.167	-11.3	97	-0.03
36 C	4-Chloro-3-methylphenol	0.369	0.380	-3.0	89	-0.01
37	2-Methylnaphthalene	0.758	0.759	-0.1	89	-0.05
38 I	Acenaphthene-d10	1.000	1.000	0.0	86	-0.04
39	1,2,4,5-Tetrachlorobenzene	0.541	0.541	0.0	90	-0.04
40 P	Hexachlorocyclopentadiene	0.349	0.325	6.9	84	-0.05
41 S	2,4,6-Tribromophenol	0.201	0.232	-15.4	105	-0.03
42 C	2,4,6-Trichlorophenol	0.383	0.408	-6.5	92	-0.03
43	2,4,5-Trichlorophenol	0.407	0.419	-2.9	91	-0.01
44 S	2-Fluorobiphenyl	1.325	1.313	0.9	89	-0.05

Data Path : Z:\HPCHEM1\BNA G\Data\BG051115\
 Data File : BG017049.D
 Acq On : 11 May 2015 12:35
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 12 04:13:19 2015
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 12 04:00:00 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.432	1.417	1.0	89	-0.04
46	2-Chloronaphthalene	1.134	1.128	0.5	89	-0.04
47	2-Nitroaniline	0.395	0.458	-15.9	102	-0.02
48	Acenaphthylene	1.975	1.974	0.1	90	-0.04
49	Dimethylphthalate	1.493	1.542	-3.3	93	-0.03
50	2,6-Dinitrotoluene	0.310	0.360	-16.1	101	-0.03
51 C	Acenaphthene	1.175	1.136	3.3	87	-0.04
52	3-Nitroaniline	0.353	0.397	-12.5	96	-0.02
53 P	2,4-Dinitrophenol	0.142	0.194	-36.6#	130	-0.03
54	Dibenzofuran	1.669	1.685	-1.0	90	-0.04
55 P	4-Nitrophenol	0.299	0.334	-11.7	99	0.02
56	2,4-Dinitrotoluene	0.397	0.497	-25.2#	110	-0.03
57	Fluorene	1.478	1.490	-0.8	91	-0.03
58	2,3,4,6-Tetrachlorophenol	0.351	0.380	-8.3	94	-0.02
59	Diethylphthalate	1.577	1.653	-4.8	95	-0.04
60	4-Chlorophenyl-phenylether	0.742	0.758	-2.2	94	-0.04
61	4-Nitroaniline	0.409	0.458	-12.0	99	-0.02
62	Azobenzene	1.633	1.578	3.4	88	-0.04
63 I	Phenanthrene-d10	1.000	1.000	0.0	84	-0.03
64	4,6-Dinitro-2-methylphenol	0.099	0.138	-39.4#	126	-0.03
65 c	n-Nitrosodiphenylamine	0.611	0.626	-2.5	92	-0.03
66	4-Bromophenyl-phenylether	0.201	0.209	-4.0	94	-0.03
67	Hexachlorobenzene	0.212	0.227	-7.1	98	-0.03
68	Atrazine	0.220	0.238	-8.2	95	-0.03
69 C	Pentachlorophenol	0.137	0.143	-4.4	90	-0.02
70	Phenanthrene	1.029	1.047	-1.7	91	-0.03
71	Anthracene	1.038	1.066	-2.7	92	-0.03
72	Carbazole	0.987	1.008	-2.1	90	-0.03
73	Di-n-butylphthalate	1.280	1.347	-5.2	93	-0.04
74 C	Fluoranthene	1.194	1.241	-3.9	91	-0.03
75 I	Chrysene-d12	1.000	1.000	0.0	82	-0.03
76	Benzidine	0.680	0.661	2.8	81	-0.03
77	Pyrene	1.180	1.222	-3.6	91	-0.03
78 S	Terphenyl-d14	0.853	0.943	-10.6	95	-0.03
79	Butylbenzylphthalate	0.561	0.583	-3.9	90	-0.03
80	Benzo(a)anthracene	1.073	1.142	-6.4	92	-0.03
81	3,3'-Dichlorobenzidine	0.419	0.446	-6.4	92	-0.03
82	Chrysene	1.005	1.064	-5.9	91	-0.03
83	Bis(2-ethylhexyl)phthalate	0.767	0.814	-6.1	91	-0.04
84 c	Di-n-octyl phthalate	1.289	1.364	-5.8	90	-0.05
85	Indeno(1,2,3-cd)pyrene	1.198	1.268	-5.8	95	-0.06
86 I	Perylene-d12	1.000	1.000	0.0	82	-0.05
87	Benzo(b)fluoranthene	1.114	1.175	-5.5	92	-0.04

Data Path : Z:\HPCHEM1\BNA G\Data\BG051115\
Data File : BG017049.D
Acq On : 11 May 2015 12:35
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 12 04:13:19 2015
Quant Method : Z:\HPCHEM1\BNA G\METHODS\8270-BG050515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 12 04:00:00 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.072	1.106	-3.2	91	-0.04
89 C	Benzo(a)pyrene	1.039	1.080	-3.9	91	-0.05
90	Dibenzo(a,h)anthracene	1.061	1.114	-5.0	94	-0.06
91	Benzo(a,h,i)perylene	1.011	1.058	-4.6	95	-0.06

(#) = Out of Range

SPCC's out = 0 CCC's out = 0