


Data Path : Z:\HPCHEM1\BNA G\Data\BG051115\
 Data File : BG017048.D
 Acq On : 11 May 2015 11:11
 Operator : TP/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG050515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 05 16:13:58 2015



AutoFind: Scans 2536, 2537, 2538; Background Corrected with Scan 2530

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.1	27513	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	49.3	28767	PASS
70	69	0.00	2	0.0	6	PASS
127	198	10	80	53.5	31278	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	58410	PASS
199	198	5	9	7.9	4596	PASS
275	198	10	60	27.7	16170	PASS
365	198	1	100	5.3	3099	PASS
441	442	0.01	24	15.1	6559	PASS
442	198	50	100	74.5	43506	PASS
443	442	15	24	19.1	8310	PASS