

Data Path : Z:\HPCHEM1\BNA G\DATA\BG051116\
 Data File : BG021852.D
 Acq On : 11 May 2016 20:20
 Operator : UM/SJ
 Sample : H2937-19
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 D9X97

Quant Time: May 12 07:32:52 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG050916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 12 07:33:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.18	152	83254	20.00	ng/ul	0.00
18) Naphthalene-d8	11.00	136	405511	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.80	164	230397	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.55	188	473884	20.00	ng/ul	0.00
76) Chrysene-d12	21.84	240	472711	20.00	ng/ul	0.01
84) Perylene-d12	25.20	264	472994	20.00	ng/ul	0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.57	96	5354	3.10	ng/uL	0.00
5) Phenol-d5	7.34	99	108651	17.59	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.50	67	56438	18.42	ng/ul	0.00
9) 2-Chlorophenol-d4	7.71	132	103503	21.59	ng/ul	0.00
13) 4-Methylphenol-d8	8.89	113	84058	16.20	ng/ul	0.00
19) Nitrobenzene-d5	9.36	128	51783	19.42	ng/ul	0.00
22) 2-Nitrophenol-d4	10.08	143	55555	32.99	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.74	165	468	0.10	ng/ul	0.13
29) 4-Chloroaniline-d4	11.14	131	121102	23.61	ng/ul	0.00
44) Dimethylphthalate-d6	14.20	166	326417	20.17	ng/ul	0.00
47) Acenaphthylene-d8	14.50	160	449469	19.56	ng/ul	0.00
52) 4-Nitrophenol-d4	15.00	143	43403	15.65	ng/ul	0.00
58) Fluorene-d10	15.80	176	299216	17.56	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.91	200	38106	29.03	ng/ul	0.00
71) Anthracene-d10	17.65	188	413200	17.91	ng/ul	0.00
77) Pyrene-d10	19.93	212	440777	19.92	ng/ul	0.00
88) Benzo(a)pyrene-d12	24.97	264	420958	18.73	ng/ul	0.04

Target Compounds


						Ovalue
45) Dimethylphthalate	14.25	163	54699	3.32	ng/ul	98
82) Bis(2-ethylhexyl)phthalate	21.71	149	44647	4.52	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG051116\
Data File : BG021852.D
Acq On : 11 May 2016 20:20
Operator : UM/SJ
Sample : H2937-19
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9X97

Quant Time: May 12 07:32:52 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG050916.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 12 07:33:29 2016
Response via : Initial Calibration









#45
 Dimethylphthalate
 Concen: 3.32 ng/ul
 RT: 14.25 min Scan# 1942
 Delta R.T. 0.01 min
 Lab File: BG021852.D
 Acq: 11 May 2016 20:20

Instrument :
 BNA_G
 ClientSampleId :
 D9X97

Tot Ion	Ratio	Lower	Upper
163	100		
194	4.2	3.9	5.9
164	11.5	8.7	13.1



#82
 Bis(2-ethylhexyl)phthalate
 Concen: 4.52 ng/ul
 RT: 21.71 min Scan# 3211
 Delta R.T. 0.01 min
 Lab File: BG021852.D
 Acq: 11 May 2016 20:20

Tgt Ion	Ratio	Lower	Upper
149	100		
167	31.8	20.5	30.7#
279	5.4	3.7	5.5

