

Data Path : Z:\HPCHEM1\BNA G\DATA\BG051116\
 Data File : BG021860.D
 Acq On : 12 May 2016 1:41
 Operator : UM/SJ
 Sample : H2936-06MSD
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 D9WF4MSD

Manual Integrations
 APPROVED

sohil
 5/12/2016 7:26:57 PM

Quant Time: May 12 09:26:08 2016
 Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG050916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu May 12 07:33:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.19	152	74768	20.00	ng/ul	0.00
18) Naphthalene-d8	11.01	136	360518	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.80	164	205714	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.55	188	435169	20.00	ng/ul	0.00
76) Chrysene-d12	21.83	240	424593	20.00	ng/ul	0.00
84) Perylene-d12	25.18	264	428145	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.57	96	4387	2.83	ng/uL	0.00
5) Phenol-d5	7.33	99	90365	16.29	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.50	67	48009	17.45	ng/ul	0.00
9) 2-Chlorophenol-d4	7.71	132	88488	20.55	ng/ul	0.00
13) 4-Methylphenol-d8	8.88	113	61434	13.18	ng/ul	0.00
19) Nitrobenzene-d5	9.35	128	44813	18.91	ng/ul	0.00
22) 2-Nitrophenol-d4	10.08	143	46508	31.07	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.62	165	82400	19.47	ng/ul	0.00
29) 4-Chloroaniline-d4	11.13	131	48553	10.65	ng/ul	0.00
44) Dimethylphthalate-d6	14.20	166	265812	18.40	ng/ul	0.00
47) Acenaphthylene-d8	14.50	160	362963	17.69	ng/ul	0.00
52) 4-Nitrophenol-d4	15.00	143	29091	11.75	ng/ul	0.00
58) Fluorene-d10	15.79	176	244566	16.08	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.91	200	28086	23.30	ng/ul	0.00
71) Anthracene-d10	17.65	188	344296	16.25	ng/ul	0.00
77) Pyrene-d10	19.92	212	370533	18.64	ng/ul	0.00
88) Benzo(a)pyrene-d12	24.95	264	343941	16.91	ng/ul	0.02

Target Compounds

					Ovalue	
6) Phenol	7.36	94	91658	15.53	ng/ul#	70
10) 2-Chlorophenol	7.74	128	83983	19.34	ng/ul#	77
15) N-Nitroso-di-n-propylamine	8.98	70	57042	16.45	ng/ul#	74
33) 4-Chloro-3-methylphenol	12.26	107	76107	17.45	ng/ul#	67
45) Dimethylphthalate	14.25	163	46938	3.19	ng/ul	96
50) Acenaphthene	14.87	153	247849	16.97	ng/ul	96
53) 4-Nitrophenol	15.01	109	13836m	12.26	ng/ul	
55) 2,4-Dinitrotoluene	15.16	165	71834	17.86	ng/ul#	55
69) Pentachlorophenol	17.19	266	47593	23.73	ng/ul	92
70) Phenanthrene	17.59	178	42156	1.77	ng/ul	96
75) Fluoranthene	19.59	202	91844	3.40	ng/ul#	95
78) Pyrene	19.95	202	538969m	21.94	ng/ul	
79) Butylbenzylphthalate	20.82	149	9015	1.66	ng/ul#	86
81) Benzo(a)anthracene	21.82	228	52929	2.24	ng/ul	95
82) Bis(2-ethylhexyl)phthalate	21.70	149	49401	5.57	ng/ul	95
83) Chrysene	21.88	228	64633m	2.77	ng/ul	
86) Benzo(b)fluoranthene	24.12	252	114953	4.69	ng/ul#	93
87) Benzo(k)fluoranthene	24.17	252	32489m	1.24	ng/ul	
89) Benzo(a)pyrene	25.02	252	64701	2.65	ng/ul#	94
90) Indeno(1,2,3-cd)pyrene	29.01	276	46800	1.62	ng/ul#	92
92) Benzo(a,h,i)perylene	30.21	276	40115m	1.68	ng/ul	

Data Path : Z:\HPCHEM1\BNA G\DATA\BG051116\
Data File : BG021860.D
Acq On : 12 May 2016 1:41
Operator : UM/SJ
Sample : H2936-06MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9WF4MSD

Manual Integrations
APPROVED

sohil
5/12/2016 7:26:57 PM

Quant Time: May 12 09:26:08 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG050916.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 12 07:33:29 2016
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA G\DATA\BG051116\
Data File : BG021860.D
Acq On : 12 May 2016 1:41
Operator : UM/SJ
Sample : H2936-06MSD
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9WF4MSD

Manual Integrations
APPROVED

sohil
5/12/2016 7:26:57 PM

Quant Time: May 12 09:26:08 2016
Quant Method : Z:\HPCHEM1\BNA G\METHODS\SOM02.2-EPA-BG050916.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu May 12 07:33:29 2016
Response via : Initial Calibration

