




Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051120\
Data File : BG045291.D
Acq On : 11 May 2020 11:08
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 02:25:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 11 13:27:41 2020
Response via : Initial Calibration



TIC: BG045291.D

(70) Pentachlorophenol (C)




17.021min (-0.007) 0.00ng

response 53525

Ion	Exp%	Act%
265.70	100	100
268.00	64.40	60.73
264.00	65.30	66.16
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051120\
 Data File : BG045291.D
 Acq On : 11 May 2020 11:08
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 02:25:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 11 13:27:41 2020
 Response via : Initial Calibration



TIC: BG045291.D

(77) Benzidine

19.600min (-0.001) 319472.24ng

response 441583

Ion	Exp%	Act%
184.00	100	100
185.00	15.80	14.32
183.00	14.10	12.30
0.00	0.00	0.00