




Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051120\
Data File : BG045309.D
Acq On : 12 May 2020 00:04
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 02:19:28 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 11 13:27:41 2020
Response via : Initial Calibration



TIC: BG045309.D

(70) Pentachlorophenol (C)




17.022min (-0.007) 26459.74ng

response 55598

Ion	Exp%	Act%
265.70	100	100
268.00	64.40	60.45
264.00	65.30	65.09
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA G\DATA\BG051120\
 Data File : BG045309.D
 Acq On : 12 May 2020 00:04
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 12 02:19:28 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA G\METHODS\8270-BG041520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 11 13:27:41 2020
 Response via : Initial Calibration



TIC: BG045309.D

(77) Benzidine

19.595min (-0.007) 224122.50ng

response 411177

Ion	Exp%	Act%
184.00	100	100
185.00	15.80	14.33
183.00	14.10	12.81
0.00	0.00	0.00