

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051222\
 Data File : BG053528.D
 Acq On : 12 May 2022 9:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Manual Integrations

APPROVED

Reviewed By : Christian Giraldo 05/13/2022
 Supervised By : Jagrut Upadhyay 05/13/2022

Quant Time: May 13 06:35:48 2022

Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed May 11 06:00:46 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.088	152	23195	20.000	ng	0.00
21) Naphthalene-d8	10.896	136	101815	20.000	ng	0.00
39) Acenaphthene-d10	14.708	164	82292	20.000	ng	0.00
64) Phenanthrene-d10	17.457	188	204674	20.000	ng	0.00
76) Chrysene-d12	21.740	240	198449	20.000	ng	0.00
86) Perylene-d12	25.018	264	213930	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.650	112	109285	79.962	ng	0.00
7) Phenol-d6	7.248	99	174059	83.916	ng	0.00
23) Nitrobenzene-d5	9.251	82	192452	80.463	ng	0.00
42) 2,4,6-Tribromophenol	16.200	330	100481	82.788	ng	0.00
45) 2-Fluorobiphenyl	13.334	172	447904	79.535	ng	0.00
79) Terphenyl-d14	20.060	244	857800	80.506	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.529	88	28244	39.387	ng	99
3) Pyridine	3.929	79	72247	41.254	ng	94
4) n-Nitrosodimethylamine	3.846	42	34431	40.360	ng	# 97
6) Aniline	7.406	93	97476	42.430	ng	96
8) 2-Chlorophenol	7.653	128	56751	39.720	ng	96
9) Benzaldehyde	7.218	77	52737m	40.022	ng	
10) Phenol	7.277	94	83988	41.401	ng	96
11) bis(2-Chloroethyl)ether	7.500	93	61789	40.840	ng	91
12) 1,3-Dichlorobenzene	7.976	146	66931	39.709	ng	91
13) 1,4-Dichlorobenzene	8.123	146	67853	39.893	ng	98
14) 1,2-Dichlorobenzene	8.440	146	65509	41.015	ng	97
15) Benzyl Alcohol	8.323	79	75378	43.468	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.610	45	105656	42.907	ng	98
17) 2-Methylphenol	8.528	107	58391	42.586	ng	96
18) Hexachloroethane	9.169	117	24728	38.284	ng	97
19) n-Nitroso-di-n-propyla...	8.887	70	63489	42.661	ng	94
20) 3+4-Methylphenols	8.857	107	81103	41.882	ng	97
22) Acetophenone	8.904	105	108352	38.884	ng	# 97
24) Nitrobenzene	9.292	77	91867	39.684	ng	97
25) Isophorone	9.815	82	167627	41.643	ng	100
26) 2-Nitrophenol	10.009	139	38617	41.187	ng	98
27) 2,4-Dimethylphenol	10.061	122	57528	40.770	ng	93
28) bis(2-Chloroethoxy)met...	10.291	93	92641	40.016	ng	98
29) 2,4-Dichlorophenol	10.549	162	71655	42.076	ng	88
30) 1,2,4-Trichlorobenzene	10.761	180	77926	40.274	ng	97
31) Naphthalene	10.948	128	204621	39.509	ng	98
32) Benzoic acid	10.232	122	28013m	34.631	ng	
33) 4-Chloroaniline	11.054	127	89634	41.326	ng	97
34) Hexachlorobutadiene	11.230	225	58203	40.566	ng	95
35) Caprolactam	11.824	113	25230	40.769	ng	# 88
36) 4-Chloro-3-methylphenol	12.176	107	85915	42.176	ng	100
37) 2-Methylnaphthalene	12.546	142	157732	40.810	ng	97
38) 1-Methylnaphthalene	12.764	142	153058m	40.589	ng	
40) 1,2,4,5-Tetrachloroben...	12.916	216	103980	40.086	ng	99
41) Hexachlorocyclopentadiene	12.893	237	31738	35.557	ng	92
43) 2,4,6-Trichlorophenol	13.151	196	70235	41.481	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051222\
 Data File : BG053528.D
 Acq On : 12 May 2022 9:55
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 SSTDCCC040


Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 05/13/2022
 Supervised By : Jagrut Upadhyay 05/13/2022

Quant Time: May 13 06:35:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG050522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 11 06:00:46 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	13.234	196	73694	40.894	ng	97
46) 1,1'-Biphenyl	13.545	154	221388	39.468	ng	97
47) 2-Chloronaphthalene	13.592	162	171726	39.075	ng	99
48) 2-Nitroaniline	13.792	65	69360	42.355	ng	96
49) Acenaphthylene	14.438	152	280916	40.057	ng	99
50) Dimethylphthalate	14.162	163	248450	39.730	ng	99
51) 2,6-Dinitrotoluene	14.279	165	51420	40.360	ng	97
52) Acenaphthene	14.779	154	166862	39.925	ng	96
53) 3-Nitroaniline	14.614	138	56609	41.123	ng	97
54) 2,4-Dinitrophenol	14.831	184	27691	37.813	ng	91
55) Dibenzofuran	15.108	168	299872	39.946	ng	98
56) 4-Nitrophenol	14.931	139	40758	41.454	ng	92
57) 2,4-Dinitrotoluene	15.072	165	74516	40.351	ng	# 83
58) Fluorene	15.760	166	242337	40.227	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.337	232	73241	40.057	ng	96
60) Diethylphthalate	15.519	149	260053	40.258	ng	98
61) 4-Chlorophenyl-phenyle...	15.742	204	137007	40.757	ng	99
62) 4-Nitroaniline	15.777	138	58929	41.273	ng	96
63) Azobenzene	16.036	77	274411	41.485	ng	98
65) 4,6-Dinitro-2-methylph...	15.842	198	51129	41.923	ng	100
66) n-Nitrosodiphenylamine	15.959	169	218499	40.660	ng	97
67) 4-Bromophenyl-phenylether	16.641	248	98130	41.052	ng	96
68) Hexachlorobenzene	16.770	284	104818	40.039	ng	98
69) Atrazine	16.899	200	88409	39.080	ng	97
70) Pentachlorophenol	17.117	266	51521	39.070	ng	94
71) Phenanthrene	17.498	178	420666	40.420	ng	98
72) Anthracene	17.592	178	416867	40.752	ng	99
73) Carbazole	17.857	167	396657	39.565	ng	98
74) Di-n-butylphthalate	18.403	149	449227	39.842	ng	99
75) Fluoranthene	19.507	202	533573	39.596	ng	98
77) Benzidine	19.678	184	170685	39.905	ng	99
78) Pyrene	19.866	202	527509	41.062	ng	99
80) Butylbenzylphthalate	20.741	149	193067	40.984	ng	98
81) Benzo(a)anthracene	21.716	228	517268	40.867	ng	99
82) 3,3'-Dichlorobenzidine	21.628	252	130886	40.777	ng	95
83) Chrysene	21.787	228	480661	40.659	ng	100
84) Bis(2-ethylhexyl)phtha...	21.605	149	274131	42.161	ng	98
85) Di-n-octyl phthalate	22.832	149	463100	42.120	ng	100
87) Indeno(1,2,3-cd)pyrene	28.771	276	610434	40.969	ng	# 96
88) Benzo(b)fluoranthene	23.972	252	514432	40.625	ng	99
89) Benzo(k)fluoranthene	24.043	252	495168	39.794	ng	100
90) Benzo(a)pyrene	24.859	252	435596	40.415	ng	99
91) Dibenzo(a,h)anthracene	28.836	278	496674	40.455	ng	99
92) Benzo(g,h,i)perylene	29.952	276	497407	40.872	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 8.088 min Scan# 80
Delta R.T. -0.001 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:152 Resp: 2319
Ion Ratio Lower Upper
152 100
150 137.4 131.6 197.4
115 61.1 50.9 76.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#2
1,4-Dioxane
Concen: 39.387 ng
RT: 3.529 min Scan# 24
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 88 Resp: 28244
Ion Ratio Lower Upper
88 100
58 73.7 59.2 88.8
43 31.0 24.3 36.5









#3
Pyridine
Concen: 41.254 ng
RT: 3.929 min Scan# 91
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 79 Resp: 7224
Ion Ratio Lower Upper
79 100
52 73.4 54.6 81.8
51 37.8 32.5 48.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#4
n-Nitrosodimethylamine
Concen: 40.360 ng
RT: 3.846 min Scan# 78
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 42 Resp: 34431
Ion Ratio Lower Upper
42 100
74 118.1 97.0 145.4
44 5.6 7.0 10.4





#5
2-Fluorophenol
Concen: 79.962 ng
RT: 5.650 min Scan# 388
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:112 Resp: 109285
Ion Ratio Lower Upper
112 100
64 76.9 55.1 82.7
63 39.3 30.4 45.6

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#6
Aniline
Concen: 42.430 ng
RT: 7.406 min Scan# 684
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 93 Resp: 97476
Ion Ratio Lower Upper
93 100
66 44.1 32.9 49.3
65 23.1 17.8 26.8









#7
Phenol-d6
Concen: 83.916 ng
RT: 7.248 min Scan# 61
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 174059
Ion Ratio Lower Upper
99 100
42 21.6 17.0 25.6
71 44.5 33.8 50.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#8
2-Chlorophenol
Concen: 39.720 ng
RT: 7.653 min Scan# 726
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:128 Resp: 56751
Ion Ratio Lower Upper
128 100
130 31.6 10.5 50.5
64 58.5 35.1 75.1









#9
Benzaldehyde
Concen: 40.022 ng m
RT: 7.218 min Scan# 61
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 5273
Ion Ratio Lower Upper
77 100
105 82.4 59.6 99.6
106 81.0 56.8 96.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#10
Phenol
Concen: 41.401 ng
RT: 7.277 min Scan# 662
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 94 Resp: 83988
Ion Ratio Lower Upper
94 100
65 35.5 14.8 54.8
66 49.5 25.2 65.2









#11
bis(2-Chloroethyl)ether
Concen: 40.840 ng
RT: 7.500 min Scan# 702
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion: 93 Resp: 61789
Ion Ratio Lower Upper
93 100
63 93.6 63.6 103.6
95 32.6 10.1 50.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#12
1,3-Dichlorobenzene
Concen: 39.709 ng
RT: 7.976 min Scan# 781
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:146 Resp: 66931
Ion Ratio Lower Upper
146 100
148 64.0 46.1 69.1
75 40.8 27.6 41.4





#13

1,4-Dichlorobenzene

Concen: 39.893 ng

RT: 8.123 min Scan# 80

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:146 Resp: 6785

Ion Ratio Lower Upper

146 100

148 62.8 51.5 77.3





111 43.9 33.9 50.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/13/2022

Supervised By :Jagrut Upadhyay 05/13/2022



#14

1,2-Dichlorobenzene

Concen: 41.015 ng

RT: 8.440 min Scan# 860

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55




Tgt Ion:146 Resp: 65509


Ion Ratio Lower Upper

146 100

148 62.4 52.6 78.8

111 43.6 34.5 51.7





#15
Benzyl Alcohol
Concen: 43.468 ng
RT: 8.323 min Scan# 842
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion: 79 Resp: 75378
Ion Ratio Lower Upper
79 100
108 63.4 53.7 80.5
77 63.2 52.2 78.4

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#16
2,2'-oxybis(1-Chloropropane)
Concen: 42.907 ng
RT: 8.610 min Scan# 889
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 45 Resp: 105656
Ion Ratio Lower Upper
45 100
77 14.0 0.0 32.7
79 9.8 0.0 29.7









#17
2-Methylphenol
Concen: 42.586 ng
RT: 8.528 min Scan# 81
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 5839
Ion Ratio Lower Upper
107 100
108 111.6 85.4 128.0
77 58.7 46.2 69.2
79 54.2 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#18
Hexachloroethane
Concen: 38.284 ng
RT: 9.169 min Scan# 984
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:117 Resp: 24728
Ion Ratio Lower Upper
117 100
119 99.8 75.0 112.6
201 111.7 89.0 133.6






#19
n-Nitroso-di-n-propylamine
Concen: 42.661 ng
RT: 8.887 min Scan# 91
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion: 70 Resp: 63489
Ion Ratio Lower Upper
70 100
42 64.4 47.3 70.9
101 9.1 7.0 10.4
130 17.2 14.6 22.0


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#20
3+4-Methylphenols
Concen: 41.882 ng
RT: 8.857 min Scan# 931
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:107 Resp: 81103
Ion Ratio Lower Upper
107 100
108 90.1 67.5 107.5
77 42.5 20.3 60.3
79 34.4 16.7 56.7





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.896 min Scan# 1018
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:136 Resp: 101815
Ion Ratio Lower Upper
136 100
137 10.9 8.5 12.7
54 9.6 7.4 11.2
68 5.9 4.9 7.3

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#22
Acetophenone
Concen: 38.884 ng
RT: 8.904 min Scan# 939
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55


Tgt Ion:105 Resp: 108352
Ion Ratio Lower Upper
105 100
71 2.6 1.1 1.7#
51 33.2 25.4 38.0
120 19.5 17.3 25.9





#23
Nitrobenzene-d5
Concen: 80.463 ng
RT: 9.251 min Scan# 91
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion: 82 Resp: 19245
Ion Ratio Lower Upper
82 100
128 35.4 27.9 41.9
54 52.4 41.2 61.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#24
Nitrobenzene
Concen: 39.684 ng
RT: 9.292 min Scan# 1005
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion: 77 Resp: 91867
Ion Ratio Lower Upper
77 100
123 36.6 27.4 41.0
65 14.1 11.4 17.0





#25
Isophorone
Concen: 41.643 ng
RT: 9.815 min Scan# 1094
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G





ClientSampleId :

SSTDCCC040

Tgt Ion: 82 Resp: 16762
Ion Ratio Lower Upper
82 100
95 7.7 6.2 9.4
138 14.7 11.6 17.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#26
2-Nitrophenol
Concen: 41.187 ng
RT: 10.009 min Scan# 1127
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:139 Resp: 38617
Ion Ratio Lower Upper
139 100
109 45.5 36.1 54.1
65 71.8 55.8 83.6





#27
2,4-Dimethylphenol
Concen: 40.770 ng
RT: 10.061 min Scan# 1136
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:122 Resp: 57528
Ion Ratio Lower Upper
122 100
107 127.2 109.4 164.0
121 56.1 48.7 73.1

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#28
bis(2-Chloroethoxy)methane
Concen: 40.016 ng
RT: 10.291 min Scan# 1175
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 93 Resp: 92641
Ion Ratio Lower Upper
93 100
95 34.7 26.9 40.3
123 10.3 7.7 11.5





#29

2,4-Dichlorophenol

Concen: 42.076 ng

RT: 10.549 min Scan# 1219

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040





Tgt Ion:	162	Resp:	7165
Ion Ratio	Lower	Upper	
162	100		
164	57.6	46.6	86.6
98	34.9	22.8	62.8

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/13/2022

Supervised By :Jagrut Upadhyay 05/13/2022



#30

1,2,4-Trichlorobenzene

Concen: 40.274 ng




RT: 10.761 min Scan# 1255


Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55


Tgt Ion:	180	Resp:	77926
Ion Ratio	Lower	Upper	
180	100		
182	98.2	76.6	115.0
145	28.3	25.0	37.4





#31
Naphthalene
Concen: 39.509 ng
RT: 10.948 min Scan# 1165
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:128 Resp: 20462
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 13.1 11.1 16.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#32
Benzoic acid
Concen: 34.631 ng m
RT: 10.232 min Scan# 1165
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:122 Resp: 28013
Ion Ratio Lower Upper
122 100
105 133.3 111.8 151.8
77 105.5 109.4 149.4#






#33
4-Chloroaniline
Concen: 41.326 ng
RT: 11.054 min Scan# 1105
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion	Ratio	Lower	Upper
127	100		
129	34.0	26.3	39.5
65	38.0	31.1	46.7
92	20.2	19.1	28.7


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#34
Hexachlorobutadiene
Concen: 40.566 ng
RT: 11.230 min Scan# 1335
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion	Ratio	Lower	Upper
225	100		
223	61.9	48.9	73.3
227	59.6	52.8	79.2






#35
Caprolactam
Concen: 40.769 ng
RT: 11.824 min Scan# 1436
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:113 Resp: 25230
Ion Ratio Lower Upper
113 100
55 214.9 169.8 209.8
56 152.9 141.2 181.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#36
4-Chloro-3-methylphenol
Concen: 42.176 ng
RT: 12.176 min Scan# 1496
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:107 Resp: 85915
Ion Ratio Lower Upper
107 100
144 23.1 18.5 27.7
142 69.2 55.0 82.4









#37
2-Methylnaphthalene
Concen: 40.810 ng
RT: 12.546 min Scan# 1559
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:142 Resp: 15773
Ion Ratio Lower Upper
142 100
141 87.2 72.3 108.5
115 44.3 33.7 50.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#38
1-Methylnaphthalene
Concen: 40.589 ng m
RT: 12.764 min Scan# 1596
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:142 Resp: 153058
Ion Ratio Lower Upper
142 100
141 92.6 76.4 114.6
116 5.1 3.3 4.9#






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.708 min Scan# 1929
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:164 Resp: 82292
Ion Ratio Lower Upper
164 100
162 101.5 82.3 123.5
160 43.5 34.2 51.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#40
1,2,4,5-Tetrachlorobenzene
Concen: 40.086 ng
RT: 12.916 min Scan# 1622
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:216 Resp: 103980
Ion Ratio Lower Upper
216 100
214 78.1 62.9 94.3
179 20.0 16.1 24.1
108 16.6 12.2 18.2





#41
Hexachlorocyclopentadiene
Concen: 35.557 ng
RT: 12.893 min Scan# 10
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:237 Resp: 31738
Ion Ratio Lower Upper
237 100
235 70.1 43.5 83.5
272 12.9 0.0 30.1

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#42
2,4,6-Tribromophenol
Concen: 82.788 ng
RT: 16.200 min Scan# 2181
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:330 Resp: 100481
Ion Ratio Lower Upper
330 100
332 97.9 76.0 114.0
141 34.5 27.1 40.7





#43
2,4,6-Trichlorophenol
Concen: 41.481 ng
RT: 13.151 min Scan# 1662
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:196 Resp: 7023
Ion Ratio Lower Upper
196 100
198 93.0 74.1 111.1
200 28.2 24.7 37.1

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#44
2,4,5-Trichlorophenol
Concen: 40.894 ng
RT: 13.234 min Scan# 1676
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55


Tgt Ion:196 Resp: 73694
Ion Ratio Lower Upper
196 100
198 91.7 76.8 115.2
97 50.3 39.8 59.6
132 24.6 18.4 27.6





#45
2-Fluorobiphenyl
Concen: 79.535 ng
RT: 13.334 min Scan# 1729
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:172 Resp: 447904
Ion Ratio Lower Upper
172 100
171 34.8 28.5 42.7
170 22.9 18.9 28.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#46
1,1'-Biphenyl
Concen: 39.468 ng
RT: 13.545 min Scan# 1729
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:154 Resp: 221388
Ion Ratio Lower Upper
154 100
153 39.4 21.6 61.6
76 12.6 0.0 32.6





#47
2-Chloronaphthalene
Concen: 39.075 ng
RT: 13.592 min Scan# 1737
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:162 Resp: 171720
Ion Ratio Lower Upper
162 100
127 39.6 30.6 45.8
164 33.3 26.5 39.7

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#48
2-Nitroaniline
Concen: 42.355 ng
RT: 13.792 min Scan# 1771
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion: 65 Resp: 69360
Ion Ratio Lower Upper
65 100
92 63.6 51.5 77.3
138 82.8 71.3 106.9






#49
Acenaphthylene
Concen: 40.057 ng
RT: 14.438 min Scan# 1881
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:152 Resp: 280910
Ion Ratio Lower Upper
152 100
151 20.1 16.4 24.6
153 12.9 10.8 16.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#50
Dimethylphthalate
Concen: 39.730 ng
RT: 14.162 min Scan# 1834
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:163 Resp: 248450
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.6
164 10.4 8.0 12.0






#51
2,6-Dinitrotoluene
Concen: 40.360 ng
RT: 14.279 min Scan# 1854
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:165 Resp: 51420
Ion Ratio Lower Upper
165 100
63 70.6 54.1 81.1
89 70.0 54.8 82.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#52
Acenaphthene
Concen: 39.925 ng
RT: 14.779 min Scan# 1939
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Tgt Ion:154 Resp: 166862
Ion Ratio Lower Upper
154 100
153 120.6 92.7 139.1
152 55.5 42.6 64.0





#53
3-Nitroaniline
Concen: 41.123 ng
RT: 14.614 min Scan# 1911
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:138 Resp: 56609
Ion Ratio Lower Upper
138 100
108 10.7 9.2 13.8
92 123.5 101.9 152.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#54
2,4-Dinitrophenol
Concen: 37.813 ng
RT: 14.831 min Scan# 1948
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:184 Resp: 27691
Ion Ratio Lower Upper
184 100
63 66.1 61.1 91.7
154 71.1 60.6 91.0






#55
Dibenzofuran
Concen: 39.946 ng
RT: 15.108 min Scan# 1995
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:168 Resp: 299872
Ion Ratio Lower Upper
168 100
139 39.0 30.5 45.7
169 14.1 10.3 15.5


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#56
4-Nitrophenol
Concen: 41.454 ng
RT: 14.931 min Scan# 1965
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:139 Resp: 40758
Ion Ratio Lower Upper
139 100
109 104.1 88.5 128.5
65 125.8 93.9 133.9





#57
2,4-Dinitrotoluene
Concen: 40.351 ng
RT: 15.072 min Scan# 1991
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:165 Resp: 74510
Ion Ratio Lower Upper
165 100
63 56.1 34.6 52.0
89 88.9 60.3 90.5
182 2.4 2.2 3.4

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#58
Fluorene
Concen: 40.227 ng
RT: 15.760 min Scan# 2106
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55


Tgt Ion:166 Resp: 242337
Ion Ratio Lower Upper
166 100
165 95.9 75.0 112.6
167 13.5 10.8 16.2





#59
2,3,4,6-Tetrachlorophenol
Concen: 40.057 ng
RT: 15.337 min Scan# 2034
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:232 Resp: 7324
Ion Ratio Lower Upper
232 100
131 40.8 30.6 46.0
130 2.2 1.8 2.8
166 28.9 21.1 31.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#60
Diethylphthalate
Concen: 40.258 ng
RT: 15.519 min Scan# 2065
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:149 Resp: 260053
Ion Ratio Lower Upper
149 100
177 23.5 18.4 27.6
150 12.6 11.2 16.8









#61
4-Chlorophenyl-phenylether
Concen: 40.757 ng
RT: 15.742 min Scan# 2105
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:204 Resp: 13700
Ion Ratio Lower Upper
204 100
206 33.2 27.5 41.3
141 56.5 45.3 67.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#62
4-Nitroaniline
Concen: 41.273 ng
RT: 15.777 min Scan# 2109
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:138 Resp: 58929
Ion Ratio Lower Upper
138 100
92 53.8 40.5 80.5
108 125.2 103.7 143.7






#63
Azobenzene
Concen: 41.485 ng
RT: 16.036 min Scan# 2153
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion: 77 Resp: 27441
Ion Ratio Lower Upper
77 100
182 22.4 3.8 43.8
105 17.1 0.0 37.8
51 30.1 10.7 50.7


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.457 min Scan# 2395
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:188 Resp: 204674
Ion Ratio Lower Upper
188 100
94 6.5 5.4 8.0
80 7.5 6.8 10.2





#65
4,6-Dinitro-2-methylphenol
Concen: 41.923 ng
RT: 15.842 min Scan# 2112
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:198 Resp: 51129
Ion Ratio Lower Upper
198 100
51 47.5 28.1 68.1
105 40.3 20.3 60.3

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#66
n-Nitrosodiphenylamine
Concen: 40.660 ng
RT: 15.959 min Scan# 2140
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55


Tgt Ion:169 Resp: 218499
Ion Ratio Lower Upper
169 100
168 67.1 55.9 83.9
167 37.7 30.6 46.0





#67
4-Bromophenyl-phenylether
Concen: 41.052 ng
RT: 16.641 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:248 Resp: 98130
Ion Ratio Lower Upper
248 100
250 99.3 75.2 112.8
141 64.2 50.0 75.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#68
Hexachlorobenzene
Concen: 40.039 ng
RT: 16.770 min Scan# 2278
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:284 Resp: 104818
Ion Ratio Lower Upper
284 100
142 28.3 23.6 35.4
249 31.6 25.8 38.6





#69
 Atrazine
 Concen: 39.080 ng
 RT: 16.899 min Scan# 21
 Delta R.T. -0.008 min
 Lab File: BG053528.D
 Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :





SSTDCCC040

Tgt Ion:200 Resp: 88409
 Ion Ratio Lower Upper
 200 100
 173 22.1 3.7 43.7
 215 47.1 25.2 65.2

Manual Integrations




APPROVED


Reviewed By :Christian Giraldo 05/13/2022
 Supervised By :Jagrut Upadhyay 05/13/2022



#70
 Pentachlorophenol
 Concen: 39.070 ng
 RT: 17.117 min Scan# 2337
 Delta R.T. -0.002 min
 Lab File: BG053528.D
 Acq: 12 May 2022 9:55

Tgt Ion:266 Resp: 51521
 Ion Ratio Lower Upper
 266 100
 268 64.4 56.2 84.2
 264 62.6 53.0 79.4









#71
Phenanthrene
Concen: 40.420 ng
RT: 17.498 min Scan# 2404
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :
BNA_G
ClientSampleId :
SSTDCCC040

Tgt Ion:178 Resp: 420660
Ion Ratio Lower Upper
178 100
176 18.2 14.1 21.1
179 15.8 12.1 18.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#72
Anthracene
Concen: 40.752 ng
RT: 17.592 min Scan# 2418
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:178 Resp: 416867
Ion Ratio Lower Upper
178 100
176 18.1 15.2 22.8
179 16.0 13.0 19.4






#73
Carbazole
Concen: 39.565 ng
RT: 17.857 min Scan# 2463
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:167 Resp: 39665
Ion Ratio Lower Upper
167 100
166 21.4 18.2 27.2
139 12.2 10.1 15.1


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#74
Di-n-butylphthalate
Concen: 39.842 ng
RT: 18.403 min Scan# 2556
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Tgt Ion:149 Resp: 449227
Ion Ratio Lower Upper
149 100
150 9.3 7.7 11.5
104 6.9 5.4 8.0





#75
Fluoranthene
Concen: 39.596 ng
RT: 19.507 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 53357
Ion Ratio Lower Upper
202 100
101 6.7 0.0 26.9
203 18.2 0.0 37.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.740 min Scan# 3124
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Tgt Ion:240 Resp: 198449
Ion Ratio Lower Upper
240 100
120 6.1 4.6 7.0
236 24.7 20.2 30.4





#77
Benzidine
Concen: 39.905 ng
RT: 19.678 min Scan# 21
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:184 Resp: 170685
Ion Ratio Lower Upper
184 100
185 14.9 11.8 17.6
183 11.8 8.9 13.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#78
Pyrene
Concen: 41.062 ng
RT: 19.866 min Scan# 2805
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:202 Resp: 527509
Ion Ratio Lower Upper
202 100
200 21.1 16.6 24.8
203 18.9 14.6 21.8





#79

Terphenyl-d14

Concen: 80.506 ng

RT: 20.060 min Scan# 21

Delta R.T. -0.002 min

Lab File: BG053528.D


Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040



Tgt Ion:244 Resp: 857800

Ion Ratio Lower Upper

244 100

212 7.7 6.5 9.7




122 6.3 5.0 7.4

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/13/2022

Supervised By :Jagrut Upadhyay 05/13/2022



#80

Butylbenzylphthalate

Concen: 40.984 ng

RT: 20.741 min Scan# 2954

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55




Tgt Ion:149 Resp: 193067


Ion Ratio Lower Upper

149 100

91 76.8 59.6 89.4


206 23.1 18.6 28.0





#81
Benzo(a)anthracene
Concen: 40.867 ng
RT: 21.716 min Scan# 3122
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:228 Resp: 51726
Ion Ratio Lower Upper
228 100
226 27.5 21.6 32.4
229 20.0 16.2 24.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#82
3,3'-Dichlorobenzidine
Concen: 40.777 ng
RT: 21.628 min Scan# 3105
Delta R.T. 0.004 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Tgt Ion:252 Resp: 130886
Ion Ratio Lower Upper
252 100
254 66.6 49.8 74.8
126 6.2 6.0 9.0





#83
Chrysene
Concen: 40.659 ng
RT: 21.787 min Scan# 31
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:228 Resp: 48066
Ion Ratio Lower Upper
228 100
226 29.5 23.4 35.2
229 19.5 15.5 23.3




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#84
Bis(2-ethylhexyl)phthalate
Concen: 42.161 ng
RT: 21.605 min Scan# 3101
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:149 Resp: 274131
Ion Ratio Lower Upper
149 100
167 26.3 20.2 30.4
279 4.7 3.6 5.4





#85

Di-n-octyl phthalate

Concen: 42.120 ng

RT: 22.832 min Scan# 31

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040

Tgt Ion:149 Resp: 463100

Ion Ratio Lower Upper

149 100

167 1.6 1.3 1.9





43 11.1 9.0 13.4

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 05/13/2022

Supervised By :Jagrut Upadhyay 05/13/2022



#86

Perylene-d12

Concen: 20.000 ng

RT: 25.018 min Scan# 3682

Delta R.T. -0.002 min

Lab File: BG053528.D

Acq: 12 May 2022 9:55




Tgt Ion:264 Resp: 213930


Ion Ratio Lower Upper

264 100

260 23.5 19.0 28.6


265 21.0 17.4 26.2





#87
Indeno(1,2,3-cd)pyrene
Concen: 40.969 ng
RT: 28.771 min Scan# 41
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040



Tgt Ion:276 Resp: 610434
Ion Ratio Lower Upper
276 100
138 0.0 5.0 7.4
277 25.4 20.3 30.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#88
Benzo(b)fluoranthene
Concen: 40.625 ng
RT: 23.972 min Scan# 3504
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Tgt Ion:252 Resp: 514432
Ion Ratio Lower Upper
252 100
253 22.9 18.0 27.0
125 5.8 4.6 6.8






#89
Benzo(k)fluoranthene
Concen: 39.794 ng
RT: 24.043 min Scan# 3516
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55

Instrument :

BNA_G

ClientSampleId :

SSTDCCC040






Tgt Ion:252 Resp: 495168
Ion Ratio Lower Upper
252 100
253 23.2 18.4 27.6
125 5.9 4.8 7.2


Manual Integrations

APPROVED



Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022




#90
Benzo(a)pyrene
Concen: 40.415 ng
RT: 24.859 min Scan# 3655
Delta R.T. -0.008 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Tgt Ion:252 Resp: 435596
Ion Ratio Lower Upper
252 100
253 21.8 17.8 26.8
125 6.3 4.3 6.5





#91
Dibenzo(a,h)anthracene
Concen: 40.455 ng
RT: 28.836 min Scan# 41
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55




Instrument :
BNA_G
ClientSampleId :
SSTDCCC040




Tgt Ion:278 Resp: 496674
Ion Ratio Lower Upper
278 100
139 8.4 7.2 10.8
279 23.7 18.6 27.8

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 05/13/2022
Supervised By :Jagrut Upadhyay 05/13/2022



#92
Benzo(g,h,i)perylene
Concen: 40.872 ng
RT: 29.952 min Scan# 4522
Delta R.T. -0.002 min
Lab File: BG053528.D
Acq: 12 May 2022 9:55



Tgt Ion:276 Resp: 497407
Ion Ratio Lower Upper
276 100
277 22.5 17.3 25.9
138 11.9 9.6 14.4

