

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051223\
 Data File : BG057422.D
 Acq On : 13 May 2023 1:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 03:13:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 22:04:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.483	0.471	2.5	113	0.00
3	Pyridine	1.554	1.467	5.6	104	0.00
4	n-Nitrosodimethylamine	0.730	0.653	10.5	99	0.00
5 S	2-Fluorophenol	1.169	1.176	-0.6	113	0.00
6	Aniline	2.252	2.122	5.8	105	0.00
7 S	Phenol-d6	1.775	1.747	1.6	111	0.00
8	2-Chlorophenol	1.235	1.208	2.2	111	0.00
9	Benzaldehyde	1.042	0.871	16.4	106	0.00
10 C	Phenol	1.731	1.700	1.8	109	0.00
11	bis(2-Chloroethyl)ether	1.305	1.216	6.8	106	0.00
12	1,3-Dichlorobenzene	1.364	1.327	2.7	111	0.00
13 C	1,4-Dichlorobenzene	1.370	1.358	0.9	112	0.00
14	1,2-Dichlorobenzene	1.322	1.275	3.6	111	0.00
15	Benzyl Alcohol	1.479	1.384	6.4	101	0.00
16	2,2'-oxybis(1-Chloropropane	1.645	1.431	13.0	99	0.00
17	2-Methylphenol	1.248	1.220	2.2	109	0.00
18	Hexachloroethane	0.494	0.478	3.2	111	0.00
19 P	n-Nitroso-di-n-propylamine	1.354	1.211	10.6	101	0.00
20	3+4-Methylphenols	1.708	1.690	1.1	111	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
22	Acetophenone	0.572	0.531	7.2	105	0.00
23 S	Nitrobenzene-d5	0.477	0.441	7.5	103	0.00
24	Nitrobenzene	0.481	0.444	7.7	103	0.00
25	Isophorone	0.924	0.829	10.3	101	0.00
26 C	2-Nitrophenol	0.184	0.195	-6.0	117	0.00
27	2,4-Dimethylphenol	0.322	0.325	-0.9	110	0.00
28	bis(2-Chloroethoxy)methane	0.463	0.426	8.0	103	0.00
29 C	2,4-Dichlorophenol	0.342	0.353	-3.2	115	0.00
30	1,2,4-Trichlorobenzene	0.394	0.385	2.3	111	0.00
31	Naphthalene	1.069	1.036	3.1	109	0.00
32	Benzoic acid	0.170	0.189	-11.2	117	0.00
33	4-Chloroaniline	0.480	0.489	-1.9	112	0.00
34 C	Hexachlorobutadiene	0.293	0.282	3.8	110	0.00
35	Caprolactam	0.117	0.119	-1.7	111	0.00
36 C	4-Chloro-3-methylphenol	0.389	0.399	-2.6	115	0.00
37	2-Methylnaphthalene	0.841	0.838	0.4	112	0.00
38	1-Methylnaphthalene	0.792	0.778	1.8	111	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	0.710	0.718	-1.1	112	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.321	-11.8	117	0.00
42 S	2,4,6-Tribromophenol	0.284	0.306	-7.7	121	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.455	-5.1	114	0.00
44	2,4,5-Trichlorophenol	0.509	0.521	-2.4	111	0.00
45 S	2-Fluorobiphenyl	1.485	1.463	1.5	110	0.00
46	1,1'-Biphenyl	1.484	1.474	0.7	111	0.00
47	2-Chloronaphthalene	1.104	1.102	0.2	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051223\
 Data File : BG057422.D
 Acq On : 13 May 2023 1:13
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 13 03:13:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 22:04:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.375	0.371	1.1	108	0.00
49	Acenaphthylene	1.793	1.790	0.2	111	0.00
50	Dimethylphthalate	1.644	1.564	4.9	109	0.00
51	2,6-Dinitrotoluene	0.336	0.342	-1.8	113	0.00
52 C	Acenaphthene	1.120	1.114	0.5	111	0.00
53	3-Nitroaniline	0.336	0.346	-3.0	115	0.00
54 P	2,4-Dinitrophenol	0.182	0.189	-3.8	109	0.00
55	Dibenzofuran	1.882	1.859	1.2	111	0.00
56 P	4-Nitrophenol	0.224	0.226	-0.9	105	0.00
57	2,4-Dinitrotoluene	0.480	0.505	-5.2	117	0.00
58	Fluorene	1.551	1.538	0.8	112	0.00
59	2,3,4,6-Tetrachlorophenol	0.461	0.482	-4.6	114	0.00
60	Diethylphthalate	1.669	1.578	5.5	107	0.00
61	4-Chlorophenyl-phenylether	0.934	0.922	1.3	111	0.00
62	4-Nitroaniline	0.339	0.349	-2.9	112	0.00
63	Azobenzene	1.610	1.463	9.1	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.132	-10.9	118	0.00
66 c	n-Nitrosodiphenylamine	0.543	0.534	1.7	112	0.00
67	4-Bromophenyl-phenylether	0.244	0.243	0.4	114	0.00
68	Hexachlorobenzene	0.237	0.245	-3.4	119	0.00
69	Atrazine	0.207	0.208	-0.5	116	0.00
70 C	Pentachlorophenol	0.137	0.139	-1.5	111	0.00
71	Phenanthrene	1.025	1.003	2.1	113	0.00
72	Anthracene	1.052	1.024	2.7	112	0.00
73	Carbazole	0.885	0.867	2.0	113	0.00
74	Di-n-butylphthalate	1.019	1.024	-0.5	114	0.00
75 C	Fluoranthene	1.275	1.257	1.4	113	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	121	0.00
77	Benzidine	0.447	0.422	5.6	129	0.00
78	Pyrene	1.453	1.378	5.2	114	0.00
79 S	Terphenyl-d14	1.226	1.156	5.7	114	0.00
80	Butylbenzylphthalate	0.461	0.481	-4.3	121	0.00
81	Benzo(a)anthracene	1.425	1.395	2.1	115	0.00
82	3,3'-Dichlorobenzidine	0.456	0.463	-1.5	118	0.00
83	Chrysene	1.341	1.314	2.0	116	0.00
84	Bis(2-ethylhexyl)phthalate	0.682	0.690	-1.2	117	0.00
85 c	Di-n-octyl phthalate	1.137	1.173	-3.2	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	1.458	1.417	2.8	118	0.00
88	Benzo(b)fluoranthene	1.307	1.256	3.9	116	0.00
89	Benzo(k)fluoranthene	1.329	1.251	5.9	115	0.00
90 C	Benzo(a)pyrene	1.277	1.216	4.8	116	-0.01
91	Dibenzo(a,h)anthracene	1.214	1.184	2.5	118	-0.01
92	Benzo(g,h,i)perylene	1.185	1.152	2.8	118	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051223\
Data File : BG057422.D
Acq On : 13 May 2023 1:13
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 13 03:13:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 12 22:04:03 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0