

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051315\
 Data File : BG017117.D
 Acq On : 13 May 2015 20:57
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 14 04:02:52 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 16:12:13 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.445	0.490	-10.1	129	0.00
3	Pyridine	1.346	1.468	-9.1	121	0.00
4	n-Nitrosodimethylamine	1.021	0.999	2.2	108	0.00
5 S	2-Fluorophenol	1.127	1.149	-2.0	112	0.00
6	Aniline	2.189	2.396	-9.5	120	0.00
7 S	Phenol-d6	1.582	1.703	-7.6	118	0.00
8	2-Chlorophenol	1.349	1.375	-1.9	113	0.00
9	Benzaldehyde	1.041	1.079	-3.7	117	0.00
10 C	Phenol	1.678	1.830	-9.1	121	0.00
11	bis(2-Chloroethyl)ether	1.258	1.367	-8.7	118	0.00
12	1,3-Dichlorobenzene	1.506	1.452	3.6	111	0.00
13 C	1,4-Dichlorobenzene	1.520	1.505	1.0	112	0.00
14	1,2-Dichlorobenzene	1.452	1.483	-2.1	114	0.00
15	Benzyl Alcohol	1.537	1.608	-4.6	116	0.00
16	2,2'-oxybis(1-Chloropropane	2.040	2.243	-10.0	124	0.00
17	2-Methylphenol	1.216	1.270	-4.4	119	0.00
18	Hexachloroethane	0.635	0.632	0.5	112	0.00
19 P	n-Nitroso-di-n-propylamine	1.378	1.467	-6.5	118	0.00
20	3+4-Methylphenols	1.691	1.775	-5.0	115	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
22	Acetophenone	0.485	0.494	-1.9	115	0.00
23 S	Nitrobenzene-d5	0.428	0.441	-3.0	114	0.00
24	Nitrobenzene	0.419	0.433	-3.3	118	0.00
25	Isophorone	0.790	0.835	-5.7	117	0.00
26 C	2-Nitrophenol	0.187	0.195	-4.3	116	0.00
27	2,4-Dimethylphenol	0.308	0.310	-0.6	115	0.00
28	bis(2-Chloroethoxy)methane	0.384	0.396	-3.1	116	0.00
29 C	2,4-Dichlorophenol	0.291	0.298	-2.4	113	0.00
30	1,2,4-Trichlorobenzene	0.330	0.322	2.4	109	0.00
31	Naphthalene	0.979	0.985	-0.6	112	0.00
32	Benzoic acid	0.209	0.223	-6.7	118	0.00
33	4-Chloroaniline	0.463	0.474	-2.4	112	0.00
34 C	Hexachlorobutadiene	0.209	0.204	2.4	109	0.00
35	Caprolactam	0.146	0.158	-8.2	126	0.00
36 C	4-Chloro-3-methylphenol	0.368	0.393	-6.8	118	0.00
37	2-Methylnaphthalene	0.776	0.771	0.6	115	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
39	1,2,4,5-Tetrachlorobenzene	0.566	0.570	-0.7	110	0.00
40 P	Hexachlorocyclopentadiene	0.345	0.343	0.6	109	0.00
41 S	2,4,6-Tribromophenol	0.242	0.242	0.0	110	0.00
42 C	2,4,6-Trichlorophenol	0.401	0.411	-2.5	109	0.00
43	2,4,5-Trichlorophenol	0.414	0.432	-4.3	114	0.00
44 S	2-Fluorobiphenyl	1.363	1.391	-2.1	112	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051315\
 Data File : BG017117.D
 Acq On : 13 May 2015 20:57
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 14 04:02:52 2015
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 13 16:12:13 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.441	1.465	-1.7	111	0.00
46	2-Chloronaphthalene	1.141	1.175	-3.0	113	0.00
47	2-Nitroaniline	0.444	0.474	-6.8	118	0.00
48	Acenaphthylene	1.953	2.028	-3.8	112	0.00
49	Dimethylphthalate	1.564	1.583	-1.2	111	0.00
50	2,6-Dinitrotoluene	0.347	0.360	-3.7	111	0.00
51 C	Acenaphthene	1.154	1.179	-2.2	113	0.00
52	3-Nitroaniline	0.384	0.412	-7.3	117	0.00
53 P	2,4-Dinitrophenol	0.201	0.204	-1.5	105	0.00
54	Dibenzofuran	1.715	1.728	-0.8	111	0.00
55 P	4-Nitrophenol	0.310	0.327	-5.5	117	0.00
56	2,4-Dinitrotoluene	0.483	0.514	-6.4	116	0.00
57	Fluorene	1.518	1.518	0.0	108	0.00
58	2,3,4,6-Tetrachlorophenol	0.380	0.392	-3.2	111	0.00
59	Diethylphthalate	1.680	1.667	0.8	111	0.00
60	4-Chlorophenyl-phenylether	0.780	0.773	0.9	107	0.00
61	4-Nitroaniline	0.429	0.464	-8.2	117	0.00
62	Azobenzene	1.605	1.670	-4.0	114	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	112	0.00
64	4,6-Dinitro-2-methylphenol	0.137	0.146	-6.6	112	0.00
65 c	n-Nitrosodiphenylamine	0.614	0.640	-4.2	113	0.00
66	4-Bromophenyl-phenylether	0.218	0.218	0.0	110	0.00
67	Hexachlorobenzene	0.230	0.230	0.0	109	0.00
68	Atrazine	0.225	0.231	-2.7	109	0.00
69 C	Pentachlorophenol	0.144	0.147	-2.1	109	0.00
70	Phenanthrene	1.031	1.065	-3.3	111	0.00
71	Anthracene	1.045	1.082	-3.5	113	0.00
72	Carbazole	0.960	1.025	-6.8	116	0.00
73	Di-n-butylphthalate	1.285	1.366	-6.3	116	0.00
74 C	Fluoranthene	1.239	1.255	-1.3	110	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	115	0.00
76	Benzidine	0.653	0.643	1.5	108	0.00
77	Pyrene	1.227	1.254	-2.2	114	0.00
78 S	Terphenyl-d14	0.960	0.976	-1.7	114	0.00
79	Butylbenzylphthalate	0.558	0.574	-2.9	115	0.00
80	Benzo(a)anthracene	1.146	1.144	0.2	111	0.00
81	3,3'-Dichlorobenzidine	0.444	0.450	-1.4	114	0.00
82	Chrysene	1.068	1.079	-1.0	113	0.00
83	Bis(2-ethylhexyl)phthalate	0.793	0.810	-2.1	112	0.00
84 c	Di-n-octyl phthalate	1.320	1.362	-3.2	116	0.00
85	Indeno(1,2,3-cd)pyrene	1.277	1.291	-1.1	112	-0.01
86 I	Perylene-d12	1.000	1.000	0.0	111	0.00
87	Benzo(b)fluoranthene	1.166	1.159	0.6	111	0.00

Data Path : Z:\HPCHEM1\BNA_G\Data\BG051315\
Data File : BG017117.D
Acq On : 13 May 2015 20:57
Operator : TP/IZ
Sample : SSTDCCC040
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 14 04:02:52 2015
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\8270-BG051315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 13 16:12:13 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.107	1.173	-6.0	115	0.00
89 C	Benzo(a)pyrene	1.071	1.099	-2.6	113	0.00
90	Dibenzo(a,h)anthracene	1.099	1.120	-1.9	113	-0.01
91	Benzo(g,h,i)perylene	1.035	1.054	-1.8	112	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0