




Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051424\
Data File : BG061182.D
Acq On : 14 May 2024 9:14
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 14 11:43:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061182.D\data.ms

(70) Pentachlorophenol (C)




16.965min (-0.021) 14808.23 ng

response 57357

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	59.80	63.89
264.00	64.40	62.37
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051424\
Data File : BG061182.D
Acq On : 14 May 2024 9:14
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: May 14 11:43:44 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration



TIC: BG061182.D\data.ms

(77) Benzidine

19.544min (-0.021) 0.00 ng

response 353643

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.00	13.87
183.00	13.60	12.61
0.00	0.00	0.00