

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051424\
 Data File : BG061189.D
 Acq On : 14 May 2024 14:11
 Operator : MA/JU
 Sample : PB160877BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 PB160877BL

Quant Time: May 14 14:50:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 05:11:25 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	18265	20.000	ng	-0.03
21) Naphthalene-d8	10.721	136	69036	20.000	ng	#-0.04
39) Acenaphthene-d10	14.557	164	58929	20.000	ng	-0.03
64) Phenanthrene-d10	17.307	188	165625	20.000	ng	-0.03
76) Chrysene-d12	21.567	240	177184	20.000	ng	-0.03
86) Perylene-d12	24.687	264	205146	20.000	ng	-0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.527	112	144088	160.436	ng	-0.02
7) Phenol-d6	7.107	99	198079	149.091	ng	-0.03
23) Nitrobenzene-d5	9.087	82	127610	91.632	ng	-0.03
42) 2,4,6-Tribromophenol	16.050	330	153985	130.730	ng	-0.03
45) 2-Fluorobiphenyl	13.183	172	347032	87.081	ng	-0.03
79) Terphenyl-d14	19.928	244	934813	87.746	ng	-0.03


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051424\
Data File : BG061189.D
Acq On : 14 May 2024 14:11
Operator : MA/JU
Sample : PB160877BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
PB160877BL

Quant Time: May 14 14:50:18 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 05:11:25 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.930 min Scan# 84
Delta R.T. -0.031 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11




Instrument :
BNA_G
ClientSampleId :
PB160877BL

Tgt Ion:152 Resp: 18265
Ion Ratio Lower Upper
152 100
150 149.3 138.3 207.5
115 58.4 48.0 72.0



#5
2-Fluorophenol
Concen: 160.436 ng
RT: 5.527 min Scan# 432
Delta R.T. -0.020 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11

Tgt Ion:112 Resp: 144088
Ion Ratio Lower Upper
112 100
64 74.4 55.7 83.5
63 45.6 38.1 57.1









#7
Phenol-d6
Concen: 149.091 ng
RT: 7.107 min Scan# 705
Delta R.T. -0.026 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11




Instrument :
BNA_G
ClientSampleId :
PB160877BL


Tgt Ion: 99 Resp: 198079
Ion Ratio Lower Upper
99 100
42 34.6 29.2 43.8
71 47.9 39.7 59.5



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.721 min Scan# 1316
Delta R.T. -0.037 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11

Tgt Ion: 136 Resp: 69036
Ion Ratio Lower Upper
136 100
137 9.4 7.8 11.6
54 11.5 8.1 12.1
68 4.9 3.0 4.4#





#23

Nitrobenzene-d5

Concen: 91.632 ng

RT: 9.087 min Scan# 1043

Delta R.T. -0.031 min

Lab File: BG061189.D

Acq: 14 May 2024 14:11

Instrument :

BNA_G

ClientSampleId :

PB160877BL





Tgt Ion: 82 Resp: 127610

Ion Ratio Lower Upper

82 100

128 36.9 32.7 49.1

54 79.7 65.6 98.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 14.557 min Scan# 1969

Delta R.T. -0.031 min

Lab File: BG061189.D

Acq: 14 May 2024 14:11




Tgt Ion:164 Resp: 58929


Ion Ratio Lower Upper

164 100

162 113.2 79.4 119.2

160 42.5 37.0 55.6








#42
2,4,6-Tribromophenol
Concen: 130.730 ng
RT: 16.050 min Scan# 21
Delta R.T. -0.026 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11




Instrument :
BNA_G
ClientSampleId :
PB160877BL


Tgt Ion:330 Resp: 153985
Ion Ratio Lower Upper
330 100
332 97.4 78.6 117.8
141 25.1 20.5 30.7



#45
2-Fluorobiphenyl
Concen: 87.081 ng
RT: 13.183 min Scan# 1735
Delta R.T. -0.026 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11

Tgt Ion:172 Resp: 347032
Ion Ratio Lower Upper
172 100
171 37.2 29.0 43.6
170 24.0 20.2 30.4





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.307 min Scan# 2441

Delta R.T. -0.026 min

Lab File: BG061189.D

Acq: 14 May 2024 14:11

Instrument :

BNA_G

ClientSampleId :

PB160877BL





Tgt Ion:188 Resp: 165625

Ion Ratio Lower Upper

188 100

94 4.6 3.6 5.4

80 5.7 4.2 6.2



#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.567 min Scan# 3162

Delta R.T. -0.026 min

Lab File: BG061189.D

Acq: 14 May 2024 14:11




Tgt Ion:240 Resp: 177184


Ion Ratio Lower Upper

240 100

120 4.3 3.2 4.8


236 25.7 21.0 31.6








#79
Terphenyl-d14
Concen: 87.746 ng
RT: 19.928 min Scan# 21
Delta R.T. -0.026 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11

Instrument :
BNA_G
ClientSampleId :
PB160877BL



Tgt Ion:244 Resp: 934813
Ion Ratio Lower Upper
244 100
212 7.6 5.7 8.5
122 5.3 4.2 6.2



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.687 min Scan# 3693
Delta R.T. -0.043 min
Lab File: BG061189.D
Acq: 14 May 2024 14:11

Tgt Ion:264 Resp: 205146
Ion Ratio Lower Upper
264 100
260 25.1 17.7 26.5
265 21.0 16.4 24.6

