

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051516\
 Data File : BG021890.D
 Acq On : 15 May 2016 12:54
 Operator : UM/SJ
 Sample : H3096-02MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Manual Integrations
 APPROVED

sohil
 5/16/2016 7:03:46 PM

Quant Time: May 16 13:55:56 2016
 Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG050916.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon May 16 01:27:47 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	47346	20.00	ng/ul	0.00
18) Naphthalene-d8	10.99	136	251165	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.79	164	155214	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.53	188	326501	20.00	ng/ul	0.00
76) Chrysene-d12	21.81	240	322096	20.00	ng/ul	0.00
84) Perylene-d12	25.15	264	321893	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.54	96	3346	3.41	ng/uL	0.00
5) Phenol-d5	7.31	99	76588	21.80	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.48	67	37570	21.57	ng/ul	0.00
9) 2-Chlorophenol-d4	7.69	132	72165	26.46	ng/ul	0.00
13) 4-Methylphenol-d8	8.87	113	62186	21.07	ng/ul	0.00
19) Nitrobenzene-d5	9.33	128	38383	23.24	ng/ul	0.00
22) 2-Nitrophenol-d4	10.06	143	40303	38.64	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.60	165	79651	27.01	ng/ul	0.00
29) 4-Chloroaniline-d4	11.12	131	75087m	23.63	ng/ul	0.00
44) Dimethylphthalate-d6	14.19	166	277977	25.50	ng/ul	0.00
47) Acenaphthylene-d8	14.48	160	368838	23.83	ng/ul	0.00
52) 4-Nitrophenol-d4	14.98	143	30857m	16.51	ng/ul	0.00
58) Fluorene-d10	15.77	176	248470	21.65	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.89	200	29173	32.26	ng/ul	0.00
71) Anthracene-d10	17.63	188	353579	22.25	ng/ul	0.00
77) Pyrene-d10	19.91	212	378254	25.09	ng/ul	0.00
88) Benzo(a)pyrene-d12	24.92	264	345375	22.58	ng/ul	0.00

Target Compounds

						Qvalue
6) Phenol	7.34	94	78517	21.01	ng/ul#	71
10) 2-Chlorophenol	7.73	128	71614	26.05	ng/ul#	76
15) N-Nitroso-di-n-propylamine	8.97	70	51446	23.42	ng/ul#	73
28) Naphthalene	11.03	128	40266	3.20	ng/ul	98
33) 4-Chloro-3-methylphenol	12.24	107	79730	26.25	ng/ul#	67
34) 2-Methylnaphthalene	12.63	142	15128	1.63	ng/ul	96
35) 1-Methylnaphthalene	12.84	142	13142	1.43	ng/ul#	87
45) Dimethylphthalate	14.23	163	69008	6.21	ng/ul	97
50) Acenaphthene	14.85	153	280796	25.48	ng/ul	99
53) 4-Nitrophenol	14.99	109	12437	14.60	ng/ul#	45
54) Dibenzofuran	15.19	168	32254	2.18	ng/ul	89
55) 2,4-Dinitrotoluene	15.14	165	75756	24.97	ng/ul#	55
59) Fluorene	15.83	166	26961	2.11	ng/ul#	97
69) Pentachlorophenol	17.18	266	53563	35.59	ng/ul	95
70) Phenanthrene	17.57	178	461015	25.83	ng/ul	96
72) Anthracene	17.66	178	113336	6.25	ng/ul	97
73) Carbazole	17.94	167	93596m	5.95	ng/ul	
75) Fluoranthene	19.58	202	538021	26.53	ng/ul	96
78) Pyrene	19.93	202	887917m	47.65	ng/ul	
79) Butylbenzylphthalate	20.80	149	20252	4.92	ng/ul#	79
81) Benzo(a)anthracene	21.79	228	206674	11.53	ng/ul	97
82) Bis(2-ethylhexyl)phthalate	21.69	149	356478	53.01	ng/ul	97
83) Chrysene	21.86	228	190417	10.74	ng/ul	97

Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051516\
Data File : BG021890.D
Acq On : 15 May 2016 12:54
Operator : UM/SJ
Sample : H3096-02MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Manual Integrations
APPROVED

sohil
5/16/2016 7:03:46 PM

Quant Time: May 16 13:55:56 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG050916.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon May 16 01:27:47 2016
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
86) Benzo(b)fluoranthene	24.09	252	262733	14.26	ng/ul#	97
87) Benzo(k)fluoranthene	24.15	252	83347m	4.23	ng/ul	
89) Benzo(a)pyrene	24.98	252	171640	9.36	ng/ul#	95
90) Indeno(1,2,3-cd)pyrene	28.95	276	119497	5.50	ng/ul#	90
91) Dibenzo(a,h)anthracene	29.01	278	31375m	1.73	ng/ul	
92) Benzo(g,h,i)perylene	30.13	276	104962	5.85	ng/ul#	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_G\DATA\BG051516\
Data File : BG021890.D
Acq On : 15 May 2016 12:54
Operator : UM/SJ
Sample : H3096-02MS
Misc :
ALS Vial : 5 Sample Multiplier: 1


Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Manual Integrations
APPROVED

sohil
5/16/2016 7:03:46 PM

Quant Time: May 16 13:55:56 2016
Quant Method : Z:\HPCHEM1\BNA_G\METHODS\SOM02.2-EPA-BG050916.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon May 16 01:27:47 2016
Response via : Initial Calibration





#6

Phenol

Concen: 21.01 ng/ul

RT: 7.34 min Scan# 766

Delta R.T. 0.00 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54

Instrument :

BNA_G

ClientSampleId :

D9XF7MS

Tgt Ion: 94 Resp: 78517
Ion Ratio Lower Upper

94 100


65 19.7 26.1 39.1#

66 22.7 35.0 52.4#




Manual Integrations
APPROVED

sohil

5/16/2016 7:03:46 PM



Abundance Ion 94.00 (93.70 to 94.70): BGC
Ion 65.00 (64.70 to 65.70): BGC
Ion 66.00 (65.70 to 66.70): BGC



#10

2-Chlorophenol

Concen: 26.05 ng/ul

RT: 7.73 min Scan# 832

Delta R.T. 0.00 min

Lab File: BG021890.D


Acq: 15 May 2016 12:54

Tgt Ion: 128 Resp: 71614
Ion Ratio Lower Upper



128 100


64 33.0 47.0 70.6#

130 31.3 27.8 41.6



Abundance Ion 128.00 (127.70 to 128.70): E
Ion 64.00 (63.70 to 64.70): BGC
Ion 130.00 (129.70 to 130.70): E









#15
N-Nitroso-di-n-propylamine
Concen: 23.42 ng/ul
RT: 8.97 min Scan# 1043
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Tgt Ion	Ratio	Lower	Upper
70	100		
42	32.5	43.4	65.0#
101	11.8	6.5	9.7#
130	28.6	13.8	20.6#




Manual Integrations
APPROVED


sohil
5/16/2016 7:03:46 PM



#28
Naphthalene
Concen: 3.20 ng/ul
RT: 11.03 min Scan# 1395
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.3	12.5
127	14.3	10.8	16.2






#33
 4-Chloro-3-methylphenol
 Concen: 26.25 ng/ul
 RT: 12.24 min Scan# 1601
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS




Tgt Ion	Ratio	Lower	Upper
107	100		
144	31.3	17.2	25.8#
142	94.0	52.0	78.0#

Manual Integrations
 APPROVED

sohil
 5/16/2016 7:03:46 PM




Abundance Ion 107.00 (106.70 to 107.70): E
 Ion 144.00 (143.70 to 144.70): E
 Ion 142.00 (141.70 to 142.70): E






#34
 2-Methylnaphthalene
 Concen: 1.63 ng/ul
 RT: 12.63 min Scan# 1667
 Delta R.T. 0.00 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Lower	Upper
142	100		
141	93.0	71.3	106.9



Abundance Ion 142.00 (141.70 to 142.70): E
 Ion 141.00 (140.70 to 141.70): E









#35
 1-Methylnaphthalene
 Concen: 1.43 ng/ul
 RT: 12.84 min Scan# 1703
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Tgt Ion	Ratio	Lower	Upper
142	100		
141	83.1	77.0	115.4
116	2.3	4.2	6.4#




Manual Integrations
 APPROVED


sohil
 5/16/2016 7:03:46 PM



#45
 Dimethylphthalate
 Concen: 6.21 ng/ul
 RT: 14.23 min Scan# 1939
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.4	3.9	5.9
164	9.3	8.7	13.1









#50
Acenaphthene
Concen: 25.48 ng/ul
RT: 14.85 min Scan# 2045
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Tgt Ion	Ratio	Resp	Lower	Upper
153	100	280796		
152	47.1	36.5	54.7	
154	84.5	67.3	100.9	




Manual Integrations
APPROVED


sohil
5/16/2016 7:03:46 PM



#53
4-Nitrophenol
Concen: 14.60 ng/ul
RT: 14.99 min Scan# 2069
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Resp	Lower	Upper
109	100	12437		
139	211.4	75.9	113.9#	
65	131.6	99.8	149.8	









#54
Dibenzenofuran
Concen: 2.18 ng/ul
RT: 15.19 min Scan# 2102
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Tgt Ion	Ratio	Resp	Lower	Upper
168	100	32254		
139	36.9	35.2	52.8	




Manual Integrations
APPROVED


sohil
5/16/2016 7:03:46 PM



#55
2,4-Dinitrotoluene
Concen: 24.97 ng/ul
RT: 15.14 min Scan# 2094
Delta R.T. -0.01 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Resp	Lower	Upper
165	100	75756		
63	24.3	48.1	72.1#	
182	2.6	2.2	3.2	





#59

Fluorene

Concen: 2.11 ng/ul

RT: 15.83 min Scan# 2212

Delta R.T. 0.00 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54

Instrument :

BNA_G





ClientSampleId :

D9XF7MS

Tgt Ion:	166	Resp:	26961
Ion Ratio	Lower	Upper	
166	100		
165	99.5	81.1	121.7
167	16.6	10.6	16.0#

Manual Integrations
APPROVED

sohil
5/16/2016 7:03:46 PM



#69

Pentachlorophenol

Concen: 35.59 ng/ul




RT: 17.18 min Scan# 2441


Delta R.T. 0.00 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54

Tgt Ion:	266	Resp:	53563
Ion Ratio	Lower	Upper	
266	100		
264	71.5	56.7	85.1
268	71.0	50.6	76.0









#70
Phenanthrene
Concen: 25.83 ng/ul
RT: 17.57 min Scan# 2508
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	461015		
179	16.8	11.8	17.8	
176	19.6	14.8	22.2	




Manual Integrations
APPROVED


sohil
5/16/2016 7:03:46 PM



#72
Anthracene
Concen: 6.25 ng/ul
RT: 17.66 min Scan# 2523
Delta R.T. -0.01 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	113336		
179	16.3	12.9	19.3	
176	17.0	15.6	23.4	









#73
 Carbazole
 Concen: 5.95 ng/ul m
 RT: 17.94 min Scan# 2570
 Delta R.T. 0.00 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Tgt Ion	Ratio	Resp	Lower	Upper
167	100	93596		
166	23.0	19.0	28.4	
139	12.1	12.2	18.4	




Manual Integrations
 APPROVED


sohil
 5/16/2016 7:03:46 PM



#75
 Fluoranthene
 Concen: 26.53 ng/ul
 RT: 19.58 min Scan# 2849
 Delta R.T. 0.00 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	538021		
101	12.6	8.9	13.3	
100	10.0	7.0	10.4	









#78
 Pyrene
 Concen: 47.65 ng/ul m
 RT: 19.93 min Scan# 2910
 Delta R.T. 0.00 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
101	15.6	9.8	14.8#
100	14.1	9.3	13.9#




Manual Integrations
 APPROVED


sohil
 5/16/2016 7:03:46 PM



#79
 Butylbenzylphthalate
 Concen: 4.92 ng/ul
 RT: 20.80 min Scan# 3058
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
91	59.8	65.8	98.8#
206	21.8	15.5	23.3









#81
Benzo(a)anthracene
Concen: 11.53 ng/ul
RT: 21.79 min Scan# 3226
Delta R.T. -0.01 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Instrument :
BNA_G
ClientSampleId :
D9XF7MS

Tgt Ion	Ratio	Lower	Upper
228	100		
229	21.8	16.6	25.0
226	28.4	21.0	31.4




Manual Integrations
APPROVED


sohil
5/16/2016 7:03:46 PM



#82
Bis(2-ethylhexyl)phthalate
Concen: 53.01 ng/ul
RT: 21.69 min Scan# 3208
Delta R.T. 0.00 min
Lab File: BG021890.D
Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Lower	Upper
149	100		
167	27.6	20.5	30.7
279	4.6	3.7	5.5





#83

Chrysene

Concen: 10.74 ng/ul

RT: 21.86 min Scan# 3238

Delta R.T. -0.01 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54

Instrument :

BNA_G


ClientSampleId :

D9XF7MS

Tgt Ion:	228	Resp:	190417
Ion Ratio	Lower	Upper	
228	100		
226	30.2	23.4	35.2
229	22.4	16.2	24.2

Manual Integrations
APPROVED




sohil
5/16/2016 7:03:46 PM



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#86

Benzo(b)fluoranthene

Concen: 14.26 ng/ul


RT: 24.09 min Scan# 3617

Delta R.T. 0.00 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54



Tgt Ion:	252	Resp:	262733
Ion Ratio	Lower	Upper	
252	100		
253	23.0	18.1	27.1
125	12.7	7.6	11.4#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#87

Benzo(k)fluoranthene

Concen: 4.23 ng/ul m

RT: 24.15 min Scan# 3627

Delta R.T. -0.01 min

Lab File: BG021890.D

Acq: 15 May 2016 12:54

Instrument :

BNA_G


ClientSampleId :

D9XF7MS




Tgt Ion:	252	Resp:	83347
Ion Ratio	Lower	Upper	
252	100		
253	22.9	17.8	26.8
125	15.4	7.7	11.5#

Manual Integrations
APPROVED

sohil
5/16/2016 7:03:46 PM



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#89

Benzo(a)pyrene

Concen: 9.36 ng/ul

RT: 24.98 min Scan# 3769




Delta R.T. -0.01 min


Lab File: BG021890.D

Acq: 15 May 2016 12:54

Tgt Ion:	252	Resp:	171640
Ion Ratio	Lower	Upper	
252	100		
253	22.4	17.4	26.0
125	14.7	8.2	12.2#

Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#90
 Indeno(1,2,3-cd)pyrene
 Concen: 5.50 ng/ul
 RT: 28.95 min Scan# 4445
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.0	15.8	23.6#
277	21.6	19.0	28.6




Manual Integrations
 APPROVED


sohil
 5/16/2016 7:03:46 PM



#91
 Dibenzo(a,h)anthracene
 Concen: 1.73 ng/ul m
 RT: 29.01 min Scan# 4454
 Delta R.T. -0.02 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.7	12.1	18.1#
279	33.6	19.8	29.8#





#92
 Benzo(g,h,i)perylene
 Concen: 5.85 ng/ul
 RT: 30.13 min Scan# 4646
 Delta R.T. -0.01 min
 Lab File: BG021890.D
 Acq: 15 May 2016 12:54

Instrument :
 BNA_G
 ClientSampleId :
 D9XF7MS

Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.1	13.9	20.9#
277	24.6	19.8	29.6

Manual Integrations
 APPROVED

sohil
 5/16/2016 7:03:46 PM

