


Data Path : Z:\HPCHEM1\BNA G\DATA\BG051517\
Data File : BG026956.D
Acq On : 15 May 2017 16:29
Operator : SJ/MA
Sample : DFTPP62
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_G
ClientSampleId :
DFTPP62

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_G\METHODS\SOM-EPA-BG042717.M
Title : SVOA CALIBRATION
Last Update : Wed May 10 01:49:01 2017



AutoFind: Scans 2500, 2501, 2502; Background Corrected with Scan 2493

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.0	69438	PASS
68	69	0.00	2	0.8	566	PASS
69	198	0.00	100	36.9	73329	PASS
70	69	0.00	2	0.6	440	PASS
127	198	10	80	56.3	111816	PASS
197	198	0.00	2	0.4	711	PASS
198	198	100	100	100.0	198592	PASS
199	198	5	9	6.5	12957	PASS
275	198	10	60	24.5	48562	PASS
365	198	1	100	3.2	6440	PASS
441	442	0.01	100	16.4	22560	PASS
442	198	50	100	69.3	137584	PASS
443	442	15	24	18.6	25645	PASS

m/z	Abundance
39.90	284.0
43.10	221.0
44.00	770.0
44.90	173.0
49.00	223.0
50.90	178.0
56.10	151.0
68.90	204.0
83.10	206.0
83.90	474.0
85.80	284.0
93.00	166.0
207.00	177.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

m/z	Abundance
37.00	330.0
38.00	521.0
39.00	4863.0
40.00	334.0
43.00	184.0
44.00	1024.0
45.10	243.0
49.00	875.0
50.00	18552.0
51.00	78528.0
52.00	4608.0
55.00	733.0
56.00	2949.0
57.00	7167.0
58.00	347.0
60.00	192.0
61.00	1021.0
62.00	1644.0
63.00	3080.0
64.00	566.0
65.00	1988.0
65.90	277.0
67.00	275.0
69.00	83128.0
70.00	360.0
73.00	790.0
74.00	7731.0
75.00	13887.0
76.00	5063.0
77.00	105496.0
78.00	7690.0
79.00	5951.0
80.00	4737.0
81.00	7117.0
82.00	1926.0
83.00	1653.0
84.00	367.0
85.00	1059.0
85.90	1731.0
87.00	1467.0
87.90	556.0
89.00	209.0
91.00	1701.0
92.00	1504.0
93.00	9291.0
94.00	824.0
95.00	396.0
96.00	528.0
97.10	378.0
98.00	7679.0
99.00	6876.0
99.90	640.0
101.00	3984.0
101.90	190.0
102.90	1263.0
104.00	2326.0
105.00	2860.0
106.00	773.0
107.00	31848.0
108.00	5234.0
110.00	61432.0
111.00	9240.0
111.90	1356.0
113.00	243.0
114.90	162.0
116.10	1492.0
117.00	20944.0
117.90	1791.0
118.90	208.0
120.00	618.0
120.80	177.0
122.00	2064.0
123.00	2662.0
123.90	1617.0
125.00	1201.0
127.00	120744.0
128.00	10928.0
129.00	40896.0
130.00	3318.0
131.00	500.0
132.00	401.0
132.80	152.0
134.00	989.0
135.00	3035.0
135.90	1521.0
137.00	1816.0
137.90	426.0
139.00	226.0
139.90	460.0
140.90	5502.0
142.00	1957.0
143.00	1387.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

144.00	515.0
145.00	183.0
145.90	976.0
147.00	3514.0
148.00	5654.0
149.00	1345.0
149.90	320.0
151.00	802.0
151.70	502.0
153.00	2032.0
154.00	1237.0
155.00	3106.0
156.00	4317.0
157.00	864.0
158.00	938.0
158.90	606.0
160.00	1702.0
161.00	2342.0
161.90	633.0
163.00	247.0
164.00	438.0
165.00	2037.0
165.90	1771.0
167.00	10367.0
168.00	5734.0
169.00	903.0
170.10	476.0
170.80	537.0
171.90	972.0
173.00	1091.0
173.90	2033.0
175.00	3944.0
175.90	765.0
176.90	2069.0
178.00	738.0
179.00	7334.0
180.00	5108.0
181.00	2023.0
182.10	222.0
182.90	323.0
184.00	655.0
185.00	4804.0
186.00	29088.0
187.00	7744.0
188.00	1229.0
188.90	2512.0
189.80	430.0
191.00	647.0
191.90	2777.0
193.00	2097.0
194.20	322.0
196.00	6068.0
197.90	201280.0
198.90	14417.0
200.00	1033.0
201.50	1034.0
203.00	1562.0
204.00	6391.0
205.00	11020.0
206.00	51928.0
207.00	7638.0
208.00	1547.0
209.00	423.0
210.30	693.0
211.00	1728.0
211.90	519.0
213.00	176.0
214.90	390.0
216.00	871.0
217.00	12201.0
218.00	1360.0
220.00	658.0
221.00	7060.0
223.00	3383.0
224.00	28784.0
225.00	5898.0
226.20	729.0
227.00	11612.0
228.00	2087.0
228.90	2498.0
230.00	344.0
231.00	945.0
232.00	232.0
232.90	192.0
234.00	777.0
235.00	619.0
235.90	277.0
237.10	1263.0
238.90	333.0
240.00	407.0
241.00	835.0
242.00	1757.0
243.00	1530.0
244.00	21728.0
245.00	2455.0
246.00	4085.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

246.90	799.0
247.90	184.0
248.90	561.0
250.10	221.0
250.90	261.0
251.90	452.0
252.90	828.0
255.00	106680.0
256.00	15847.0
257.00	1282.0
258.00	5967.0
258.90	1010.0
260.00	202.0
260.80	160.0
263.90	269.0
265.00	2396.0
266.00	322.0
269.90	358.0
270.90	229.0
271.90	325.0
272.90	3081.0
274.00	8447.0
275.00	45160.0
276.00	6611.0
277.00	3862.0
278.00	891.0
278.70	152.0
283.00	363.0
284.00	554.0
285.10	799.0
288.80	204.0
289.70	204.0
291.90	189.0
292.10	197.0
293.00	1132.0
294.10	245.0
295.00	422.0
296.00	11360.0
297.00	1553.0
302.20	249.0
303.00	1088.0
303.80	395.0
304.20	390.0
308.10	224.0
310.60	186.0
313.40	297.0
313.90	665.0
315.00	1380.0
315.90	651.0
317.20	165.0
321.00	626.0
322.20	318.0
323.00	4541.0
324.00	945.0
327.00	939.0
328.10	450.0
332.10	296.0
333.10	397.0
334.00	2485.0
335.00	868.0
339.10	152.0
340.10	190.0
341.00	416.0
345.90	825.0
350.20	226.0
352.00	1520.0
353.10	867.0
354.00	1300.0
359.00	213.0
365.00	5733.0
365.90	538.0
370.00	330.0
371.00	408.0
372.00	2025.0
373.10	299.0
382.90	502.0
383.80	268.0
390.00	275.0
390.90	185.0
401.10	156.0
402.00	730.0
403.00	993.0
403.90	353.0
404.90	302.0
420.90	782.0
422.10	1015.0
423.00	6090.0
424.00	1220.0
424.90	245.0
435.40	188.0
437.70	226.0
439.00	274.0
439.40	298.0
441.00	18088.0
442.00	109584.0
443.00	23368.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

444.102184.0

444.80259.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

m/z	Abundance
36.90	216.0
38.00	570.0
39.00	5065.0
39.90	564.0
41.00	547.0
43.90	672.0
45.00	162.0
49.10	950.0
50.00	18304.0
51.00	82304.0
52.00	4193.0
53.00	155.0
55.00	585.0
55.90	2576.0
57.00	5922.0
58.00	280.0
60.90	1218.0
62.00	1025.0
63.00	4069.0
64.00	485.0
65.00	1975.0
66.00	262.0
67.10	249.0
69.00	85928.0
69.90	600.0
71.20	178.0
72.20	162.0
73.00	751.0
74.00	8096.0
75.00	15955.0
76.00	5147.0
77.00	105688.0
78.00	7739.0
79.00	6584.0
80.00	4262.0
81.00	6419.0
82.00	2105.0
82.90	2373.0
83.90	689.0
85.00	1460.0
86.00	1784.0
86.90	1404.0
88.00	391.0
91.00	1601.0
92.00	1747.0
93.00	9650.0
93.90	753.0
95.00	314.0
96.00	473.0
97.00	220.0
98.00	7631.0
99.00	6525.0
100.00	674.0
101.00	4761.0
101.90	432.0
103.00	1464.0
103.90	2445.0
105.00	2543.0
106.00	920.0
107.00	34424.0
108.00	5506.0
110.00	64080.0
110.90	9799.0
112.00	1084.0
112.90	459.0
115.00	363.0
116.00	2039.0
117.00	23336.0
117.90	1979.0
119.00	308.0
120.00	418.0
121.00	233.0
121.90	2428.0
123.00	3183.0
124.00	1840.0
125.00	1577.0
127.00	128664.0
128.00	9960.0
129.00	43400.0
129.90	3866.0
131.00	717.0
132.00	541.0
132.60	257.0
133.10	176.0
133.90	1282.0
135.00	3911.0
136.00	1444.0
137.00	2072.0
137.90	558.0
139.90	715.0
141.00	6183.0
142.00	1912.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

142.90	1502.0
144.00	556.0
145.00	437.0
145.90	1261.0
147.00	3788.0
148.00	5587.0
148.90	1580.0
150.00	326.0
151.00	610.0
151.60	401.0
152.90	1947.0
154.00	1608.0
155.00	3727.0
156.00	5342.0
156.90	895.0
157.90	1106.0
159.00	1136.0
160.00	1564.0
161.00	2934.0
162.00	885.0
163.00	250.0
164.10	294.0
164.90	2241.0
166.00	1448.0
167.00	11500.0
168.00	5810.0
168.90	1213.0
170.10	435.0
170.90	457.0
171.90	1045.0
172.90	1285.0
173.90	2148.0
175.00	5695.0
175.80	1182.0
176.90	1992.0
178.00	730.0
178.90	9192.0
180.00	5591.0
181.00	2953.0
182.00	464.0
183.20	323.0
184.00	1064.0
185.00	4452.0
186.00	32256.0
187.00	8700.0
187.90	1037.0
189.00	1929.0
190.10	314.0
190.90	921.0
192.00	2699.0
193.00	3151.0
194.00	701.0
195.00	590.0
196.00	6325.0
196.70	2133.0
197.90	228992.0
199.00	14811.0
200.00	967.0
201.50	952.0
203.00	1571.0
204.00	8173.0
205.00	13337.0
206.00	59864.0
207.00	7707.0
208.00	1751.0
209.10	722.0
210.40	1214.0
211.00	2431.0
212.10	276.0
213.10	197.0
214.90	926.0
215.90	1389.0
217.00	14099.0
218.00	1514.0
218.80	175.0
219.40	208.0
221.00	8715.0
222.00	2256.0
223.00	3509.0
224.00	30952.0
225.00	8319.0
226.10	979.0
227.00	13270.0
228.00	1980.0
229.00	2287.0
230.00	323.0
231.00	1366.0
232.00	153.0
232.90	227.0
234.00	1047.0
234.90	628.0
235.90	577.0
236.90	962.0
238.00	255.0
238.90	327.0
240.00	507.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

240.90	905.0
242.00	2246.0
243.00	1774.0
244.00	25160.0
245.00	3099.0
246.00	5585.0
247.00	760.0
248.00	364.0
248.90	1332.0
250.00	208.0
250.70	405.0
252.20	437.0
253.00	869.0
255.00	126344.0
256.00	17864.0
257.00	2129.0
258.00	7077.0
259.00	1321.0
260.10	216.0
261.00	168.0
262.10	175.0
265.00	3071.0
266.00	438.0
269.80	376.0
271.10	259.0
272.10	633.0
273.00	4371.0
274.00	10193.0
275.00	57608.0
276.00	8648.0
276.90	5055.0
278.00	671.0
278.90	261.0
281.00	241.0
283.00	407.0
284.00	434.0
285.10	1152.0
286.10	214.0
287.90	247.0
288.90	462.0
290.00	378.0
291.10	261.0
292.10	280.0
293.00	978.0
294.00	597.0
295.00	383.0
296.00	16134.0
297.00	2374.0
297.80	176.0
301.00	250.0
302.10	322.0
303.00	1938.0
304.00	525.0
307.90	224.0
309.80	252.0
313.10	326.0
314.00	919.0
314.90	1626.0
316.00	1181.0
316.90	163.0
321.10	551.0
322.00	444.0
323.00	6574.0
324.10	1004.0
327.00	719.0
328.00	490.0
328.90	160.0
331.90	454.0
333.00	628.0
334.00	3960.0
335.00	765.0
341.00	517.0
342.00	249.0
346.00	1174.0
347.00	299.0
350.90	192.0
352.00	2022.0
352.90	1299.0
354.00	1483.0
355.00	445.0
365.00	7675.0
365.90	1222.0
369.90	296.0
370.90	325.0
372.00	3206.0
373.00	617.0
382.90	633.0
384.00	361.0
389.90	409.0
391.00	253.0
401.00	239.0
402.00	1441.0
403.00	1606.0
403.90	518.0
414.80	187.0
421.00	1376.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

422.00	1282.0
423.00	8839.0
424.10	2160.0
430.40	194.0
435.00	220.0
437.30	164.0
439.10	190.0
439.80	552.0
441.00	25400.0
442.00	164160.0
443.00	29288.0
444.00	3277.0
444.90	196.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

m/z	Abundance
37.00	232.0
38.10	332.0
39.10	2872.0
39.90	300.0
40.80	219.0
43.90	733.0
49.00	741.0
50.00	11528.0
51.00	48016.0
52.00	2333.0
52.90	176.0
54.90	619.0
56.00	1959.0
57.00	3875.0
58.00	269.0
60.90	400.0
62.00	935.0
63.00	2299.0
64.00	330.0
65.00	1126.0
67.20	156.0
68.10	1698.0
68.90	51544.0
70.10	362.0
73.00	764.0
74.00	4878.0
75.00	6992.0
76.00	2833.0
77.00	65112.0
78.00	4604.0
78.90	3661.0
80.00	2861.0
81.00	3999.0
81.90	751.0
83.00	1078.0
84.00	202.0
85.00	678.0
86.00	1106.0
86.90	512.0
87.80	296.0
89.10	221.0
91.00	1509.0
92.00	1237.0
93.00	6277.0
94.00	649.0
95.10	219.0
96.00	435.0
97.90	4911.0
99.00	4045.0
100.00	307.0
101.00	2565.0
101.80	299.0
103.00	766.0
103.90	1946.0
105.00	1330.0
106.00	886.0
107.00	18736.0
108.00	3362.0
110.00	38480.0
111.00	4535.0
112.00	491.0
113.00	292.0
115.00	276.0
116.00	1512.0
117.00	11903.0
117.90	1192.0
119.00	196.0
119.90	436.0
121.10	151.0
121.90	1832.0
123.00	1749.0
123.90	841.0
125.00	1116.0
125.70	583.0
127.00	86040.0
128.00	6213.0
129.00	29368.0
130.00	2135.0
130.90	536.0
132.00	225.0
132.90	216.0
133.90	766.0
135.00	3546.0
136.00	1090.0
137.00	1374.0
138.00	280.0
140.00	439.0
140.90	3729.0
142.00	1868.0
143.00	810.0
144.90	364.0
146.00	801.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

147.00	1874.0
148.00	3700.0
149.00	842.0
149.80	183.0
151.00	696.0
151.70	256.0
152.90	1136.0
154.00	1179.0
155.00	1681.0
156.00	3456.0
157.10	699.0
157.90	748.0
159.00	484.0
160.00	1385.0
160.90	1555.0
162.00	412.0
164.00	277.0
164.90	1829.0
166.00	1195.0
167.00	6492.0
167.90	3389.0
169.00	809.0
169.90	317.0
170.90	635.0
172.00	1086.0
173.00	1044.0
174.00	1887.0
175.00	3193.0
176.00	660.0
177.00	1221.0
178.00	572.0
178.90	6441.0
180.00	3960.0
181.00	1906.0
182.20	302.0
182.80	154.0
184.00	415.0
185.00	2850.0
186.00	21488.0
187.00	6470.0
188.00	588.0
188.90	1174.0
189.90	225.0
191.00	772.0
191.90	1497.0
193.00	1962.0
194.10	482.0
195.00	420.0
196.00	5217.0
197.90	165504.0
198.90	9643.0
200.00	1015.0
201.40	942.0
203.00	1333.0
204.00	6494.0
205.00	11224.0
206.00	40648.0
207.00	5410.0
208.00	1491.0
208.90	316.0
209.90	662.0
210.50	1063.0
211.00	1717.0
211.80	196.0
214.90	517.0
215.90	683.0
216.90	9749.0
217.90	1377.0
219.10	500.0
221.00	5566.0
223.00	2191.0
224.00	20480.0
225.00	4546.0
226.10	534.0
227.00	9647.0
227.90	1401.0
229.00	1905.0
230.00	259.0
230.90	658.0
232.10	350.0
234.00	397.0
235.00	830.0
235.90	271.0
237.00	895.0
238.80	313.0
240.00	345.0
241.00	506.0
242.00	1166.0
243.00	1390.0
244.00	18240.0
245.00	2239.0
245.90	4007.0
246.90	508.0
247.80	180.0
249.00	836.0
250.60	204.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62

251.10	209.0
253.00	691.0
255.00	93568.0
256.00	13428.0
257.00	897.0
258.00	4800.0
259.00	982.0
265.00	2027.0
265.90	377.0
267.00	248.0
270.90	308.0
271.90	544.0
273.00	3151.0
274.00	8002.0
275.00	42920.0
276.00	6019.0
277.00	2925.0
278.00	649.0
278.90	210.0
282.00	171.0
283.00	565.0
284.10	414.0
284.90	512.0
285.80	319.0
289.00	229.0
291.80	193.0
293.00	1096.0
293.80	183.0
294.60	185.0
296.00	11576.0
297.00	1627.0
297.80	294.0
300.90	165.0
302.00	245.0
303.00	1451.0
304.00	333.0
306.80	181.0
308.00	228.0
309.20	153.0
310.00	182.0
313.90	733.0
315.00	1295.0
316.00	845.0
321.00	861.0
322.10	284.0
323.00	4308.0
324.00	1050.0
327.00	977.0
328.00	483.0
332.00	451.0
333.10	526.0
334.00	2282.0
335.00	711.0
335.80	165.0
339.20	168.0
341.00	591.0
341.80	185.0
346.00	873.0
350.80	181.0
352.00	1954.0
353.10	1005.0
354.00	758.0
355.00	258.0
365.00	5912.0
365.90	898.0
369.80	150.0
370.90	280.0
372.00	2345.0
373.10	459.0
382.90	672.0
383.90	157.0
390.00	225.0
391.00	297.0
401.10	206.0
401.90	859.0
403.00	1475.0
404.00	755.0
421.00	917.0
422.00	1195.0
423.00	7389.0
424.00	1640.0
424.90	206.0
430.20	310.0
436.20	303.0
437.60	201.0
438.30	169.0
439.20	321.0
441.00	24192.0
442.00	139008.0
443.00	24280.0
444.00	2698.0

Instrument :
BNA_G
ClientSampleId :
DFTPP62