

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051523\
 Data File : BG057436.D
 Acq On : 15 May 2023 17:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 16 02:15:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 22:04:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.483	0.463	4.1	118	0.00
3	Pyridine	1.554	1.487	4.3	113	0.00
4	n-Nitrosodimethylamine	0.730	0.628	14.0	101	0.00
5 S	2-Fluorophenol	1.169	1.188	-1.6	122	0.00
6	Aniline	2.252	2.195	2.5	116	0.00
7 S	Phenol-d6	1.775	1.820	-2.5	123	0.00
8	2-Chlorophenol	1.235	1.245	-0.8	122	0.00
9	Benzaldehyde	1.042	0.878	15.7	115	0.00
10 C	Phenol	1.731	1.727	0.2	118	0.00
11	bis(2-Chloroethyl)ether	1.305	1.264	3.1	118	0.00
12	1,3-Dichlorobenzene	1.364	1.377	-1.0	123	0.00
13 C	1,4-Dichlorobenzene	1.370	1.408	-2.8	124	0.00
14	1,2-Dichlorobenzene	1.322	1.322	0.0	123	0.00
15	Benzyl Alcohol	1.479	1.428	3.4	111	0.00
16	2,2'-oxybis(1-Chloropropane	1.645	1.508	8.3	111	0.00
17	2-Methylphenol	1.248	1.248	0.0	119	0.00
18	Hexachloroethane	0.494	0.488	1.2	121	0.00
19 P	n-Nitroso-di-n-propylamine	1.354	1.256	7.2	112	0.00
20	3+4-Methylphenols	1.708	1.757	-2.9	123	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
22	Acetophenone	0.572	0.551	3.7	117	0.00
23 S	Nitrobenzene-d5	0.477	0.457	4.2	115	0.00
24	Nitrobenzene	0.481	0.455	5.4	113	0.00
25	Isophorone	0.924	0.837	9.4	110	0.00
26 C	2-Nitrophenol	0.184	0.195	-6.0	126	0.00
27	2,4-Dimethylphenol	0.322	0.331	-2.8	120	0.00
28	bis(2-Chloroethoxy)methane	0.463	0.441	4.8	115	0.00
29 C	2,4-Dichlorophenol	0.342	0.356	-4.1	125	0.00
30	1,2,4-Trichlorobenzene	0.394	0.398	-1.0	124	0.00
31	Naphthalene	1.069	1.063	0.6	120	0.00
32	Benzoic acid	0.170	0.179	-5.3	120	0.00
33	4-Chloroaniline	0.480	0.487	-1.5	120	0.00
34 C	Hexachlorobutadiene	0.293	0.295	-0.7	124	0.00
35	Caprolactam	0.117	0.117	0.0	117	0.00
36 C	4-Chloro-3-methylphenol	0.389	0.400	-2.8	124	0.00
37	2-Methylnaphthalene	0.841	0.863	-2.6	125	0.00
38	1-Methylnaphthalene	0.792	0.805	-1.6	123	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
40	1,2,4,5-Tetrachlorobenzene	0.710	0.762	-7.3	125	0.00
41 P	Hexachlorocyclopentadiene	0.287	0.328	-14.3	125	0.00
42 S	2,4,6-Tribromophenol	0.284	0.315	-10.9	131	0.00
43 C	2,4,6-Trichlorophenol	0.433	0.468	-8.1	123	0.00
44	2,4,5-Trichlorophenol	0.509	0.534	-4.9	119	0.00
45 S	2-Fluorobiphenyl	1.485	1.546	-4.1	122	0.00
46	1,1'-Biphenyl	1.484	1.542	-3.9	121	0.00
47	2-Chloronaphthalene	1.104	1.138	-3.1	120	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051523\
 Data File : BG057436.D
 Acq On : 15 May 2023 17:35
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_G
 LabSampleId :
 SSTDCCC040

Quant Time: May 16 02:15:14 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri May 12 22:04:03 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.375	0.381	-1.6	116	0.00
49	Acenaphthylene	1.793	1.861	-3.8	121	0.00
50	Dimethylphthalate	1.644	1.587	3.5	116	0.00
51	2,6-Dinitrotoluene	0.336	0.367	-9.2	127	0.00
52 C	Acenaphthene	1.120	1.142	-2.0	119	0.00
53	3-Nitroaniline	0.336	0.348	-3.6	122	0.00
54 P	2,4-Dinitrophenol	0.182	0.178	2.2	108	0.00
55	Dibenzofuran	1.882	1.948	-3.5	122	0.00
56 P	4-Nitrophenol	0.224	0.235	-4.9	114	0.00
57	2,4-Dinitrotoluene	0.480	0.497	-3.5	121	0.00
58	Fluorene	1.551	1.586	-2.3	121	0.00
59	2,3,4,6-Tetrachlorophenol	0.461	0.500	-8.5	124	0.00
60	Diethylphthalate	1.669	1.620	2.9	115	0.00
61	4-Chlorophenyl-phenylether	0.934	0.936	-0.2	117	0.00
62	4-Nitroaniline	0.339	0.349	-2.9	117	0.00
63	Azobenzene	1.610	1.495	7.1	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	122	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.133	-11.8	122	0.00
66 c	n-Nitrosodiphenylamine	0.543	0.553	-1.8	120	0.00
67	4-Bromophenyl-phenylether	0.244	0.252	-3.3	121	0.00
68	Hexachlorobenzene	0.237	0.253	-6.8	126	0.00
69	Atrazine	0.207	0.214	-3.4	122	0.00
70 C	Pentachlorophenol	0.137	0.137	0.0	112	0.00
71	Phenanthrene	1.025	1.028	-0.3	119	0.00
72	Anthracene	1.052	1.059	-0.7	118	0.00
73	Carbazole	0.885	0.892	-0.8	120	0.00
74	Di-n-butylphthalate	1.019	1.043	-2.4	120	0.00
75 C	Fluoranthene	1.275	1.280	-0.4	119	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	122	0.00
77	Benzidine	0.447	0.493	-10.3	152#	0.00
78	Pyrene	1.453	1.428	1.7	119	0.00
79 S	Terphenyl-d14	1.226	1.194	2.6	119	0.00
80	Butylbenzylphthalate	0.461	0.501	-8.7	128	0.00
81	Benzo(a)anthracene	1.425	1.421	0.3	118	0.00
82	3,3'-Dichlorobenzidine	0.456	0.469	-2.9	120	0.00
83	Chrysene	1.341	1.341	0.0	120	0.00
84	Bis(2-ethylhexyl)phthalate	0.682	0.726	-6.5	124	0.00
85 c	Di-n-octyl phthalate	1.137	1.208	-6.2	125	0.00
86 I	Perylene-d12	1.000	1.000	0.0	125	0.00
87	Indeno(1,2,3-cd)pyrene	1.458	1.475	-1.2	123	0.00
88	Benzo(b)fluoranthene	1.307	1.305	0.2	121	0.00
89	Benzo(k)fluoranthene	1.329	1.277	3.9	117	0.00
90 C	Benzo(a)pyrene	1.277	1.247	2.3	119	0.00
91	Dibenzo(a,h)anthracene	1.214	1.223	-0.7	122	0.00
92	Benzo(g,h,i)perylene	1.185	1.198	-1.1	123	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_G\Data\BG051523\
Data File : BG057436.D
Acq On : 15 May 2023 17:35
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_G
LabSampleId :
SSTDCCC040

Quant Time: May 16 02:15:14 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_G\Methods\8270-BG042723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri May 12 22:04:03 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0